



Xylose-Lysine-Deoxycholate Agar

MH031

Intended use

Recommended as a selective medium for the isolation and enumeration of *Salmonella* Typhi and other *Salmonella* species from pharmaceutical products in accordance with the microbial limit testing by harmonized methodology of USP/EP/BP/JP/IP.

Composition**

Ingredients	Gms / Litre
Xylose	3.500
L-lysine	5.000
Lactose monohydrate	7.500
Sucrose	7.500
Sodium chloride	5.000
Yeast extract	3.000
Sodium deoxycholate	2.500
Sodium thiosulphate	6.800
Ferric ammonium citrate	0.800
Phenol red	0.080
Agar	13.500
pH after heating (at 25°C)	7.4±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 54.8 grams (the equivalent weight of dehydrated medium per litre) in 1000 ml purified/ distilled water. Heat with frequent agitation until the medium boils. DO NOT HEAT IN AN AUTOCLAVE. Transfer immediately to a water bath at 50°C. After cooling, pour into sterile Petri plates. It is advisable not to prepare large volumes, which will require prolonged heating and may produce precipitate.

Note: Slight precipitation in the medium may occur, which is inheritant property of the medium, and does not affect the performance of the medium.

Principle And Interpretation

Enterobacteriaceae is a family of gram-negative, non-spore-forming bacilli that contains more than 100 species of bacteria that normally inhabit the intestines of humans and animals. Members forming part of the normal intestinal microflora are referred to as coliforms. The clinically significant genera of *Enterobacteriaceae* include *Cedecea*, *Citrobacter*, *Edwardsiella*, *Enterobacter*, *Escherichia*, *Ewingella*, *Hafnia*, *Klebsiella*, *Kluyvera*, *Proteus*, *Salmonella*, *Shigella* and *Yersinia* (8).

The *Salmonellae* are the most complex of all the *Enterobacteriaceae*. Human *Salmonella* infections are most commonly caused by ingestion of food, water or milk, contaminated by human or animal excreta (6). A large number of media have been developed for the selective isolation and identification of enteric bacilli including *Salmonella*.

Xylose Lysine Deoxycholate Agar is a selective as well as differential medium formulated by Taylor (9-13) for the isolation and identification of enteric pathogens especially *Shigellae* from stool samples. It is also used for pharmaceutical testing and non-sterile product testing for the detection (or absence) of *Salmonella* after enrichment in Rappaport Vassiliadis *Salmonella* Enrichment Broth (MH1491) in accordance with the harmonized method of USP/EP/BP/JP/IP (1,2,3,5,14). Deoxycholate, ferric ammonium citrate and sodium thiosulphate are selective agents that inhibit gram-positive microorganisms. Essential nutrients, growth factors for growth of microorganism are provided by yeast extract. Xylose, sucrose and lactose are the fermentable sugars in this medium. Xylose is fermented by almost all the enteric bacteria except *Shigellae*, which enable the differentiation of *Shigellae* from *Salmonellae*. *Salmonellae* metabolize the xylose and decarboxylate lysine and thus change the pH to alkaline and mimic *Shigellae* reaction. However to prevent this reaction by lysine positive coliforms, lactose and sucrose are added in excess to produce acid and hence nonpathogenic H₂S producers do not decarboxylate lysine. Sodium thiosulphate helps in reactivation of sulphur containing compounds and prevents the desiccation of these compounds during storage. It also forms the substrate for enzyme thiosulphate reductase, which breaks it to form H₂S. Thiosulphate and ferric ammonium citrate are the H₂S indicators in the medium. Sodium thiosulphate is also

inactivator of halogens, mercurial and aldehyde and can minimize its toxicity in the testing sample, if any during microbial limit tests. Sodium chloride maintains the osmotic equilibrium in this medium. Phenol red is the pH indicator. Degradation of fermentable sugars proceed concurrently and generates acids, which cause pH indicator to give various shades of colour, causing a color change in the colonies and in the medium from red to yellow on prolonged incubation. Hydrogen sulfide production results in colonies with black centers under alkaline conditions, which can be inhibited by acid production by carbohydrate fermentation. Alkaline condition causes the color of the medium to change back to red. This medium is an ideal medium for screening samples containing mixed flora of enteric pathogens as recovery of *Salmonellae* and *Shigellae* is not conspicuous by even profuse growth of other species (4,7).

Type of specimen

Pharmaceutical samples; Clinical samples ; faeces

Specimen Collection and Handling

For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per guidelines (1,2,3,5,14). For clinical samples, follow appropriate techniques for sample collection and processing as per guidelines (4,6). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/ protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Slight precipitation in the medium may occur, which is inheritant property of the medium, and does not affect the performance of the medium.
2. This medium is general purpose medium and may not support the growth of fastidious organisms.
3. Some *Proteus* strains may give red to yellow colouration with most colonies developing black centers, giving rise to false positive reactions. Non-enterics like *Pseudomonas* and *Providencia* may exhibit red colonies.
4. *S.Paratyphi A*, *S.Choleraesuis*, *S.Pullorum* and *S.Gallinarum* may form red colonies without H₂S, thus resembling *Shigella* species.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to pink homogeneous free flowing powder

Gelling

Firm, comparable with 1.35% Agar gel

Colour and Clarity of prepared medium

Red coloured clear to slightly opalescent gel forms in Petri plates

pH

7.20-7.60

Cultural Response

Growth Promotion is carried out in accordance with the harmonized method of USP/EP/BP/JP/IP. Cultural response was observed after an incubation at 30-35°C for specified time. Recovery rate is considered as 100% for bacteria growth on Soyabean Casein Digest Agar.

Growth promoting properties

Growth of microorganism comparable to that previously obtained with previously tested and approved lot of medium occurs at the specified temperature for not more than the shortest period of time specified inoculating ≤ 100 cfu (at 30-35°C for ≤ 18 hours).

Indicative properties

Colonies are comparable in appearance and indication reaction to those previously obtained with previously tested and approved lot of medium occurs for the specified temperature for a period of time within the range specified inoculating ≤ 100 cfu (at 30-35°C for 18-72 hours).

Inhibitory properties

No growth of the test microorganism occurs for the specified temp for not less than longest period of time specified inoculating ≥ 100 cfu (at 30-35°C for ≥ 72 hours).

Cultural Response

MH031: Cultural characteristics observed after incubation at 30-35 °C for 18-48 hours. Recovery rate is considered as 100% for bacteria growth on Soyabean Casein Digest Agar.

Organism	Inoculum (CFU)	Growth	Observed Lot value (CFU)	Recovery	Colour of Colony	Incubation temperature
Growth Promoting + Indicative						
<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50 -100	luxuriant	25 -100	≥ 50 %	red with black centres	18 -72 hrs
<i>Salmonella Abony</i> NCTC 6017 (00029*)	50 -100	good-luxuriant	25 -100	≥ 50 %	red with black centres	18 -72 hrs
Additional Microbiological testing						
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	fair	10 -30	20 -30 %	yellow	18 -72 hrs
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	fair	10 -30	20 -30 %	yellow	18 -72 hrs
<i>Escherichia coli</i> NCTC 9002	50 -100	fair	10 -30	20 -30 %	yellow	18 -72 hrs
<i>Proteus vulgaris</i> ATCC 13315	50 -100	good-luxuriant	25 -100	≥ 50 %	grey with black centres	18 -72 hrs
<i>Salmonella Paratyphi A</i> ATCC 9150	50 -100	good-luxuriant	25 -100	≥ 50 %	red	18 -72 hrs
<i>Salmonella Paratyphi B</i> ATCC 8759	50 -100	good-luxuriant	25 -100	≥ 50 %	red with black centres	18 -72 hrs
<i>Salmonella Enteritidis</i> ATCC 13076 (00030*)	50 -100	good-luxuriant	25 -100	≥ 50 %	red with black centres	18 -72 hrs
<i>Salmonella Typhi</i> ATCC 6539	50 -100	good-luxuriant	25 -100	≥ 50 %	red with black centres	18 -72 hrs
<i>Shigella dysenteriae</i> ATCC 13313	50 -100	good-luxuriant	25 -100	≥ 50 %	red	18 -72 hrs
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50 -100	fair-good	15 -40	30 -40 %	red	18 -72 hrs
<i>Shigella sonnei</i> ATCC 25931	50 -100	fair-good	15 -40	30 -40 %	red	18 -72 hrs
# <i>Klebsiella aerogenes</i> ATCC 13048 (00097*)	50 -100	fair	10 -30	20 -30 %	yellow	18 -72 hrs
<i>Enterobacter cloacae</i> ATCC 13047	50 -100	fair	10 -30	20 -30 %	yellow	18 -72 hrs
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	$\geq 10^3$	inhibited	0	0%		≥ 72 hrs
<i>Staphylococcus aureus</i> ATCC 6538 subsp. <i>aureus</i> (00032*)	$\geq 10^3$	inhibited	0	0%		≥ 72 hrs
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	$\geq 10^3$	inhibited	0	0%		≥ 72 hrs

Key: (#) Formerly known as *Enterobacter aerogenes* (*) Corresponding WDCM numbers

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,6).

Reference

1. British Pharmacopoeia, 2016, The Stationery office British Pharmacopoeia.
2. European Pharmacopoeia, 2017, European Dept. for the quality of Medicines.
3. Indian Pharmacopoeia, 2018 Ministry of Health and Family Welfare, Govt. of India.
4. Isenberg H.D., Kominos S. and Siegal M., 1969, Appl. Microbiol., 18:656.
5. Japanese Pharmacopoeia, 2016.
6. Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4th Ed., J. B. Lippincott Company.
7. McCarthy M.D., 1966, N.Z. J. Med. Lab. Technol., 20:127.
8. Pelczar M. J. Jr., Reid R. D., Chan E. C. S., 1977, Microbiology, 4th Ed., Tata McGraw-Hill Publishing Company Ltd, New Delhi.
9. Taylor W. L., 1965, Am. J. Clin. Pathol., 44:471-475.
10. Taylor W. L. and Harris B., 1965, Am. J. Clin. Pathol., 44:476.
11. Taylor W. L. and Harris B., 1967, Am. J. Clin. Pathol., 48:350.
12. Taylor W. L. and Schelhart B., 1967, Am. J. Clin. Pathol., 48:356.
13. Taylor W. L. and Schelhart B., 1968, Am. J. Clin. Pathol., 16:1387.2. ,
14. The United States Pharmacopoeia, 2019, The United States Pharmacopoeial Convention. Rockville, MD.

Revision : 03 / 2019

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.