



Soybean-Casein Digest Agar (Casein Soyabean Digest Agar)

MH290

Intended use

Recommended as a general purpose medium used for cultivation of a wide variety of microorganisms from pharmaceutical products in accordance with harmonized method of USP/EP/BP/JP/IP (Medium 2).

Composition**

Ingredients	Gms / Litre
Tryptone #	15.000
Soya peptone ##	5.000
Sodium chloride	5.000
Agar	15.000
pH after sterilization (at 25°C)	7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Pancreatic digest of casein

Papaic digest of soyabean (soybean)

Directions

Suspend 40 grams in 1000 ml purified/ distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes or as per validated cycle. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

Principle And Interpretation

Various pharmacopoeias recommend Soybean Casein Digest Agar as sterility testing medium. It is also used in validation of sterility checking procedure in accordance with the microbial limit testing harmonized methodology of USP/EP/BP/JP/IP (7,2,1,5,3). This medium is used in microbial limit test and antimicrobial preservative- effective test. Gunn et al (5) used this medium for the growth of fastidious organisms and study of haemolytic reaction after addition of 5% v/v blood.

The combination of tryptone and soya peptone makes these media nutritious by providing amino acids and long chain peptides for the growth of microorganisms. Natural sugars of soy enhance growth of microorganism. Sodium chloride maintains the osmotic balance in the medium. Agar is the solidifying agent

The total aerobic count is considered to be equal to the number of colony forming units found on this medium, if colonies of fungi are detected on this medium they are counted along with total aerobic count.

Type of specimen

Pharmaceutical samples; Clinical samples

Specimen Collection and Handling

For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per pharmaceutical guidelines (7,2,1,5,3). For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,6). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Biochemical characterization is necessary to be performed on colonies from pure cultures for further identification.
2. This medium is general purpose medium and may not support the growth of fastidious organisms.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Light yellow coloured clear to slightly opalescent gel forms in Petri plates

pH

7.10-7.50

Growth Promotion Test

Growth Promotion was carried out in accordance with the harmonized method of USP/EP/BP/JP, and growth was observed after an incubation at 30-35°C for 18-24 hours. Recovery rate is considered 100% for bacteria growth on Blood Agar and fungus growth on Sabouraud Dextrose Agar.

Growth promoting properties

Growth of microorganism comparable to that previously obtained with previously tested and approved lot of medium occurs at the specified temperature for not more than the shortest period of time specified inoculating ≤ 100 cfu (at 30-35°C for 18 hours).

Cultural Response

Organism	Inoculum (CFU)	Observed Lot value (CFU)	Recovery	Incubation period
<i>Bacillus subtilis subsp. spizizenii</i> ATCC 6633 (00003*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Staphylococcus aureus subsp. aureus</i> ATCC 25923 (00034*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Staphylococcus aureus subsp. aureus</i> ATCC 6538 (00032*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Escherichia coli</i> NCTC 9002	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Salmonella</i> Abony NCTC 6017 (00029*)	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Micrococcus luteus</i> ATCC 9341	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Streptococcus pneumoniae</i> ATCC 6305	50 -100	35 -100	≥ 70 %	18 -24 hrs
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	35 -100	≥ 70 %	18 -24 hrs

<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	35 -100	>=70 %	<=5 d
<i>Candida albicans</i> ATCC 2091 (00055*)	50 -100	35 -100	>=70 %	<=5 d
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50 -100	25 -70	50-70 %	<=5 d

Key : (#) Formerly known as *Aspergillus niger*, (*) Corresponding WDCM numbers

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,6).

Reference

1. British Pharmacopoeia, 2016, The Stationery office British Pharmacopoeia
2. European Pharmacopoeia, 2017, European Dept. for the quality of Medicines.
3. Indian Pharmacopoeia, 2018, Govt. of India, the controller of Publication, Delhi, India.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
5. Japanese Pharmacopoeia, 2016.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
7. The United States Pharmacopoeia, 2019, The United States Pharmacopoeial Convention. Rockville, MD.




In vitro diagnostic medical device

Revision : 04 / 2019



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.