

## Bismuth Sulphite Agar, Granulated

**M027**

### Intended Use:

Bismuth Sulphite Agar, Granulated is recommended for the selective isolation and preliminary identification of *Salmonella* Typhi and other Salmonellae from pathological materials, sewage, water supplies, food etc.

### Composition\*\*

Ingredients	Gms / Litre
Peptone	10.000
HM Peptone B #	5.000
Dextrose (Glucose)	5.000
Disodium phosphate	4.000
Ferrous sulphate	0.300
Bismuth sulphite indicator	8.000
Brilliant green	0.025
Agar	20.000
Final pH ( at 25°C)	7.7±0.2

\*\*Formula adjusted, standardized to suit performance parameters

# - Equivalent to Beef Extract

### Directions

Suspend 52.33 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. DO NOT STERILIZE IN AUTOCLAVE or by fractional sterilization since overheating may destroy the selectivity of the medium.

The sensitivity of the medium depends largely upon uniform dispersion of precipitated bismuth sulphite in the final gel, which should be dispersed before pouring into sterile Petri plates.

### Principle And Interpretation

The Salmonellae constitute the most taxonomically complex group of bacteria among *Enterobacteriaceae* (1). Human Salmonella infections are most commonly caused by ingestion of food, water or milk contaminated by human or animal excreta. Humans are the only reservoirs of *S. Typhi* (2). Four clinical types of *Salmonella* infections may be distinguished (3) namely gastroenteritis, bacteremia or septicemia, enteric fever and a carrier state. Of the various media employed for the isolation and preliminary identification of Salmonellae, particularly *Salmonella* Typhi; Bismuth Sulphite Agar is the most productive (4).

Bismuth Sulphite Agar is a modification of original Wilson and Blair Medium (5-7). It is also recommended by various Associations (8-13) for the isolation and preliminary identification of *Salmonella* Typhi and other Salmonellae from pathological materials, sewage, water, food and other products.

*S. Typhi*, *S. Enteritidis* and *S. Typhimurium* typically grow as black colonies with a surrounding metallic sheen resulting from hydrogen sulphide production and reduction of sulphite to black ferric sulphide. *Salmonella* Paratyphi A grows as light green colonies. Bismuth Sulphite Agar may be inhibitory to some strains of *Salmonella* species and therefore should not be used as the sole selective medium for these organisms. Also this medium favors use of larger inoculum as compared to other selective media, as it has unique inhibitory action towards gram-positive organisms and coliforms.

Peptone and HM Peptone B serve as sources as carbon, nitrogen, long chain amino acids, vitamins and essential growth factors. Dextrose is the carbon source. Disodium phosphate maintains the osmotic equilibrium. Bismuth sulphite indicator along with brilliant green inhibits the intestinal gram-positive and gram-negative bacteria. Ferrous sulphate aids in detection of hydrogen sulphide production. Clinical samples can be directly used to inoculate Bismuth Sulphite Agar. In case of food samples, pre enrichment of the sample is done prior to inoculation.

### Type of specimen

Clinical samples : faeces, urine, blood and other pathological material , Foodstuff , water samples

## Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (16,17).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (10,12,15).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(9)

After use, contaminated materials must be sterilized by autoclaving before discarding.

## Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

## Limitations

1. DO NOT AUTOCLAVE OR OVERHEAT THE MEDIUM, as it destroys the selectivity of the medium.
2. *S.Typhi* and *S.Arizonae* exhibit typical brown colonies , with or without metallic sheen.
- 3.This medium is highly selective and must be used in parallel with less selective media for isolation.
4. With certain *Salmonella* species, typical black colonies with metallic sheen is observed near heavy inoculation and isolated colonies may show green colonies.
5. *Shigella* species are mostly inhibited on this medium; exceptions being *S. flexneri* and *S. sonnei* (14)
6. Some *Salmonella* like *S. Sendai*, *S. Berta*, *S. Gallinarum*, *S. Abortus-equi* are also inhibited (14).

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Light yellow to greenish yellow homogeneous granular medium

### Gelling

Firm, comparable with 2.0% agar gel.

### Colour and Clarity of prepared medium

Greenish yellow coloured, opalescent with flocculent precipitate forms in Petri plates.

### Reaction

Reaction of 5.23% w/v aqueous solution at 25°C. pH : 7.7±0.2

### pH

7.50-7.90

### Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 40-48 hours.

### Cultural Response

Organism	Inoculum (CFU)	Growth	Recovery	Colour of Colony
<b>Cultural Response</b> <i>Enterobacter aerogenes</i> ATCC 13048 (00175*)	50-100	none-poor	≤10%	brown-green (depends on the inoculum density)
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	≥10 <sup>3</sup>	inhibited	0%	
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	none-poor	≤10%	brown-green (depends on the inoculum density)
<i>Salmonella Enteritidis</i> ATCC 13076 (00030*)	50-100	good-luxuriant	≥50%	black with metallic sheen
<i>Salmonella Typhi</i> ATCC 6539	50-100	good-luxuriant	≥50%	black with metallic sheen

<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50-100	good-luxuriant	>=50%	black with metallic sheen
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50-100	none-poor	<=10%	brown
<i>Escherichia coli</i> ATCC 8739 (00012*)	50-100	none-poor	<=10%	brown to green, depends on inoculum density
<i>Escherichia coli</i> NCTC 9002	50-100	none-poor	<=10%	brown to green (depends on inoculum density)
<i>Salmonella Abony</i> NCTC 6017 (00029*)	50-100	good-luxuriant	>=50%	black with metallic sheen

Key : \*Corresponding WDCM numbers.

## Storage and Shelf Life

Store below 30°C in a tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (16,17).

## Reference


- Tindall B. J., Crimont P. A. D., Gorry G. M., EUZESY B. P., 2005, Int. J. Sys. Evol. Microbiol., 55:521
- Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4 th Ed., J. B. Lippincott Company
- Mandell G. L., Douglas R. G. Jr., Bennet J. E., (Eds.), 1985, Principles and Practice of Infectious Diseases, 2nd Ed., 660-669, John Wiley & Sons New York.
- Gunter and Tuft, 1939, J. Lab. Clin. Med., 24:461.
- Wilson and Blair, 1926, J. Pathol. Bacteriol., 29:310.
- Wilson and Blair, 1927, J. Hyg., 26:374
- Wilson and Blair, 1931, J. Hyg., 31:138
- Washington J. A., 1981, Laboratory Procedures in Clinical Microbiology, Springer-Verlag, New York.
- Greenberg A. E., Clesceri L. S. and Eaton A. D., (Eds.), 1998, Standard Methods for the Examination of Water and Wastewater, 20th Ed., APHA, Washington D.C.,
- FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, D.C.
- Murray P. R., Baron J. H., Pfaller M. A., Tenover F. C. and Tenover R. H., (Eds.). 1999, Manual of Clinical Microbiology, 7th Ed., American Society for Microbiology, Washington, D.C.
- Downes F. P. and Ito K., (Eds.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th Ed., APHA, Washington, D.C.
- Indian Pharmacopoeia, 1996, Ministry of Health and Family Welfare, Govt. of India, Volume 2.
- MacFaddin J. F., 2000, (Ed.), Biochemical Tests for Identification of Medical Bacteria, 3rd Edition, Lippincott, Williams & Wilkins, New York.
- American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
- Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.



In vitro diagnostic medical device



CE Marking




Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
23 Vadhani Industrial Estate,  
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.