

<i>Staphylococcus aureus</i> <i>subsp.aureus</i> ATCC 6538 (00032*)	>=10 ³	inhibited	0%	
<i>Salmonella Paratyphi B</i> ATCC 8759	50 -100	luxuriant	>=50 %	colourless
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	luxuriant	>=50 %	pink to red with bile precipitate
<i>Escherichia coli</i> NCTC 9002	50 -100	luxuriant	>=50 %	pink to red with bile precipitate
# <i>Klebsiella aerogenes</i> ATCC 13048 (00175*)	50 -100	luxuriant	>=50 %	pink to red
<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50 -100	luxuriant	>=50 %	colourless
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50 -100	none - poor	<=10 %	colourless to pale pink
<i>Salmonella Enteritidis</i> ATCC 13076 (00030*)	50 -100	luxuriant	>=50 %	colourless
<i>Staphylococcus aureus</i> <i>subsp.aureus</i> ATCC 25923 (00034*)	>=10 ³	inhibited	0%	

Key :- * Corresponding WDCM numbers

Storage and Shelf Life

Store below 30°C in a tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal







User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,15).

Reference

- Murray P. R, Baron E, J., Jorgensen J. H., Pfaller M. A., Tenover F. C., Tenover F. C., (Eds.), 8th Ed., 2003, Manual of Clinical Microbiology, ASM, Washington, D.C.
- Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
- Salinger Y., and Tortorello M.L. Fifth Edition, 2015, 015 American Public Health Association, Washington, D.C.
- FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, D.C.
- Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- The United States Pharmacopoeia, 2015 The United States Pharmacopoeial Convention, Rockville, M.D.
- Williams, (Ed.), 2005, Official Methods of Analysis of the Association of Official Analytical Chemists, 19th Ed., AOAC, Washington, D.C.
- Medrek T. F and Barnes Ella M., 1962, J. Appl. Bacteriol., 25(2),159-168
- Barnes Ella M. and Shrimpton D. H., 1957, J. Appl. Bacteriol., 20(2),273-285.
- Thornley Margaret J., 1957, J. Appl. Bacteriol., 20(2), 273-285.
- Eddy B. P., 1960, J. Appl. Bacteriol., 23(2).216-249.
- MacConkey A., 1905, J. Hyg., 5:333.
- MacConkey A., 1900, The Lancet, ii:20.
- British Pharmacopoeia, 2015, The Stationery office British Pharmacopoeia.

+ U G P D G T I * & % N K P K E C N / K E T Q D K Q N Q I [P 2 T Q E G F W T G U * C P F D Q Q M P F ' F K V

Revision : 01 / 2018

	In vitro diagnostic medical device
	CE Marking
	Storage temperature
	Do not use if package is damaged
	HiMedia Laboratories Pvt. Limited, 23 Vadhani Industrial Estate, LBS Marg, Mumbai-86, MS, India
	CE Partner 4U ,Esdoomlaan 13, 3951 DB Maarn The Netherlands, www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.