

Mannitol Salt Agar, Granulated

GM118

Intended Use:

Mannitol Salt Agar, Granulated is used as a selective media for the isolation of pathogenic Staphylococci from clinical and non-clinical samples..

Composition**

Ingredients	Gms / Litre
Proteose peptone	10.000
HM peptone B #	1.000
Sodium chloride	75.000
D-Mannitol	10.000
Phenol red	0.025
Agar	15.000
Final pH (at 25°C)	7.4±0.2

**Formula adjusted, standardized to suit performance parameters

- Equivalent to Beef extract

Directions

Suspend 111.02 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. If desired, add 5% v/v Egg Yolk Emulsion (FD045). Mix well and pour into sterile Petri plates.

Note : This product contains 7.5% Sodium chloride as one of its ingredients. On repeated exposure to air and absorption moisture sodium chloride has tendency to form lumps, therefore we strongly recommend storage in tightly closed containers in dry place away from bright light .

Principle And Interpretation

Staphylococci are widespread in nature, although they are mainly found on the skin, skin glands and mucous membranes of mammals and birds. The coagulase-positive species i.e *Staphylococcus aureus* is well documented as a human opportunistic pathogen. The ability to clot plasma continues to be the most widely used and accepted criterion for the identification of pathogenic staphylococci associated with acute infections (1). Staphylococci have the unique ability of growing on a high salt containing media (2). Isolation of coagulase-positive staphylococci on Phenol Red Mannitol Agar supplemented with 7.5% NaCl was studied by Chapman (3). The resulting Mannitol Salt Agar Base is recommended for the isolation of coagulase-positive staphylococci from cosmetics, milk, food and other specimens (1, 4-7). The additional property of lipase activity of *Staphylococcus aureus* can be detected by the addition of the Egg Yolk Emulsion (FD045). The lipase activity can be visualized as yellow opaque zones around the colonies (8).

HM peptone B and proteose peptone supply essential growth factors and trace nutrients to the growing bacteria. Sodium chloride serves as an inhibitory agent against bacteria other than staphylococci. Mannitol is the fermentable carbohydrate, fermentation of which leads to acid production, detected by phenol red indicator.

S.aureus ferment mannitol and produce yellow coloured colonies surrounded by yellow zones. Coagulase-negative strains of *S.aureus* are usually mannitol non-fermenters and therefore produce pink to red colonies surrounded by red-purple zones. Presumptive coagulase-positive yellow colonies of *S. aureus* should be confirmed by performing the coagulase test [tube or slide](1). Lipase activity of *S.aureus* can be detected by supplementing the medium with egg yolk emulsion.

A possible *S.aureus* must be confirmed by the coagulase test. Also the organism should be subcultured to a less inhibitory medium not containing excess salt to avoid the possible interference of salt with coagulase testing or other diagnostic tests (e.g. Nutrient Broth) (M002) (9). Few strains of *S.aureus* may exhibit delayed mannitol fermentation. Negative results should therefore be re-incubated for an additional 24 hours before being discarded (9).

Type of specimen

Clinical samples: pus; Food and dairy samples, water samples.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (10,11).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (6,12).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

NA

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to pink homogeneous granular media

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Red coloured clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 11.1% w/v aqueous solution at 25°C. pH : 7.4±0.2

pH

7.20-7.60

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-72 hours. Recovery rate is considered as 100% for bacteria growth on Soybean Casein Digest Agar.

Cultural Response

Organism	Inoculum (CFU)	Growth	Observed Lot value (CFU)	Recovery	Colour of colony	Incubation temperature
Cultural Response						
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	50 -100	luxuriant	25 -100	≥50 %	yellow/white colonies surrounded by yellow zone	18 -72 hrs
<i>Escherichia coli</i> ATCC 8739 (00012*)	≥10 ³	inhibited	0	0 %		≥72 hrs
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50 -100	luxuriant	25 -100	≥50 %	yellow/white colonies surrounded by yellow zone	18 -72 hrs
<i>Staphylococcus epidermidis</i> ATCC 12228 (00036*)	50 -100	fair - good	15 -40	30 -40 %	red	18 -72 hrs
<i>Staphylococcus epidermidis</i> ATCC 14990 (00132*)	50 -100	fair - good	15 -40	30 -40 %	red	18 -72 hrs
<i>Proteus mirabilis</i> ATCC 12453	50 -100	none-poor	0 -10	0 -10 %	yellow	18 -72 hrs
<i>Escherichia coli</i> ATCC 25922 (00013*)	≥10 ³	inhibited	0	0%		≥72 hrs

<i>Escherichia coli</i> NCTC 9002 $\geq 10^3$	inhibited	0	0%	≥ 72 hrs
# <i>Klebsiella aerogenes</i> ATCC 13048 (00175*) $\geq 10^3$	inhibited	0	0%	≥ 72 hrs

Key : *Corresponding WDCM numbers. #- Formerly known as *Enterobacter aerogenes*

Storage and Shelf Life

Store below 30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (10,11).

Reference

- Murray P. R., Baron J. H., Pfaller M. A., Tenover J. C. and Tenover F. C., (Ed.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C. ,
- Koch P. K., 1942, Zentralbl. Bakteriol. Parasitenkd. Abt. I Orig. 149:122.
- Chapman G. H., 1945, J. Bacteriol., 50:201.
- Hitchins A. D., Tran T. and McCarron J. E., 1995, FDA Bacteriological Analytical Manual, 8th Ed., AOAC International, Gaithersburg, Md.
- Davis J. G., 1959, Milk testing, 2nd Ed., Dairy Industries Ltd, London.
- Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
- Silverton R. E. and Anderson M. J., 1961, Handbook of Medical Laboratory Formulae, Butterworths, London.
- Gunn B. A., Dunkelberg W. E. and Creitz J. R., 1972, Am. J. Clin. Pathol., 57:236.
- MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore
- Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology,
- American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.


Revision : 02/ 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.