



Phenol Red Indicator

I010

Intended Use:

It is recommended as pH indicator which helps in monitoring of the pH changes in the cell culture.

Composition**

Ingredients

Phenol red sodium salt	0.04gm
Disilled water	100.0ml

Note: Grind the indicator in mortar and then dilute to 100ml by using disilled water.

**Formula adjusted, standardized to suit performance parameters

Principle And Interpretation

Phenol red is a pH indicator frequently used in cell biology laboratories. Its color exhibits a gradual transition from yellow to red over the pH range 6.8 to 8.2. Above pH 8.2, phenol red turns a bright pink (fuchsia) color.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Performance and Evaluation

Performace of the product is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Reddish orange coloured solution.

Clarity

Clear without any particles.

Reaction

At pH 6.8 the indicator is yellow and at pH 8.4, the indicator turns red.

Sensitivity (As per IP)

A mixture of 0.1ml of the solution and 100ml of carbone dioxide- free water is yellow. Not more than 0.1ml of 0.02M sodium hyroxide is required to change the colour of the solution to reddish violet.

Storage and Shelf Life

Store between 10- 30°C in tightly closed container and away from bright light. Use before expiry date on label. On opening, product should be properly stored in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (2,3).

Reference

1. Text Book of Medical Laboratory Technology; Praful B. Godkar
2. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
3. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual

Revision : 03 / 2018

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.