

Reference

- +DKQ ++ &KHXN 6) (OIHQEHLQ 6 :RRG :% \$SULO - ([S 0HG
2. Chemistry infolab reagents and resources ; The preparation of titration indicators; Dhanal De Lloyd, chem. Dept
 3. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
 4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual


Revision < 1 / 2015



In vitro diagnostic medical device



CE Marking



Storage temperature




Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
B /4-6 , MIDC, Dindori, Nashik MH

www.himedialabs.com



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.