



Folic Acid Assay Medium

M038

Intended use

Folic Acid Assay Medium is recommended for microbiological assay of Folic Acid using *Enterococcus hirae* ATCC 8043 (formerly *Streptococcus faecalis* ATCC 8043) as the test organism.

Composition**

Ingredients	Gms / Litre
Acicase, vitamin free#	12.000
Dextrose	40.000
Sodium citrate	20.000
L-Cystine	0.200
DL-Tryptophan	0.200
Adenine sulphate	0.020
Guanine hydrochloride	0.020
Uracil	0.020
Thiamine hydrochloride	0.002
Pyridoxine hydrochloride	0.004
Riboflavin (Vitamin B2)	0.002
Niacin	0.002
p-Amino benzoic acid (PABA)	0.0002
Biotin	0.0000008
Calcium pantothenate	0.0004
Dipotassium phosphate	1.000
Monopotassium phosphate	1.000
Magnesium sulphate	0.400
Sodium chloride	0.020
Ferrous sulphate	0.020
Manganese sulphate	0.020
Final pH (at 25°C)	6.8±0.2

**Formula adjusted, standardized to suit performance parameters

- Equivalent to Casein acid hydrolysate, vitamin free

Directions

Suspend 7.49 grams in 100 ml purified/ distilled water. Heat if necessary to dissolve the medium completely. Mix well to distribute the slight precipitate evenly. For assay, dispense 5 ml medium per assay tube (containing increasing amounts of standard or the unknown) and make total volume to 10 ml with distilled water. Sterilize by autoclaving at 15 lbs pressure (121°C) for 10 minutes. Cool immediately. Satisfactory results are obtained with Folic Acid at levels of 0,2,4,6,8 and 10mg per assay tube (10ml)

Principle And Interpretation

Folic Acid Assay Medium contains all the necessary nutrients for the growth of the test organism except folic acid. The medium contains nutrients like amino acids, carbohydrates, purine, pyrimidines, salts, and vitamins. Folic acid is required for the growth of *Enterococcus hirae*. Hence growth of this organism will occur only if the sample being assayed contains folic acid. The exact folic acid concentration in the test sample can be determined by comparing the growth obtained to that of known standard concentrations of folic acid (standard curve). Folic Acid Assay Medium is prepared according to the formula described by Capps et al (1) and is recommended for the determination of folic acid content of the pharmaceutical products. Standard reference for assay of folic acid should be referred (4).

Procedure : Stock cultures of *Enterococcus hirae* ATCC 8043 are prepared by stab inoculation of Micro Vitamin Test Culture Agar (M132). Following incubation at 35-37°C for 24 hours, the tubes are stored in the refrigerator. Transplants are made at monthly intervals. Inoculum for assay is prepared by subculturing from a stock culture of *Enterococcus hirae* ATCC 8043

into a tube containing 10 ml of Micro Vitamin Test Inoculum Broth (M133). After 24 hours incubation at 35-37°C, the cells are centrifuged under aseptic conditions, and the supernatant liquid is decanted. The cells are resuspended in 10 ml of sterile 0.85% NaCl. The cell suspension is then diluted 1:100 with sterile 0.85% NaCl. One drop of this later suspension is used to inoculate each of the assay tubes. It is essential that a standard curve be set up for each separate assay since conditions of autoclaving, temperature of incubation, etc., which influence the standard curve readings cannot be duplicated exactly from time to time. The standard curve is obtained by using folic acid at levels of 0, 2, 4, 6, 8 and 10 ng per assay tube (10 ml). Tubes are refrigerated for 15-30 minutes to stop growth before reading. Turbidimetric readings should be read after 16-18 hours incubation at 35-37°C and acidimetric after 72 hours at 35-37°C. To prepare stock solution of folic acid, 20 mg folic acid is used.

Preparation of Folic Acid Concentrations:

Dissolve 20 mg dried folic acid in 100 ml distilled water containing 20 ml ethanol. Adjust the pH of the solution to 10.0 with 0.1 N NaOH to dissolve the acid and then adjust pH to 7.0 with 0.05 N HCl. This solution contains 200 mcg folic acid per ml. Dilute 1 ml of this solution with 999 ml of distilled water to get 200 ng per ml and finally, dilute 1 ml of this solution with 999 ml of Folic Acid Buffer A (M544) to get a standard solution containing 0.2 ng folic acid per ml. use 0.0, 0.5, 1.0, 2.0, 3.0, 4.0 and 5.0 ml per assay tube.

Extreme care should be taken to avoid contamination of media or glassware used for the assay. Detergent free clean glassware should be used. Even small amount of contamination by foreign material can be lead to erroneous results.

Type of specimen

Test sample.

Specimen Collection and Handling

Refer procedure and preparation of folic acid concentration.

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Extreme care should be taken to avoid contamination of media or glassware used for the assay.
2. Detergent free clean glassware should be used.
3. Even small amount of contamination by foreign material can be lead to erroneous results.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Off-white to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light amber coloured, clear solution, which may contain a slight precipitate.

Reaction

Reaction of 7.5% w/v aqueous solution at 25°C. pH : 6.8±0.2

pH

6.60-7.00

Cultural Response Microbiological Assay of Folic Acid was carried out using *Enterococcus hirae* ATCC 8043 after an incubation at 35-37° C for 16-18 hours

Organism

Growth

Enterococcus hirae
ATCC 8043

Good growth is obtained. Gradual increase in growth with increasing conc. of standard folic acid 0, 2, 4, 6, 8, 10 ng per assay tube was recorded as equivalent increase in absorbance at 620 nm.

Storage and Shelf Life

Store between 2-8°C in a tightly closed container and use freshly prepared medium. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (2,3).

Reference

1. Capps, Hobbs and Fox, 1948, J. Bact., 55:869.
2. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
3. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
4. Official Methods of Analysis of AOAC International, 2005, 19th Ed., Vol. II, Association of Analytical Chemists, Washington, D.C.

Revision :02 / 2019

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.