



Tryptose Blood Agar Base

M097

Intended Use:

Recommended for the isolation of fastidious organisms and determining the haemolytic reactions.

Composition**

Ingredients	Gms / Litre
Tryptose	10.000
HM peptone B #	3.000
Sodium chloride	5.000
Agar	15.000
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

- Equivalent to Meat extract

Directions

Suspend 33 grams in 950 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool the autoclaved medium to 45 - 50°C and aseptically add 5% v/v sterile defibrinated blood. Mix thoroughly, avoiding air bubbles and pour into sterile Petri plates.

Principle And Interpretation

Tryptose Blood Agar Base is a tryptose based medium that can be used for the cultivation of fastidious organisms (1, 2), on supplementation with blood. This medium is devoid of dextrose and therefore useful in determining the haemolytic reactions. Tryptose Blood Agar Base is recommended by FDA (3) and APHA (4). Tryptose Blood Agar Base can be used as a general-purpose medium without supplementation of blood. These media can be used to determine the haemolytic reactions of fastidious organisms. The four different types of haemolysis observed are as follows:

- Alpha haemolysis: partial lysis of the erythrocytes surrounding colony, causing a grey green or brownish discoloration in the media.
- Beta haemolysis: complete lysis of the red blood cells surrounding a colony, causing a clearing of blood from the medium.
- Gamma haemolysis: no haemolysis and consequently, no colour change of the medium surrounding a colony. Organisms showing no haemolysis are generally termed non-hemolytic rather than gamma haemolytic.
- Alpha-prime or wide zone alpha: a small zone of intact erythrocytes immediately adjacent to the colony, with a zone of complete red cell haemolysis surrounding the zone of intact erythrocytes. This type of haemolysis may be confused with beta haemolysis (6).

Tryptose and HM peptone B provide nitrogenous and carbonaceous compounds, sulphur, vitamin B complex and trace elements essential for bacterial metabolism. Blood provides additional nutrients and serves as a base to study haemolytic reactions. This medium not only keeps the blood cells in a good state but also help in forming distinct haemolysis. Perform biochemical test for further identification (5).

Type of specimen

Clinical samples: Blood, CSF.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Some fastidious organisms may not grow.
2. Inoculum density affects inhibition zone. Heavy inoculum may result in smaller zones while scanty growth may result in enlarged zones.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Basal medium : Yellow coloured clear to slightly opalescent gel forms in Petri plates. After addition of 5% v/v sterile defibrinated blood : Cherry red coloured opaque gel forms in Petri plates.

Reaction

Reaction of 3.3% w/v aqueous solution at 25°C. pH : 7.2±0.2

pH

7.00-7.40

Cultural Response

Cultural characteristics observed with added 5% v/v sterile defibrinated blood, after an incubation at 35-37°C for 18-48 hours.

Organism	Inoculum (CFU)	Growth w/o blood	Recovery w/o blood	Growth w/ blood	Haemolysis	Recovery w/ blood
<i>Neisseria meningitidis</i> ATCC 50-100 13090		good-luxuriant	≥70%	luxuriant	none	≥70%
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	good-luxuriant	≥70 %	luxuriant	beta/gamma	≥70%
<i>Staphylococcus epidermidis</i> ATCC 12228 (00036*)	50-100	fgood-luxuriant	≥70%	luxuriant	gamma	≥70%
<i>Streptococcus pneumoniae</i> ATCC 6303	50-100	fair-good	40-50%	good	alpha	50-70%
<i>Streptococcus pyogenes</i> ATCC 19615	50-100	fair-good	40-50%	good	beta	50-70%

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

Reference

1. American Public Health Association, 1970, Diagnostic Procedures and Reagents, 5th Ed., APHA Inc., New York.
2. Casman E. P., 1942, J. Bacteriol., 43:33.
3. Casman E. P., 1947, Am. J. Clin. Pathol., 17: 281.
4. FDA Bacteriological Analytical Manual, 8th Ed., 1995, AOAC International, Gaithersburg, Md.
5. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
7. Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4th Ed., J. B. Lippincott Company.
8. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.


Revision : 03/ 2019



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.