



Urea Broth (Filter Sterilizable)

M111A

Intended use

Recommended for the identification of bacteria on the basis of urea utilization, specifically for the differentiation of *Proteus* species from *Salmonella* and *Shigella* species.

Composition**

Ingredients	Gms / Litre
Potassium dihydrogen phosphate	9.100
Yeast extract	0.100
Dipotassium hydrogen phosphate	9.500
Urea	20.000
Phenol red	0.010
Final pH (at 25°C)	6.8±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 38.7 grams in 1000 ml purified/distilled water. Mix well and sterilize by filtration. DO NOT AUTOCLAVE OR HEAT the medium. Dispense in sterile tubes or flasks as desired.

Principle And Interpretation

Urea Broth (Filter Sterilizable) was developed by Rustigian and Stuart (8). This medium is especially recommended for the differentiation of *Proteus* species from *Salmonella* and *Shigella* species in the enteric infection diagnosis (4), based on urea utilization (3, 7). Gram-negative enteric bacilli are unable to utilize urea because of less nutrients and high buffering capacity of the medium. Urea Broth becomes alkaline as the utilization of urea by the organisms liberate ammonia during the incubation, indicated by pink red colour. All urea test media rely on the alkalinity formation and so they are not specific for urease testing. The utilization of proteins may raise the pH to alkalinity due to protein hydrolysis and excess of amino acids results in false-positive reaction.

Type of specimen

Clinical samples: Faeces; Food & dairy samples; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (5,6).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1,9,10).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards (2).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Prolonged incubation may cause alkaline reaction in the medium.
2. All urea test media rely on the alkalinity formation and so they are not specific for determining the absolute rate of urease activity (7).

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to light pink coloured homogeneous free flowing powder

Colour and Clarity of prepared medium

Yellow to orange coloured clear solution without any precipitate.

Reaction

Reaction of 3.87% w/v aqueous solution at 25°C. pH : 6.8±0.2

pH

6.60-7.00

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism	Inoculum (CFU)	Urease
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	Negative reaction, no change
# <i>Klebsiella aerogenes</i> ATCC 13048 (00175*)	50-100	Negative reaction, no change
<i>Klebsiella pneumoniae</i> ATCC 13883	50-100	Positive reaction, cerise colour
<i>Proteus mirabilis</i> ATCC 12453	50-100	Positive reaction, cerise colour
<i>Proteus vulgaris</i> ATCC 13315	50-100	Positive reaction, cerise colour
<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50-100	Negative reaction, no change

Key : (*) Corresponding WDCM numbers (#) Formerly known as *Enterobacter aerogenes*.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-25°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

Reference

1. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
2. Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
3. Christensen, 1946, J. Bact., 52:461.
4. Finegold and Baron, 1986, Bailey and Scotts Diagnostic Microbiology, 7th ed., The C.V. Mosby Co., St. Louis.
5. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
7. MacFaddin J. F., 2000, Biochemical Tests for Identification of Medical Bacteria, 3rd Ed., Williams and Wilkins, Baltimore.Md.
8. Rustigian and Stuart, 1941, Proc. Soc. Exp. Biol. Med., 47:108.
9. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
10. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.


Revision : 01 / 2019



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.