



Baird Parker Agar with Sulpha

M1140

Intended use

Baird Parker Agar Base with supplements is recommended for the isolation and enumeration of coagulase positive Staphylococci from food and other materials.

Composition**

Ingredients	Gms / Litre
Tryptone	10.000
HM peptone B #	5.000
Yeast extract	1.000
Glycine	12.000
Sodium pyruvate	10.000
Lithium chloride	5.000
Sulphamethazine	0.050
Agar	20.000
Final pH (at 25°C)	7.0±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Beef extract

Directions

Suspend 63.05 grams in 950 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 50°C and aseptically add 50 ml concentrated Egg Yolk Emulsion (FD045) and 3 ml sterile 3.5% Potassium Tellurite solution (FD047) or 50 ml Egg Yolk Tellurite Emulsion (FD046). Mix well and pour into sterile Petri plates.

Principle And Interpretation

Baird Parker Agar was developed by Baird Parker (3,4) from the Tellurite-glycine formulation of Zebovitz et al (16) for isolation and enumeration of Staphylococci in food and other material since it allows a good differentiation of coagulase positive strains. A high correlation has been found between the coagulase test and the presence of clear zone of lypolysis in this medium, which is due to the lecithinase of Staphylococci that breakdown, the egg yolk. On the other hand, studies show that almost 100% of coagulase positive Staphylococci are capable of reducing tellurite, which produces black colonies, whereas other Staphylococci cannot always do so. The medium was found to be less inhibitory to *Staphylococcus aureus* than

other media at the same time being more selective (13, 5, 1). Subsequently the use of Baird-Parker Agar was officially adopted by AOAC International (7) and is recommended in the USP for use in the performance of Microbial Limit Tests (14). Recently, ISO committee has also recommended this medium for the isolation and enumeration of Staphylococci (8).

The identity of--*Staphylococcus aureus* isolated on Baird-Parker Agar must be confirmed with a coagulase reaction. Baird-Parker Agar can also be used to detect coagulase activity by adding fibrinogen plasma (6). Fibrinogen Plasma Trypsin Inhibitor supplement (FD195) dissolved in 10 ml sterile distilled water added to 90 ml sterile molten media kept at 45-50°C. On this medium coagulase positive colonies appear white to grey-black surrounded by an opaque zone due to coagulase activity within 24-48 hours incubation at 35°C. Reduction in tellurite is necessary because of absence of egg yolk emulsion. This results in translucent agar and white to grey coloured colonies of Staphylococci. For quantitative results select 20-200 colonies. Count *Staphylococcus aureus* like colonies and test them for coagulase reaction. Report *Staphylococcus aureus* per gram of food. Smith and Baird-Parker (11) found that the addition of 50 mg/l Sulphamethazine in the medium, suppresses the growth and swarming of *Proteus* species.

Tryptone, HM peptone B and yeast extract are sources of nitrogen, carbon, sulphur and vitamins. Sodium pyruvate not only protects injured cells and helps recovery but also stimulates *Staphylococcus aureus* growth without destroying selectivity. Lithium chloride and potassium tellurite inhibit most of the contaminating microflora except *Staphylococcus aureus*. The tellurite additive is toxic to egg yolk-clearing strains other than *S.aureus* and imparts a black color to the colonies. Glycine,

pyruvate enhances growth of *Staphylococcus*. With the addition of egg yolk, the medium becomes yellow, opaque. The egg yolk additive, in addition to provide enrichment, aids in the identification process by demonstrating lecithinase activity (egg yolk reaction). A clear zone and grey-black colonies on this medium are diagnostic for coagulase positive -Staphylococci. Upon further incubation, an opaque zone is developed around colonies, which can be due to lipolytic activity. When testing the medium, inoculate the material to be examined (0.1 ml per plate of diameter 90-100 mm), incubate at 37°C and take the first reading after 24-26 hours. The colonies of *Staphylococcus aureus* are black and shiny, with a fine white rim, surrounded by a clear zone. Incubate at 37°C for another 24 hours and perform the coagulase test on the colonies with the above characteristics, which have developed during the further incubation period.

The basal medium, without the egg yolk or the tellurite, is perfectly stable.

Type of specimen

Food and dairy samples

Specimen Collection and Handling:

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (2,12,15). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Plates should be used on the same day of preparation or within 48 hours, to avoid the loss of definition in the precipitated zones.
2. Colonies of some contaminating organisms may digest the coagulase halo reaction.
3. Other bacteria may grow on this media but biochemical test will differentiate coagulase positive Staphylococci from the other organisms.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 2.0% agar gel.

Colour and Clarity of prepared medium

Basal medium : Light amber coloured clear to slightly opalescent gel. After addition of Egg Yolk Tellurite Emulsion (FD046) : Yellow coloured opaque gel forms in Petri plates.

Reaction

Reaction of 6.3% w/v aqueous solution at 25°C. pH : 7.0±0.2

pH

6.80-7.20

Cultural Response

M1140: Cultural characteristics observed with added Egg Yolk Tellurite Emulsion (FD046) after an incubation at 35-37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Lecithinase activity	Colour of colony
Cultural Response <i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (00003*)	50-100	none-poor	<=10%	negative	

<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	none-poor	<=10%	negative	large brown black
<i>Micrococcus luteus</i> ATCC 10240	50-100	fair-good	30-40%	negative	very small, brown-black
<i>Proteus mirabilis</i> ATCC 25933	50-100	none-poor	<=10%	negative	brown-black w/ o swarming
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	good-luxuriant	>=50%	positive,halo or clear zone around the colony	grey-black shiny
<i>Staphylococcus epidermidis</i> ATCC 12228 (00036*)	50-100	fair-good	30-40%	negative	black

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (9,10).

Reference

1. Assoc. off. Anal. Chem., 1971, 54:401.
2. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
3. Baird-Parker A. C., 1962, J. Appl. Bacteriol., 25:12.
4. Baird-Parker A. C. and Davenport E., 1965, J. Appl. Bacteriol., 28:390.
5. Baer, 1971, J. Assoc. Off. Anal. Chem., 54:732.
6. Beckers N. J. et al, 1984, Can. J. Microbiol., 30:470.
7. Horwitz (Ed.), 2000, Official methods of analysis of AOAC International, 17th Ed., Vol. I., AOAC International, Gaithersburg, Md.
8. International Organization for Standardization (ISO), 1983, Draft ISO/DIS 6888.
9. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
10. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
11. Smith B. A. and Baird-Parker A.C., 1964, J. Appl. Bacteriol., 27:78.
12. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 215, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
13. Tardio and Baer, 1971, J. Assoc. Off. Anal. Chem., 54:728.
14. The United States Pharmacopoeia, 2008, USP31, The United States Pharmacopoeial Convention. Rockville, MD.
15. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
16. Zebovitz E., Evans J. B. and Niven C.F., 1955, J. Bacteriol., 70:686 .

Revision : 2 / 2018

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.