



# Technical Data

## Yeast Nitrogen Base w/o Amino Acids and Ammonium Sulphate M151

### Intended use

Yeast Nitrogen Base without Amino Acids and Ammonium Sulphate is recommended for use in the classification of yeasts on the basis of their ability to assimilate nitrogen and carbon compounds.

### Composition\*\*

Ingredients	Gms / Litre
Biotin	0.000002
Calcium pantothenate	0.0004
Folic acid	0.000002
Inositol	0.002
Niacin	0.0004
p-Amino benzoic acid (PABA)	0.0002
Pyridoxine hydrochloride	0.0004
Riboflavin (Vitamin B2)	0.0002
Thiamine hydrochloride	0.0004
Boric acid	0.0005
Copper sulphate	0.00004
Potassium iodide	0.0001
Ferric chloride	0.0002
Manganese sulphate	0.0004
Sodium molybdate	0.0002
Zinc sulphate	0.0004
Monopotassium phosphate	1.000
Magnesium sulphate	0.500
Sodium chloride	0.100
Calcium chloride	0.100
Final pH ( at 25°C)	4.5±02

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

A. For Carbon Assimilation tests, prepare the broth base in 10X concentration. Dissolve 1.7 grams in 100 ml distilled water. Add 5 grams ammonium sulphate, 10 mg L-histidine, 20 mg DL-methionine and 20 mg DL-tryptophan. Carbon compounds for assimilation test are added in 10X concentration singly or in combination as required.

B. For Nitrogen Assimilation tests, prepare the medium in 10X concentration. Dissolve 1.7 grams in 100 ml distilled water. Add 1 gram dextrose, 1 mg L-histidine, 2 mg DL-methionine and 2 mg DL-tryptophan. Add nitrogen compounds for assimilation test in 10X concentration singly or in combination as required. Wickerham employed the following nitrogen sources : ammonium sulphate 1gm, potassium nitrate 0.78 gm, urea 0.46 gm, asparagine 1 gm, peptone (gelatin) 1.32 gms.

For A and B, filter sterilize the 10X strength solution. Refrigerate and use as needed. Prepare final medium by aseptically pipetting 0.5 ml of the 10X sterile medium into 4.5 ml sterile distilled water. Mix well.

### Principle And Interpretation

Yeast Nitrogen Base without Amino Acids and Ammonium Sulphate is used for classifying yeasts based on carbohydrate and amino acids requirements. This medium lacks the amino acids, histidine, methionine and tryptophan and also ammonium sulphate. Yeast Nitrogen Base is prepared as per the formulations of Guenter (1), which in turn is modification of Wickerham's formulation (2). Yeast Nitrogen Base without Amino Acids and Ammonium Sulphate contains essential nutrients and vitamins necessary for cultivation of yeasts, except amino acids and a source of nitrogen and carbohydrates.

Wickerham used the following nitrogen sources - ammonium sulphate 1.0 gm/l, potassium nitrate 0.78 gm/l, urea 0.46 gm/l, asparagine 1.0 gm/l, peptone (gelatin) 1.32 gm/l. Yeasts grown on rich medium may carry a reserve of nitrogen in the form of proteins that may result in erroneous findings. To avoid this, 2 serial transfers in complete medium are recommended. After sufficient incubation, measure the growth turbidimetrically at 660 nm using spectrophotometer and compare with control.

Please refer disclaimer Overleaf.

## Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets

## Limitations :

Further biochemical tests must be carried out for confirmation

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

White to cream homogeneous free flowing powder

### Colour and Clarity of prepared medium

Colourless (at 10X concentration colour of medium is pale yellow) clear solution without any precipitate.

### Reaction

Reaction of 0.17% w/v aqueous solution at 25°C. pH : 4.5±0.2

### Cultural Response

M151: Cultural characteristics observed after an incubation at 35-37°C for 6-7 days.

Organism	Growth (Plain)	Growth w/ additions
----------	----------------	---------------------

### Cultural Response

<i>Kloeckera apiculata</i> ATCC 9774	none-poor	good
--------------------------------------	-----------	------

<i>Saccharomyces cerevisiae</i> ATCC 9763 (00053*)	none-poor	good
--	-----------	------

<i>Saccharomyces uvarum</i> ATCC 28098	none-poor	good
--	-----------	------

Key \* : Corresponding WDCM Numbers

## Storage and Shelf Life

Store dehydrated medium in a tightly closed container and use freshly prepared medium. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (3,4).

## Reference

1. Guenter, Personal communication.
2. Wickerham L. J., 1951, U.S. Dept. Agric. Tech. Bull No. 1029.
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2<sup>nd</sup> Edition.
4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 02/2018

**Disclaimer :**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.