



Clausen Medium

M552

Intended Use:

Recommended for sterility testing as per Nordic Pharmacopoeia Board.

Composition**

Ingredients	Gms / Litre
Tryptone	15.000
Soya peptone	3.000
Yeast extract	6.000
Dextrose (Glucose)	6.000
Sodium chloride	2.500
Dipotassium hydrogen phosphate	2.000
Sodium citrate	1.000
L-Cystine	0.500
L-Asparagine	1.250
Sodium dithionite	0.400
Sodium thioglycollate	0.500
Lecithin	0.300
Magnesium sulphate	0.400
Calcium chloride	0.004
Cobalt sulphate	0.001
Cupric sulphate	0.001
Ferrous sulphate	0.001
Zinc sulphate	0.001
Manganese chloride	0.002
Resazurin	0.001
Agar	0.750
Final pH (at 25°C)	7.1±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 40 grams in 1000 ml purified / distilled water containing 3 grams polysorbate 80 and 5 grams glycerol. Heat to boiling to dissolve the medium completely. Dispense as desired and sterilize by autoclaving at 118°C for 15 minutes. Place in cool dark place till use. DO NOT RESTERILIZE the medium

Note: If more than upper one-third of the medium has acquired a pink colour, the medium may be restored once by heating in a water bath or in free flowing steam until the pink colour disappears.

Principle And Interpretation

Clausen Medium was developed by Clausen (1). This medium is also called as HS-T (Dithionite Thioglycollate) Medium and is recommended for sterility testing by the Nordic Pharmacopoeia Board. Random sample selection is recognized by the Board and they refer to the process as microbial-contamination test. The Standard microbial contamination test is developed to establish the number of non-sterile units, if any in batch, is below a specific level. Random sampling in sufficient quantity of the bulk should be examined.

In the microbial contamination test for detecting the non-sterile units, two methods can be used viz. Membrane filter method and Dilution method. The test must be performed with all precautions taken to prevent laboratory contamination.

This medium is very nutritious consisting of casein enzymic hydrolysate, papaic digest of soyabean meal, yeast extract and dextrose. L-cystine and sodium thioglycollate act as reducing agents, and the essential metals help for isolating anaerobic spore-formers. Polysorbate 80 and lecithin are added in this medium to overcome the effects of cationic agents, which can exert bacteriostatic effect in vitro. This medium is clear in appearance and yellow coloured. Under aerobic conditions it turns pink. Therefore at the time of use the upper one third of the medium should be pink.

The standard microbial contamination test is passed if growth is not observed in any of the tubes. Growth is examined by the appearance of turbidity in fluid or semi fluid media and by the formation of colonies on solid media.

Type of specimen

Pharmaceutical samples for sterility testing.

Specimen Collection and Handling:

For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per guidelines (2,3,12)
After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

4 G C Y J N C D E G H Q R G P K P E Q P M C K P G T T Q V G I N Q X X G U R T E Q N Q G V E Y K B E G G o n G f a c e protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

u t oufoefe gps uif fy n o u po pg dmf s m rv e ps ufs tpmvcmf n ufs mt

Performance and Evaluation

Performance of t e medium is e pected en used as per t e direction on t e la el it in t p r y period en stored at recommended temperature

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light straw coloured, clear to slightly opalescent solution with upper 10% or less portion pink on standing.

Reaction

Reaction of 4% w/v aqueous solution containing 0.3% w/v polysorbate 80 and 0.5% w/v glycerol pH : 7.1±0.2

pH

6.90-7.30

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-48 hours.

Organism	Inoculum (CFU)	Growth
<i>Bacillus subtilis</i> ATCC 6633	50-100	luxuriant
<i>Candida albicans</i> ATCC 10231	50-100	luxuriant
<i>Clostridium sporogenes</i> ATCC 11437	50-100	luxuriant
<i>Pseudomonas aeruginosa</i> ATCC 27853	50-100	luxuriant
<i>Staphylococcus aureus</i> ATCC 25923	50-100	luxuriant
<i>Staphylococcus epidermidis</i> ATCC 12228	50-100	luxuriant
<i>Streptococcus pyogenes</i> ATCC 19615	50-100	luxuriant

Reference

1. Clausen O.G., 1973, Pharmaceutica Acta Helvetiae, 48:541.

Revision : 1 / 2011

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
