



## Lactobacillus MRS Agar

M641

### Intended use

Lactobacillus MRS Agar is recommended for cultivation of all *Lactobacillus* species.

### Composition\*\*

Ingredients	Gms / Litre
Proteose peptone	10.000
Meat extract B #	10.000
Yeast extract	5.000
Dextrose (Glucose)	20.000
Tween 80 (Polysorbate 80)	1.000
Ammonium citrate	2.000
Sodium acetate	5.000
Magnesium sulphate	0.100
Manganese sulphate	0.050
Dipotassium hydrogen phosphate	2.000
Agar	12.000
Final pH ( at 25°C)	6.5±0.2

\*\*Formula adjusted, standardized to suit performance parameters

# Equivalent to Beef extract

### Directions

Suspend 67.15 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

Lactobacilli MRS medium is based on the formulation of deMan, Rogosa and Sharpe (1) with slight modification. It supports luxuriant growth of all Lactobacilli from oral cavity (1), dairy products (2), foods (3), faeces (4) and other sources (5).

Proteose peptone and meat extract B supply nitrogenous and carbonaceous compounds. Yeast extract provides vitamin B complex and dextrose is the fermentable carbohydrate and energy source. Polysorbate 80 supplies fatty acids required for the metabolism of Lactobacilli. Sodium acetate and ammonium citrate inhibit Streptococci, moulds and many other microorganisms. Magnesium sulphate and manganese sulphate provide essential ions for multiplication of lactobacilli. Phosphates provide good buffering action in the media.

Lactobacilli are microaerophilic and generally require layer plates for aerobic cultivation on solid media. When the medium is set, another layer of un-inoculated MRS Agar is poured over the surface to produce a layer plate (5). Lactobacilli isolated on MRS Agar should be further confirmed biochemically.

### Type of specimen

Clinical samples - faeces ; Food and dairy samples

### Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (6,7).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (8,9,10).

After use, contaminated materials must be sterilized by autoclaving before discarding.

## Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets

## Limitations :

Biochemical identification is required for further isolation upto species level.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Cream to light yellow homogeneous free flowing powder

### Gelling

Firm, comparable with 1.2% Agar gel.

### Colour and Clarity of prepared medium

Medium to dark amber coloured, clear to slightly opalescent gel forms in Petri plates

### Reaction

Reaction of 6.71% w/v aqueous solution at 25°C. pH : 6.5±0.2

### pH

6.30-6.70

### Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours or longer.(with 5% CO<sub>2</sub>)

Organism	Inoculum (CFU)	Growth	Recovery
<i>Lactobacillus casei</i> ATCC 9595	50-100	luxuriant	≥50%
<i>Lactobacillus fermentum</i> ATCC 9338	50-100	luxuriant	≥50%
<i>Lactobacillus leichmannii</i> ATCC 7830	50-100	luxuriant	≥50%
<i>Lactobacillus plantarum</i> ATCC 8014	50-100	luxuriant	≥50%
<i>Lactobacillus saki</i> ATCC 15521(00015*)	50-100	luxuriant	≥70%
<i>Lactobacillus lactis</i> ATCC 19435(00016*)	50-100	luxuriant	≥70%
<i>Pediococcus pentosaceus</i> ATCC 33316(00158*)	50-100	luxuriant	≥70%
<i>Escherichia coli</i> ATCC 25922(00013*)	≥10 <sup>3</sup>	Inhibition	0%
<i>Bacillus cereus</i> ATCC 11778(00001*)	≥10 <sup>3</sup>	Inhibition	0%

Key: \* Corresponding WDCM numbers.

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (11, 12).

## Reference

- 1.deMan J., Rogosa M. and Sharpe M., 1960, J. Appl. Bacteriol., 23:130.
- 2.Marshall R.T. (Ed.), 1992, Standard Methods for the Examination of Dairy Products, 16th ed., APHA, Washington,D.C.
- 3.Downes F. P. and Ito K., (Eds.), 2001, Compendium of Methods For the Microbiological Examination of Foods, 4th Ed., APHA, Washington, D.C.
- 4.Sabine and Vaselekos, 1965, Nature, 206:960.
- 5.MacFaddin J.,1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol.1, Williams and Wilkins, Baltimore.
- 6.Isenberg, H.D. Clinical Microbiology Procedures Handb0ook. 2<sup>nd</sup> Edition.
7. Jorgensen,J.H., Pfaller , M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 03 / 2017



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
B /4-6 , MIDC, Dindori, Nashik MH

[www.himedialabs.com](http://www.himedialabs.com)



CE Partner 4U ,Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory,diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.