



## Antibiotic Assay Medium C

ME555

Antibiotic Assay Medium C is used as the broth medium in turbidimetric assay of a wide variety of antibiotics in accordance with European Pharmacopoeia .

### Composition\*\*

Ingredients	Gms / Litre
Peptone	6.000
Beef extract	1.500
Yeast extract	3.000
Sodium chloride	3.500
Glucose monohydrate	1.000
Dipotassium hydrogen phosphate	3.680
Potassium dihydrogen phosphate	1.320
pH after sterilization	*7.0±0.1

\*\*Formula adjusted, standardized to suit performance parameters

\* While assaying Josamycin & Josamycin sulphate adjust the pH to 8.0 ±0.1

### Directions

Suspend 19.9 grams of dehydrated media in 1000 ml R-water/ purified /distilled water. Heat if necessary to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Adjust the pH of the medium, using freshly prepared buffer solution as recommended by the European pharmacopoeia for the antibiotic assayed.

*Advice : Recommended for the microbiological assay of Colistimethate sodium, Dihydrostreptomycin sulphate, Erythromycin estolate, Framycetin sulphate, Gentamicin sulphate, Gramicidin, Kanamycin acid sulphate ,Kanamycin monosulphate, Neomycin sulphate, Rifamycin sodium ,Spiramycin, Streptomycin sulphate, Tylosin, Tylosin tartarate, Tyrothricin and Vancomycin hydrochloride .*

### Principle And Interpretation

This medium is used in turbidimetric assay of several antibiotics. The composition of the medium is in accordance to the specifications detailed in the European Pharmacopoeia (1). Turbidimetric methods for determining the potency of antibiotics are inherently more accurate and more precise than comparable agar diffusion procedures (2) .

Peptone, beef extract and yeast extract provide essential nutrients and growth factors for enhanced microbial growth. Sodium chloride maintains the osmotic equilibrium while phosphates are incorporated in the medium to provide good buffering action. Glucose monohydrate serves as the carbon and energy source for faster growth.

Turbidimetric antibiotic assay is based on the change or inhibition of growth of a test microorganisms in a liquid medium containing a uniform concentration of an antibiotic (3). Use of this method is appropriate only when test samples are clear .

### Quality Control

#### Appearance

Cream to yellow homogeneous free flowing powder

#### Colour and Clarity of prepared medium

Light yellow coloured clear solution without any precipitate

#### pH

6.90-7.10

#### Cultural Response

ME555: Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours .

Organism	Inoculum (CFU)	Growth	Serial dilution with
<i>Escherichia coli</i> ATCC 9637	50-100	luxuriant	Colistimethate sodium
<i>Escherichia coli</i> ATCC 10536	50-100	luxuriant	Rifamycin sodium
<i>Enterococcus hirae</i> ATCC 10541	50-100	luxuriant	Gramicidin, Tyrothricin
<i>Klebsiella pneumoniae</i> ATCC 10031	50-100	luxuriant	Streptomycin sulphate
<i>Staphylococcus aureus</i> ATCC 6538p	50-100	luxuriant	Framycetin sulphate, Genatamicin sulphate, Gramicidin, Kanamycin monosulphate, Kanamycin acid sulphate, Neomycin sulphate, Spiramycin; For Josamycin & Josamycin propionate-adjust the pH of the medium to $8.0 \pm 0.1$ , For Vancomycin hydrochloride incubate at 35-37°C.
<i>Staphylococcus aureus</i> ATCC 9144	50-100	luxuriant	Tylosin, Tylosin tartarate

## Storage and Shelf Life

Store below 30°C in tightly closed container and use freshly prepared medium . Use before expiry date on the label.

## Reference

1. European Pharmacopoeia, 2011, European Department, for the Quality of Medicines.
2. Ripper RA. Some principles of microbiological turbidimetric assays of antibiotics. J Assoc Off Anal Chem.1979 62(4):951-6.
3. Chapin-Robertson and Edberg, 1991, Measurement of Antibiotics in Human Body fluids: Techniques and significance . Antibiotics in Laboratory medicine, New York pp 305

Revision : 2 / 2015

## Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.