



Folin & Wu's phosphate,Molybdate

R007

It is used for determination of blood sugar by Folin Wu's method.

Composition**

Ingredients

Molybdic acid	35.0 gm
Sodium tungstate	5.0 gm
10% NaOH solution	200.0 ml
Phosphoric acid	125.0 ml
Distilled water	500.0 ml
Final pH (at 25°C)	0.9±0.5

**Formula adjusted, standardized to suit performance parameters

Directions

a) Deproteinisation :

- 1)Take 1 ml blood in a test tube.
- 2)Add 7 ml distilled water and 1 ml of 0.6 N H₂SO₄ and mix well. Allow it to react for 5-10 minutes.
- 3)Precipitate the proteins by adding 1 ml of 10% Sodium tungstate. Mix thoroughly and keep at room temperature for 10 minutes.
- 4)Centrifuge or filter through Whatman No. 1 filter paper.

b) Reduction and Colour Development:

- 1)Pipette out 2 ml of the above filtrate into a Folin-Wu tube labelled "Test".
- 2)Add 2 ml of standard glucose (100% in saturated benzoic acid) in to another Folin-Wu tube named "S".
- 3)Add 2.0ml Folin and WU's Alkaline Copper solution (R006) in each tube. Mix and keep in boiling water bath for 8 minutes.
- 4)Cool under running tap water.
- 5)Add 2.0 ml of Folin and WU's phosphate molybdate solution (R007) and dilute with distilled water to 25 ml mark. Mix well and keep for 5 minutes.
- 6)Read the colour of standards and test at 420 nm using distilled water as blank.

Blood glucose : (mg/dl) = OD of T/OD of S x 100.

Principle And Interpretation

Glucose and other reducing substances under alkaline conditions reduce the cupric ions of the copper reagent to cuprous ions. Cuprous oxides reacts with phosphomolybdic acid to form molybdenum blue. Intensity of molybdenum blue is directly proportional to the reducing sugars present. It is measured at 420 nm and compared with a known standard.

Quality Control

Appearance

Colourless solution

Clarity

Clear with no insoluble particles.

Reaction

Reaction of the solution at 25°C

pH

0.40-1.40

Test

Estimation of blood sugar is carried out by Folin and WU's method.

Results

The normal range for

- 1) Blood sugar fasting : 80-120 mg/dl
- 2) Blood sugar P.P : up to 140mg/dl
(2 hours after lunch)

Storage and Shelf Life

Store below 30°C in tightly closed container and away from bright light. Use before expiry date on label.

Reference

- 1.) Text book of Medical Laboratory Technology; Praful B. Godkar

Revision : 01 / 2015

**Disclaimer :**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.