

Co-Trimoxazole Ezy MIC™ Strip (TSH) (0.016-256 mcg/ml) EM083I **(Trimethoprim/ Sulfamethoxazole) (1:19)** **(Individually Packed)**

Antimicrobial Susceptibility Testing
For *In Vitro* Diagnostic use

It is a unique MIC determination paper strip which is coated with Co-Trimoxazole in a concentration gradient manner, capable of showing MICs in the range of 0.016mcg/ml to 256 mcg/ml, on testing against the test organism.

Introduction

Ezy MIC™ strip is useful for quantitative determination of susceptibility of bacteria to antibacterial agents. The system comprises of a predefined quantitative gradient which is used to determine the Minimum Inhibitory Concentration (MIC) in mcg/ml of different antimicrobial agents against microorganisms as tested on appropriate agar media, following overnight incubation.

The overuse and misuse of known effective drugs like Co-Trimoxazole has led to emergence of resistant strains among cultures that were known to be sensitive to this antibiotic. Such resistant strains can be detected using the high concentration of Co-Trimoxazole Ezy MIC™ Strip.

Ezy MIC™ Strip FEATURES AND ADVANTAGES

Ezy MIC™ strip exhibits several advantages over existing plastic strip.

1. Ezy MIC™ strip is made up of porous paper material unlike plastic non-porous material
2. Ezy MIC™ strip has MIC values printed on both sides identically.
3. The antimicrobial agent is evenly distributed on either side of the Ezy MIC™ strip and hence it can be placed by any side on the agar surface.
4. For Ezy MIC™ strips, MIC values can be read without opening the lid of the plate as most commonly translucent medium such as Mueller Hinton Agar is employed.
5. Once placed, Ezy MIC™ strip is adsorbed within 60 seconds and firmly adheres to the agar surface.
6. Unlike the plastic material, it does not form air bubbles underneath and hence there is no need to press the strip once placed.

METHOD AND USE OF EZY MIC™ STRIPS

• Type of specimen

Pure cultures should be derived from specimens obtained from patients prior to the initiation of antimicrobial therapy. Specimens can be of bacterial or fungal isolates derived from blood, urine, faeces, pus, CSF etc. Direct specimens should not be employed in this test. Refer procedure, which includes preparation of inoculum (1, 3).

• Clinical specimen collection, handling and processing

Follow appropriate techniques for handling specimens as per established guidelines. After use, contaminated materials must be sterilized by autoclaving before discarding (1, 3).

• Guidelines for preparation of the medium

Prepare the medium of choice from dehydrated powder according to the directions specified on the label. Cool the sterilized molten medium to 45-50°C and pour in sterile, dry Petri plates on a leveled surface, to a depth of 4± 0.2mm and allow to solidify. Few droplets appearing on the surface of the medium following cooling do not matter. Hence, once poured, Petri plates containing media should not be dried on laminar flow and can be used immediately for swabbing.

• Preparation of Inoculum

Use only pure cultures. Confirm by Gram-staining before starting susceptibility test. Transfer 4-5 similar colonies with a wire, needle or loop to 5 ml Tryptone Soya Broth (M011) and incubate at 35-37°C for 2-8 hours until light to moderate turbidity develops. Compare the inoculum turbidity with that of standard 0.5 McFarland. Alternatively, the inoculum can be standardized by other appropriate optical method (0.08 - 0.13 OD turbid suspension at 620 nm).

Also, direct colony suspension method can be used. Prepare a direct colony suspension, from 18-24-hour old non-selective media agar plate in broth or saline. Adjust the turbidity to that of standard 0.5 McFarland. This method is recommended for testing fastidious organisms like *Haemophilus* spp., *Neisseria* spp, streptococci and for testing staphylococci for potential Methicillin or Oxacillin resistance.

- **Test Procedure**

1. Prepare plates with suitable make of recommended medium as mentioned in Table 2 & Table 4.
2. Dip a sterile non-toxic cotton swab on a wooden applicator into the standardized inoculum and rotate the soaked swab firmly against the upper inside wall of the tube to express excess fluid. Swab the entire agar surface of the plate with the swab three times, turning the plate at 60° angle between each streaking.
3. Remove individually packed Ezy MIC™ strip from cold and keep it at room temperature for 15 minutes before opening.
4. Tear the aluminum pouch carefully at the notch provided, so that it does not damage the Ezy MIC™ Strip.
5. With the help of sterile forceps, remove the Ezy MIC™ Strip gently and place the strip at a desired position on an agar plate swabbed with the test culture.
6. DO NOT PRESS EZY MIC™ STRIP. Within 60 seconds, Ezy MIC™ strip will be adsorbed and will firmly adhere to the agar surface.
7. Ezy MIC™ strip should not be repositioned or adjusted once placed.
8. Transfer plates in the incubator under appropriate conditions.

MIC Reading:

1. Read the plates only when sufficient growth is seen.
2. Read the MIC where the ellipse intersects the MIC scale on the strip.
3. For bacteriostatic drugs such Trimethoprim/ sulfamethoxazole (Co-Trimoxazole), Tetracycline, Azithromycin, Fluconazole, Linezolid and, Chloramphenicol read MICs at 80% inhibition for homogenously sensitive strains such as QC control strains.
4. Isolated colonies, micro colonies and hazes appearing in the zone of inhibition are indicative of heteronature of the culture having resistant subpopulation in it. In such cases, consider reading for MIC determination at a point on the scale above which no resistant colonies are observed close to MIC strip (within 1-3 mm distance from the strip).
5. Since Ezy MIC™ strip has continuous gradient, MIC values “in-between” two-fold dilutions can be obtained.
6. Always round up these values to the next two-fold dilution before categorization. For example: Co-Trimoxazole showing reading of 0.75 mcg/ml should be rounded up to next concentration ie. 1.0 mcg/ml.
7. If the ellipse intersects the strip in between 2 dilutions, read the MIC as the value which is nearest to the intersection.
8. When growth occurs along the entire strip, report the MIC as \geq the highest values on the MIC strip. When the inhibition ellipse is below the strip (does not intersect the strip), report the MIC $<$ the lowest value on the MIC scale.

Warning and Precautions:

1. Ezy MIC™ Strip is intended for *In vitro* diagnostic use only.
2. Although based on simple procedure, Ezy MIC™ Strip should only be used by at least semi-trained personnel.
3. This strip is intended only for agar diffusion method and not for broth dilution method.
4. Ezy MIC™ Strip should be used strictly according to procedures described herein.
5. Performance of Ezy MIC™ Strips depends on use of proper inoculum and control cultures, recommended test medium and proper storage temperature.
6. Follow aseptic techniques and precautions against microbiological hazards should be used when handling bacterial or fungal specimen throughout the testing procedure.
7. Before using Ezy MIC™ Strips, ensure that the strip is at room temperature.
8. When applying strips be steady. Do not move the strip once in contact with agar surface, since the antibiotic instantaneously diffuse on contact with agar.
9. Place the unused strips back to recommended temperature.

INTERPRETATION & QUALITY CONTROL (As per CLSI Guidelines):**Interpretation:**

Table 1: Use following interpretive criteria for susceptibility categorization.

When testing	Incubation	Interpretative Criteria		
		≤ S	I	≥ R
<i>Enterobacterales</i> (Excluding <i>Salmonella/Shigella</i> spp.), <i>Acinetobacter</i> spp., <i>B. cepacia</i> , <i>S. maltophilia</i> , Other Non- <i>Enterobacterales</i> , <i>Staphylococcus</i> spp., <i>Salmonella</i> and <i>Shigella</i> spp.	35-37°C for 18 hrs.	2	-	4
<i>Haemophilus</i> spp., <i>S. pneumoniae</i>	35-37°C for 20-24 hrs with 5% CO ₂	0.5	1-2	4
<i>N. meningitidis</i>	35-37°C for 20-24 hrs with 5% CO ₂	0.12	0.25	0.5

Quality control

Quality control of Ezy MIC™ Strip is carried out by testing the strips with standard ATCC Cultures recommended by CLSI on suitable medium incubated appropriately.

Table 2: Following are the reference MIC values (mcg/ml) range for Co-Trimoxazole.

Organism	Medium used	Incubation	Std. Quality Control limits (mcg/ml)
<i>E.coli</i> ATCC 25922	Mueller Hinton Agar	35-37°C for 18 hrs.	≤ 0.5
<i>S.aureus</i> ATCC 29213	Mueller Hinton Agar	35-37°C for 18 hrs.	≤ 0.5
<i>P.aeruginosa</i> ATCC 27853	Mueller Hinton Agar	35-37°C for 18 hrs.	8-32
<i>E.faecalis</i> ATCC 29212	Mueller Hinton Agar	35-37°C for 18 hrs.	≤ 0.5
<i>H. influenzae</i> ATCC 49247	Haemophilus Test Medium	35-37°C for 20-24 hrs with 5% CO ₂	0.032 - 0.064 - 0.12 - 0.25
<i>S. pneumoniae</i> ATCC 49619	Mueller Hinton Agar w/ 5% Sheep Blood	35-37°C for 20-24 hrs with 5% CO ₂	0.125 - 0.25 - 0.5 - 1.0

INTERPRETATION & QUALITY CONTROL (As per EUCAST Guidelines):**Interpretation:**

Table 3: Use following interpretive criteria for susceptibility categorization for Co-Trimoxazole as per EUCAST.

When testing	Incubation	Interpretive Criteria	
		< S	≥ R
<i>Enterobacterales</i> , <i>Acinetobacter</i> spp., <i>Staphylococcus</i> spp., <i>Aeromonas</i> spp.	35°C ± 1°C for 18 ± 2hr	2	4
<i>Stenotrophomonas maltophilia</i> , <i>Burkholderia pseudomallei</i>	35°C ± 1°C for 18 ± 2hr	0.001	4
<i>Moraxella catarrhalis</i>	35°C ± 1°C for 18 ± 2hr	0.5	1
<i>Listeria monocytogenes</i>	35°C ± 1°C for 18 ± 2hr	0.06	0.06
<i>Pasteurella multocida</i> , <i>Kingella kingae</i>	35°C ± 1°C for 18 ± 2hr	0.25	0.25

<i>Achromobacter xylosoxidans</i> , <i>Brucella melitensis</i>	35°C ± 1°C for 18 ± 2hr	0.125	0.125
<i>Corynebacterium diphtheriae</i> and <i>C. ulcerans</i>	35°C ± 1°C for 18 ± 2hr with 5% CO ₂	0.5	0.5
<i>Vibrio</i> spp.	35°C ± 1°C for 18 ± 2hr	0.25	0.25
<i>Streptococcus</i> groups A, B, C and G, <i>Streptococcus pneumoniae</i>	35°C ± 1°C for 18 ± 2hr with 5% CO ₂	1	2
<i>Haemophilus</i> spp.	35°C ± 1°C for 18 ± 2hr with 5% CO ₂	0.5	1

Quality Control:

Quality control of Ezy MIC™ Strip is carried out by testing the strips with standard ATCC Cultures recommended by EUCAST on suitable medium incubated appropriately.

Table 4: Following are the reference MIC values (mcg/ml) range for Co-Trimoxazole.

Organism	Medium used	Incubation	Target MIC Value (mcg/ml)	Std. Quality Control limits (mcg/ml)
<i>E. coli</i> ATCC 25922	Mueller Hinton Agar	35°C ± 1°C for 18 ± 2hr	≤0.5	-
<i>S. aureus</i> ATCC 29213	Mueller Hinton Agar	35°C ± 1°C for 18 ± 2hr	≤0.5	-
<i>E. faecalis</i> ATCC 29212	Mueller Hinton Agar	35°C ± 1°C for 18 ± 2hr	≤0.5	-
<i>S. pneumoniae</i> ATCC 49619	Mueller Hinton Agar w/ 5% defibrinated horse blood & 20 mg/L β-NAD	35°C ± 1°C for 18 ± 2hr with 5% CO ₂	0.25 – 0.5	0.125 – 0.25 – 0.5 – 1.0
<i>H. influenzae</i> ATCC 49766	Mueller Hinton Agar w/ 5% defibrinated horse blood & 20 mg/L β-NAD	35°C ± 1°C for 18 ± 2hr with 5% CO ₂	0.032	0.016 – 0.032 – 0.064

Storage & Shelf Life:

1. Once the consignment is received, store applicators at room temperature and Ezy MIC™ Strip container at 2-8°C, for prolonged use store below -20°C or below.
2. Use before expiry date on the label.
3. Ezy MIC™ Strip left over from opened package must be kept dry.
4. Moisture should be prevented from penetrating into or forming within the package or storage container.
5. Check whether the batch number and expiry date are marked on the storage container.
6. Product performance is best within stated expiry period if correctly stored and handled.

Disposal:

After use, Ezy MIC™ Strips and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (2, 3).

Limitation of Test

Ezy MIC™ Strips provides *In vitro* MIC values, which provides only a possible insinuation of pathogens potential in *In vivo* susceptibility. These values can be considered as a guide to therapy selection only after taking into consideration several other factors; and must be the sole decision and responsibility of the physician along with the clinical experience in treating the infection. These tests are comparable to the standards as per the given specifications and set of experiment standards as far as possible. Please refer to CLSI standards for detailed limitation of susceptibility test on the clinical use of an antibiotic in various therapeutic conditions.

References:

1. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition, Vol. 1, Section 2.
2. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition, Vol. 3, Section 15.
3. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
4. Performance Standards of Antimicrobial Susceptibility Testing; 34th Edition. M100-Ed34, Vol.44, No.5, Jan-2024.
5. European Committee on Antimicrobial Susceptibility Testing Breakpoint tables for interpretation of MICs and zone diameters Version 14.0, valid from 2024-01-01.
6. European Committee on Antimicrobial Susceptibility Testing Routine and extended internal quality control as recommended by EUCAST Version 14.0, valid from 2024-01-01.

Packing:

Each Pack contains following material packed in a desiccator with desiccator.

- 1) Co-Trimoxazole Ezy MIC™ strips (Individually packed) (10/30/60/90/120/150 Strips per pack)
- 2) Package insert

Revision: 02/2024



On receipt store at -20°C



In vitro diagnostic medical device



Plot No. C-40,
Road No. 21Y, MIDC,
Wagle Industrial Estate,
Thane (W) - 400604,
Maharashtra, India



Indicates a single sterile barrier system



Do not re-use



CEpartner4U,
ESDOORNLAAN 13,
3951DB MAARN, NL
www.cepartner4u.eu



Do not use if package is damaged

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.