



Middlebrook OADC Enrichment Supplement

FD329

An enrichment supplement recommended for cultivation of *Mycobacteria*.

Composition

Per vial, sufficient for 180 ml medium

*Ingredients	Concentration
Bovine albumin fraction V	1.00 gm
Dextrose	0.40 gm
Catalase	0.0006 gm
Oleic acid	0.012 gm
Sodium chloride	0.17gm
Distilled water	20.00 ml

Directions:

Warm up the refrigerated supplement to room temperature. Shake well to form uniform suspension. Aseptically add 1 vial to 180 ml of sterile, molten, cooled (45-50 C) Middlebrook 7H9 Agar Base (M197) / Middlebrook 7H10 Agar Base (M199) / Middlebrook 7H10 Agar Base, Special (M196) / Middlebrook 7H11 Agar Base (M511) / Middlebrook 7H11 HiVeg™ Agar Base (MV511) / Middlebrook 7H11 Agar Base w/o Malachite green (M511A). Mix well and pour into sterile screw capped tubes or containers.

Note : Shake well before use. "Presence of Bovine albumin Fraction V, in the supplement may give it an opaque and sometimes cloudy appearance. It does not interfere with the biological efficacy of the product".

Type of specimen

Clinical samples - Sputum

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (1,2).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning & Precautions

In Vitro diagnostic use only. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Storage and Shelf Life

Store at 2 - 8°C. Use before expiry date on the label.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,2).

Reference

1. Isenberg (Ed.), 2004, Clinical Microbiology Procedures Handbook, Vol.3, American Society for Microbiology, Washington. D.C.
2. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

* Not For Medicinal Use

Revision : 00/2023



HiMedia Laboratories Pvt.
Limited, Plot No.C-40, Road
No.21Y, MIDC, Wagle Industrial
Area, Thane (W) -400604, MS,
India




CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



In vitro diagnostic
medical device



CE Marking



Storage temperature



Do not use if
package is damaged

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.