



HiTouch™ Flexi Plate - VJ

FL012

For enumeration (count) of *S. aureus* .

Composition**

Ingredients	Gms / Litre
Peptone special	15.000
Mannitol	10.000
Dipotassium phosphate	5.000
Lithium chloride	5.000
Glycine	10.000
Indicator dye	0.025
Agar	16.000
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Open the lid and carefully lift up the enclosed prepared medium plate so as to avoid touching the agar surface by hand. Touch the surface of agar plate onto the surface to be tested. Gently press the plate manually for upto 10 seconds. Apply constant and uniform pressure to the whole surface (ensuring that an even pressure of 25 gm/cm² is distributed over the whole plate for 10 seconds). Replace exposed medium plate back in the base plate. Close the lid. Press the sides of the lid to make sure that it is fixed in the grooves. Disinfect the surface where the sample was taken in order to remove any possible traces of agar. Incubate the plates at specified temperature. After incubation as recommended count the number of colonies which have appeared on the surface of medium. Alternative Methods of Inoculation : To use as Culture Plate (ii), Sample Dilution Plate (iii) or Swabbing Plate (iv) To use as Gravitation Settling Plate (v)

Principle And Interpretation

HiTouch Flexi Plates are specially developed for the microbial testing in food , pharmaceutical, cosmetic,dairy, hospitals, water works, environmental testing etc. These plates are handy and ready to use sterile media supplied in flexible disposable plates, 55 mm in diameter. It is grid scored on the base and is irradiated to ensure perfect sterility. Medium is filled aseptically and each plate is packed in pre-sterilized plastic bag. HiTouch Flexi Plate is then packed in plastic pouch wrapping. The unique flexible plate configuration ensures close contact even with uneven surfaces. where not only counts are obtained but it is also possible to select and differentiate between groups of microorganisms like coliforms (both *E. coli* and non *E. coli*). These plates are specially developed for microbial testing, The Flexi plate medium formula is suitable for optimal growth of environmental microorganisms and the grids enable direct reading on the plates of the number of colonies per cm².

These plates are specially developed for microbial testing, The Flexi plate medium formula is suitable for enumeration of *S. aureus* and the grids enable direct reading on the plates of the number of colonies per cm².

Quality Control

Appearance

Sterile plastic plate containing reddish pink coloured firm gel

Quantity of Medium

9ml of gel in plastic plate

Reaction

7.00- 7.40

Sterility test

Passes release criteria

Cultural Response

Cultural characteristics was observed after incubation at 35-37°C for 18-24 hours.

Organism	Growth	Colour of Colony
<i>Escherichia coli</i> ATCC 25922	Inhibited	-
<i>Staphylococcus aureus</i> ATCC25923	Luxuriant	Black with yellow halo
<i>Staphylococcus epidermidis</i> ATCC 12228	Fair	Translucent to blackish

Storage and Shelf Life

Store between 2-8°C. Use before expiry date on the label.

Reference

- 1.Hall L., Hartnett M.J., Pub. (1964) Health Rep., 79,1021.
- 2.Pryor A.K., Mc Duff C.R., Exed Housekeeper. (1969) Mass.
- 3.Lenette E.H., Spauldine E.H., Truant J.P.Manual of Clinical Microbiology, 2nd ed., (1974) American Society for Microbiology.
- 4.Desbordes J., Biodegradation microbienne des antiseptiques et conservateurs. (1977) Rev. Institut Pasteur de Lyon, 10 (n04), 291-311.
- 5.Norme francaise NF T 72-151, Antiseptiques et desinfectants utilises a retat liquide, misibles a reau (03/1981).
- 6."Guide du blonnetoyage". Journal Officiel de la Republique Francaise. Recommandation NOE 1-90 (1991).

Revision : 1 / 2011



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory,diagnostic,research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.