



HiTouch™ SCDA w/ Polysorbate 80 & Lecithin *Flexi* plate FL035G (gamma irradiated)

For determining the efficiency of sanitization of containers, equipment surfaces, water miscible cosmetics, etc. It can also be used to enumerate the organisms from water insoluble products and fatty products containing preservatives or antimicrobials.

Composition**

Ingredients	Gms / Litre
Casein enzymic hydrolystae	15.000
Papaic digest of soyabean meal	5.000
Sodium chloride	5.000
Lecithin	0.700
Polysorbate 80 (Tween 80)	5.000
Agar	15.000

**Formula adjusted, standardized to suit performance parameters

Directions

Open the pouch in the protected area . Remove the wrapping and open the lid and carefully lift up the enclosed prepared medium plate so as to avoid touching the agar surface by hand.Touch the surface of agar plate onto the surface to be tested. Gently press the plate manually for upto 10 sencond. Apply constant and uniform pressure to the whole surface (ensuring that an even pressure of 25 gm/cm2 is distributed over the whole plate for 10 seconds). Replace exposed medium plate back in the base plate. Close the lid. Press the sides of the lid to make sure that it is fixed in the grooves. Disinfect the surface where the sample was taken in order to remove any possible traces of agar. Incubate the plates at specified temperature. After incubation as recommended count the number of colonies which have appeared on the surface of medium. Alternative Methods of Inoculation : To use as Culture Plate (ii), Sample Dilution Plate (iii) or Swabbing Plate (iv) To use as Gravitation Settling Plate (v)

Principle And Interpretation

Hitouch *Flexi* Plates are specially developed for the microbial testing in food , pharmaceutical, cosmetic,dairy, hospitals, water works, environmental testing etc. These plates are handy and ready to use sterile media supplied in flexible disposable plates, 55 mm in diameter. It is grid scored on the base and is irradiated to ensure perfect sterility. Medium is filled aseptically and each plate is packed in pre-sterilized plastic bag. Hitouch *Flexi* Plateis then packed in plastic pouch wrapping. The unique flexible plate configuration ensures close contact even with uneven surfaces. where not only counts are obtained but it is also possible to select and differentiate between groups of microorganisms like coliforms (both *E. coli* and non *E. coli*). These plates are specially developed for microbial testing, The *Flexi* plate medium formula is suitable for enumeration of the organisms from water soluble products, fatty products containing preservatives or antimicrobials and the grids enable direct reading on the plates of the number of colonies per cm2. This plate can be used for determining the effeciency of sanitization of containers, equipment surfaces, water miscible cosmetic, etc.

Casein enzymic hydrolysate and papaic digest of soyabean meal provide nitrogenous compounds and other nutrients essential for microbial replication. Lecithin and polysorbate 80 (Tween 80) are neutralizers reported to inactivate residual disinfectants from where the sample is collected (4). Lecithin neutralizes quaternary ammonium compounds and polysorbate 80 neutralizes phenolic disinfectants, hexachlorophene, formalin and with lecithin ethanol (5).

Collection of samples from areas before and after the treatment with disinfectant evaluates cleaning procedures in environmental sanitation. The presence and number of microorganisms is determined by the appearance of colonies on the agar surface (6). After counting the colonies, carry out biochemical testing for identification.

Quality Control

Appearance

Sterile plastic plate containing light to medium amber coloured firm gel.

Quantity of Medium

9 ml of gel in plastic plate.

Reaction

7.10- 7.50

Sterility test

Passes release criteria

Cultural response

Cultural characteristics observed after incubation at 35-37°C for 18-24 hours.

Organism	Growth
<i>Escherichia coli</i> ATCC 25922	Luxuriant
<i>Staphylococcus aureus</i> ATCC 25923	Luxuriant
<i>Staphylococcus aureus</i> ATCC 6538	Luxuriant
<i>Pseudomonas aeruginosa</i> ATCC 27853	Luxuriant (may have green pigment)

Storage and Shelf Life

Store between 2-8°C. Use before expiry date on the label.

Reference

- Hall L., Hartnett M.J., Pub. (1964) Health Rep., 79,1021.
- Pryor A.K., Mc Duff C.R., Exed Housekeeper. (1969) Mass.
- Lenette E.H., Spauldine E.H., Truant J.P. Manual of Clinical Microbiology, 2nd ed., (1974) American Society for Microbiology.
- Desbordes J., Biodegradation microbienne des antiseptiques et conservateurs. (1977) Rev. Institut Pasteur de Lyon, 10 (n04), 291-311.
- Norme francaise NF T 72-151, Antiseptiques et desinfectants utilises a retat liquide, misibles a reau (03/1981).
- "Guide du blonettoyage". Journal Officiel de la Republique Francaise. Recommandation NOE 1-90 (1991).

Revision : 1 / 2011



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.