

Sakazakii DHL Agar, Granulated®

GM1619

Intended Use:

Recommended for the detection and isolation of pathogenic *Enterobacteriaceae* from all types of specimens.

Composition**

Ingredients	g / L
Tryptone	10.000
HM peptone #	10.000
HM extract \$	3.000
Lactose	10.000
Sucrose	10.000
L-Cysteine hydrochloride monohydrate	0.200
Sodium citrate	1.000
Sodium deoxycholate	1.500
Sodium thiosulphate	2.000
Ammonium iron (III) citrate	1.000
Neutral red	0.030
Agar	15.000
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Meat peptone, \$ Equivalent to Meat extract

Directions

Suspend 63.71 grams (the equivalent weight of dehydrated medium per litre) in 1000 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

Principle And Interpretation

Enterobacteriaceae have wide distribution. Many members form the normal gut and intestinal microflora in humans and animals. They are also found on plants and in soils and water (1). Some species occupy very limited ecological niches. They are a major component of the normal intestinal flora of humans but are relatively uncommon at other body sites. They account for nearly 50% of septicemia cases, more than 70% of urinary tract infections and a significant percentage of intestinal infections (2).

Sakazakii DHL Agar is modified Deoxycholate Agar (M065) as described by Sakazakii et al (3,4). Sakazakii DHL Agar stands for Sakazakii Deoxycholate-Hydrogen sulphide-Lactose Agar. The medium is selective for the identification and isolation of *Enterobacteriaceae* due to inclusions of sodium deoxycholate. Sodium deoxycholate inhibits gram-positive bacteria and also prevents swarming growth of *Proteus* species. However due to the low concentration of sodium deoxycholate coupled with the nutritionally rich media, fastidious strains of *Salmonella* and *Shigella* are able to grow on this medium. Sulphur is released from thiosulphate or other sulphur-containing compounds in the form of sulphide. The H₂S thus produced is detected by ferric ammonium citrate to form insoluble heavy metal sulphides that appear as a black precipitate (5). *Proteus*, *Morganella*, *Retzgerella* and *Providencia* colonies are surrounded by dark brown zones due to phenylalanine deamination. Phenylalanine is sourced from peptone that forms an iron complex with the ferric ions. The high concentration of sucrose in the medium permits the recovery of sucrose positive and lactose negative members of *Enterobacteriaceae*.

Type of specimen

Clinical- faeces; Food samples.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (7,8). For food samples, follow appropriate techniques for sample collection and processing as per guidelines (6). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use. For professional use only. Read the label before opening the container. Wear protective gloves/ protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Further biochemical identification is required for confirmation of species.
2. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
3. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to light pink colored granular medium

Gelling

Firm, comparable with 1.5% Agar gel.

Colour and Clarity of prepared medium

Red coloured clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 6.37% w/v aqueous solution at 25°C pH : 7.2±0.2

pH

7.00-7.40

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony	H ₂ S production
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	good-luxuriant	≥50%	red with bile precipitate	negative
<i>Klebsiella pneumoniae</i> ATCC 10031	50-100	good-luxuriant	≥50%	pink	negative
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50-100	good-luxuriant	≥50%	colourless	positive
<i>Salmonella</i> Enteritidis ATCC 13076 (00030*)	50-100	good-luxuriant	≥50%	colourless	positive
## <i>Proteus hauseri</i> ATCC 13315	50-100	fair-good	30-40%	pink with brownish zone	negative
<i>Proteus mirabilis</i> ATCC 25933	50-100	good-luxuriant	≥50%	colourless with brownish zone	variable
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50-100	fair-good	30-40%	colourless	negative
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	≥10 ⁴	inhibited	0%	-	-
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50-100	none-poor	≤10%	colourless	negative
<i>Bacillus cereus</i> ATCC 10876	≥10 ⁴	inhibited	0%	-	-

Key : *Corresponding WDCM numbers.

Formerly known as *Proteus vulgaris*

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference

1. Krieg N. R. and Holt J. G., (Eds.), 1984, Bergeys Manual of Systematic Bacteriology Vol. I, P-408-516. Williams and Wilkins Co. Baltimore.
2. Murray P. R., Baron J. H., Pfaller M. A., Jorgensen J. H. and Tenover F. C., (Ed.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C.
3. Sakazaki R., Namioka S., Osada A., and Yamada C. A., 1960, Japan. J. Ex. Med., 30; 13-22.
4. Sakazaki R., Tamura K., Prescott L. M., Benzie Z., Sanyal S. C., and Sinha, R., 1971, Indian J. Med. Res., 59; 1025-1034.
5. Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W.C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4th Ed., J. B. Lippincott Company.
6. Salfinger Y., and Tortorello M.L., 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
7. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
8. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 01/2026



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC,Wagle Industrial Area,
Thane (W) -400604, MS, India



IVD *In vitro diagnostic
medical device*



Storage temperature



AR Experts BV
Boeingavenue 209
1119 PD Schiphol-Rijk
The Netherlands



CE Marking



**Do not use if
package is damaged**

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.