

## Violet Red Bile Glucose Agar w/o Lactose, Granulated®

GM581

### Intended Use:

Recommended for enumeration of *Enterobacteriaceae* in raw food and clinical samples. The composition and performance criteria are in accordance with ISO 21528-1&2:2017 and ISO 11133:2014 /Amd.2 : 2020 (E).

### Composition\*\*

#### ISO Specifications-Violet Red Bile Glucose

#### Agar w/o Lactose

Ingredients	g / L
Enzymatic digest of animal tissues	7.000
Yeast extract	3.000
Sodium chloride	5.000
Bile salts No.3	1.500
Glucose	10.000
Neutral red	0.030
Crystal violet	0.002
Agar	9.000-18.000
Final pH ( at 25°C)	7.4±0.2

#### Violet Red Bile Glucose Agar w/o Lactose, Granulated®

GM581

Ingredients	g / L
Peptone \$	7.000
Yeast extract	3.000
Sodium chloride	5.000
Bile salts mixture	1.500
Glucose (Dextrose)	10.000
Neutral red	0.030
Crystal violet	0.002
Agar	12.000
Final pH ( at 25°C)	7.4±0.2

\*\*Formula adjusted, standardized to suit performance parameters

\$ -Equivalent to Enzymatic digest of animal tissues

### Directions

Suspend 38.53 gram in 1000 ml purified/distilled water. Heat to boiling to dissolve the medium completely. **DO NOT AUTOCLAVE.** Cool to 45-50°C. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

Violet Red Bile Agar, a modification of MacConkey original formulation (1) is used for the enumeration of coli-aerogenes bacterial group. Violet Red Bile Glucose Agar w/o Lactose, a modification of VRBA (M049), was designed for the enumeration of *Enterobacteriaceae* (2). It employs the selective inhibitory components crystals violet and bile salts and the indicator system glucose and neutral red. Sought bacteria will dissimilate glucose and produce purple zones around the colonies (3). ISO committee has also recommended this medium (4,5). Selectivity of VRBGA can be increased by incubation under anaerobic conditions and/or at elevated temperature, i.e. equal to or above 42°C (6-8).

Peptone and yeast extract serve as sources of carbon, nitrogen, vitamins and other essential growth nutrients. Glucose is the fermentable carbohydrate, utilization of which leads to the production of acids. Neutral red indicator detects the acidity so formed. Crystal violet and bile salts mixture help to inhibit the accompanying gram-positive and unrelated flora. Sodium chloride maintains the osmotic equilibrium. Further biochemical tests are necessary for positive identification (9).

### Type of specimen

Clinical samples - faeces ; Food and dairy samples; Water samples

### Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (10,11).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (4,5,12-14).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines & local standards (15).

After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

In Vitro diagnostic use. For professional use only. Read the label before opening the container. Wear protective gloves/ protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations

1. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.

- Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
- Over incubation may result in reverting of reaction.
- Further biochemical tests must be carried out on colonies of pure culture for confirmation.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Light yellow to pinkish beige coloured granular media.

### Gelling

Firm, comparable with 1.2% Agar gel.

### Colour and Clarity of prepared medium

Reddish purple coloured clear to slightly opalescent gel forms in Petri plates.

### Reaction

Reaction of 3.85% w/v of aqueous solution at 25°C. pH : 7.4±0.2

### pH

7.20-7.60

### Cultural Response

**Productivity** : Cultural characteristics was observed after an incubation at 35±1°C for 24±2 hours. Recovery rate is considered as 100% for bacteria growth on Soyabean Casein Digest Agar.

**Selectivity**: Cultural characteristics was observed after an incubation at 35±1°C for 24±2 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
<b>Productivity</b>				
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	luxuriant	≥50 %	pink to red colonies with or without precipitation zone
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	luxuriant	≥50 %	pink to red colonies with or without precipitation zone
<i>Salmonella</i> Enteritidis ATCC 13076 (00030*)	50 -100	luxuriant	≥50 %	pink to red colonies with or without precipitation zone
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	luxuriant	≥50 %	pink to red colonies with or without precipitation zone
<b>Selectivity</b>				
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	≥10 <sup>4</sup>	inhibited		
<i>Enterococcus faecalis</i> ATCC 19433 (00009*)	≥10 <sup>4</sup>	inhibited		

Key : (\*) Corresponding WDCM numbers.

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (10,11).

## Reference

- MacConkey A., 1905, J. Hyg., 5, 333-379.

2. Mossel D. A. A., Eclerink I., Koopmans M. and Van Rossem F., 1978, Lab. practice, 27 No. 12: 1049.
3. Corry J. E. L., Curtis G. D. W. and Baird R. M., (Ed.), 1995, Culture Media for Food Microbiology, Vol. 34, Progress in Industrial Microbiology, Elsevier, Amsterdam.
4. Microbiology of the food chain Horizontal method for the detection and enumeration of *Enterobacteriaceae* Part 2: Colony-count technique ISO 21528-2:2017.
5. Microbiology of food, animal feeding stuffs and water- Preparation, production, storage and performance testing of culture media, EN ISO 11133:2014 (E) /Amd.: 2020
6. Mossel D. A. A. and Vega C. L., 1973, Hlth. Lab. Sci., 11:303
7. Mossel D. A. A., Eclerink I., Koopmans M. and Van Rossem F., 1979, Food Protect., 42 : 470
8. Mossel D. A. A. et al, 1986, J. Appl. Bacteriol., 60:289.
9. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.
10. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition. Clinical Microbiology, 11th Edition. Vol.
11. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
12. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
13. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
14. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
15. Lipps WC, Braun-Howland EB, Baxter TE, eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.


Revision : 08/2026



HiMedia Laboratories Pvt. Limited,  
Plot No.C-40, Road No.21Y,  
MIDC, Wagle Industrial Area,  
Thane (W) -400604, MS, India



**In vitro diagnostic  
medical device**



**Storage temperature**



AR Experts BV  
Boeingavenue 209  
1119 PD Schiphol-Rijk  
The Netherlands



**CE Marking**



**Do not use if  
package is damaged**

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.