



Thymol Blue Indicator

I011

Intended use

Thymol Blue Indicator is recommended as a pH indicator.

Composition**

Ingredients

Thymol blue	0.04 gm
Distilled water	100.0 ml

**Formula adjusted, standardized to suit performance parameters

Directions

1. Thymol Blue Indicator has wide range of application so follow appropriate direction as per application protocol.

Principle And Interpretation

Thymol blue is used as a pH indicator. It transitions from red to yellow at pH 1.2–2.8 and from yellow to blue at pH 8.0–9.6. It is usually a component of Universal indicator.

Type of specimen

Biological sample

Specimen Collection and Handling

Follow appropriate techniques for handling specimens as per established guidelines

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. An indicator is not functional above its pH range because the indicator does not change color at these pH values.
2. If the substance or sample is contaminated, the color may be wrong.
3. Acid-base indicators show just one- or two-color changes.
4. Indicators measure pH at low accuracy, they only indicate sample acidity or alkalinity and not exact pH

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature

Quality Control

- **Appearance** : Dark blue coloured solution.
- **Clarity** : Clear without any particles
- **Reaction** : Reaction of solution is a colour change of red to yellow at pH 1.2 to 2.8 and from yellow to blue at pH 8.0 to 9.6 respectively

Storage and Shelf Life


Store between 10- 30°C in tightly closed container and away from bright light. Use before expiry date on label. On opening, product should be properly stored in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Reference

1. Chemistry in lab reagents and resources; The preparation of titration indicators; Dhanal De Lloyd, chem. Dept.
2. Text Book of Medical Laboratory Technology; Praful B. Godka.
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
4. INDIAN PHARMACOPOEIA 2018, VOL-I, 4.3(962)



Storage temperature



Do not use if package is damaged



In vitro diagnostic medical device



CE Marking



HiMedia Laboratories Pvt. Limited,
C-40, Road No.21Y, MIDC, Wagle
Industrial Area, Thane (W) - 400604,
MS, India



CEpartner4U,ESDOORNLAAN 13,3951
DB MAARN,The Netherlands,
www.cepartner4u.eu

Revision: 02/2022

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.