

LA1092 Salivol[®]- Container for Sputum & Saliva Sample

Intended Use

For the collection of human nucleic acid from saliva samples.

Introduction

Saliva has become an ideal translational research tool and diagnostic medium as it provides a “window” into the oral and systemic health of an individual, and like other bodily fluids, saliva can be analyzed and studied to diagnose diseases. It is being used in novel ways to provide molecular biomarkers for a variety of oral and systemic diseases and conditions. Improper collection of saliva can lead to data errors. However, by following proper saliva collection and handling procedures high quality, reproducible data can be obtained from this sample type. Saliva samples can be collected in a convenient, minimally-invasive, and repeated way by using Salivol[®].

Each unit of HiMedia’s Salivol[®] contains one saliva collection funnel with nucleic acid stabilizer and a collection tube.



**Saliva Collection
Funnel with
Stabilizer**

Salient features:

1. Reliable and cost-effective
2. Simple and non-invasive saliva collection
3. Sample collection and preservation in one convenient kit
4. Preservation of DNA/RNA in saliva samples at ambient temperature
5. Compatible with most DNA/RNA isolation methods

Directions for use:

Do not eat, drink, smoke or chew gum for 30 minutes before giving your saliva sample.

Do not remove the plastic film from the funnel lid.



1
Take the tube and Salivol®



2
Remove the red cap and place it below the Salivol®



3
Fit the tube and Salivol® tightly



4
Spit the saliva in the funnel of Salivol®



5
Press the upper portion of Salivol™ so that the solution and saliva slides down towards the tube



6
Close the tube with the help of red cap and tighten the tube



7
Tilt the tubes in upside down position for proper mixing

Warning and precautions:

Wash with water if stabilizing liquid comes in contact with eyes or skin. Please refer the product MSDS.

Please refer disclaimer Overleaf.












Sample Storage:

Saliva specimen collected in Salivol® can be stored for a week at 25°C, one month at 4°C.


Technical Assistance

At HiMedia we pride ourselves on the quality and availability of our technical support. For any kind of technical assistance, mail at mb@himedialabs.com.

Symbol

	Manufacturer		Do not use if package is damaged
	Authorized representative in the European Community		Temperature limit
	Date of manufacture (YYYY-MM)		Consult instructions for use
	Use-by date (YYYY-MM)		In vitro diagnostic medical device
	Batch code		CE marking of conformity
	Catalogue number		

Authorized representative (AR) Address:

	AR Experts B.V. Boeingavenue 209, 1119 PD, Schiphol-Rijk, The Netherlands
---	--

Identification No.: PILA1092

Rev No.: 02

Date of Issue.: 2025-12

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Reg.office : Plot No. C-40, Road No. 21Y, MIDC, Wagle Industrial Area, Thane, (West) 400604, Maharashtra, INDIA.
Customer Care No.: 00-91-22-6116 9797 Tel: 00-91-22-6147 1919, 6903 4800 Email: techhelp@himedialabs.com Website: www.himedialabs.com