



Columbia Broth

LQ006A

Intended Use

Recommended for cultivation of fastidious organisms from blood. *Sterile, in glass bottles.*

Composition**

Ingredients	g / L
Peptone special	10.000
Biopeptone	10.000
HI powder #	3.000
L-Cysteine hydrochloride	0.100
Dextrose (Glucose)	2.500
Sodium chloride	5.000
Magnesium sulphate	0.100
Ferrous sulphate	0.020
Sodium carbonate	0.600
Tris (hydroxymethyl) aminomethane	0.830
Tris (hydroxymethyl) aminomethane hydrochloride	2.860
Final pH (at 25°C)	7.5±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Heart infusion powder

Directions

Label the ready to use blood culture bottle. Remove the Aluminium foil cap. Disinfect the part of the rubber stopper which is now exposed. Draw patient's blood with the sterile or disposable needle and syringe as explained in specimen collection and disposable column. Transfer the blood sample immediately into the culture bottle by puncturing the rubber stopper with the needle and injecting the blood. Venting: Use sterile venting needle (LA038). Keep the bottle in an upright position preferably in a biological safety cabinet, place an alcohol swab over the rubber stopper and insert the venting needle with filter through it. Insertion and withdrawal of the needle should be done in a straight line. discard the needle and mix the contents by gently inverting the bottle 2-3 times. Do Not vent the bottle for anaerobic cultures. Incubate at 35-37°C for 18-48 hours. and further for seven days.

Principle And Interpretation

Morello and Ellner in 1969 devised a liquid medium for the recovery of microorganisms from blood cultures (1). This medium was devised from Columbia Blood Agar Base previously formulated by Ellner et al (2). While studying they found that Columbia Broth was superior to a commonly used general-purpose broth for faster growth of *Staphylococcus aureus*, *Escherichia coli*, viridans Streptococci and *Enterococcus* groups. In the formulation the increased concentration of cystine is provided for improved recovery of both aerobic and anaerobic microorganisms from blood specimens. It is an excellent blood culture medium (3). It differs from the agar base in that the cornstarch is omitted to reduce opalescence (4) and salts have been included. Medium contains peptone special, biopeptone and HI powder to support luxurious growth of the organisms. Dextrose is added as a carbon and energy source. The medium is buffered with tris buffer. The addition of salts was found to be beneficial for the recovery of organisms. L-Cysteine HCL is the reducing agent. Magnesium & iron are added to facilitate organism growth. Growth in tubes is indicated by presence of turbidity compared to an uninoculated control. If growth appears, cultures should be subcultured onto appropriate media.

Type of specimen

Clinical samples : Blood

Specimen Collection and Handling :

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (2,3). After use, contaminated materials must be sterilized by autoclaving before discarding.

Please refer disclaimer Overleaf.

Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
2. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
3. Recommended for Paediatric use only.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Sterile clear Columbia Broth in glass bottle.

Colour

Light amber coloured clear solution

Quantity of Medium

70 ml of medium in glass bottle. (For adult use)

Reaction

7.30- 7.70

Sterility test

Passes release criteria

Cultural response

Cultural characteristics was observed after incubation at 35-37°C for 18-48 hours.

Organism	Inoculum (CFU)	Growth	Growth under anaerobic conditions
<i>Streptococcus mitis</i> ATCC 9811	50-100	luxuriant	-
<i>Streptococcus pyogenes</i> ATCC 19615	50-100	luxuriant	-
<i>Neisseria meningitidis</i> ATCC 13090	50-100	luxuriant	-
<i>Clostridium perfringens</i> ATCC 12924	50-100	-	luxuriant
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	luxuriant	

Key : (*) Corresponding WDCM numbers.

Storage and Shelf Life

Store between 15-30°C. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (3,4).

Reference

1. Morello J. A. and Ellner P. D., 1969, Appl. Microbiol. 17:68.
2. Ellner P. D., Stoessel C. J., Drakeford E. and Vasi F., 1966, Am. J. Clin. Pathol., 45:502
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.

4.Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W.(2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 03/2023



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



In vitro diagnostic
medical device



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



Do not use if
package is damaged

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.