



Fluid Thioglycollate Medium w/ 0.05% SPS

LQ023A

Intended use

Recommended for cultivation of aerobes, anaerobes and microaerophiles. *Sterile, in glass bottles.*

Composition**

Ingredients	g / L
Tryptone	15.000
Yeast extract	5.000
Dextrose (Glucose)	5.500
Sodium chloride	2.500
L-Cystine	0.500
Sodium thioglycollate	0.500
Resazurin sodium	0.001
Agar	0.750
SPS	0.500
Final pH (at 25°C)	7.1±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Label the ready to use blood culture bottle. Remove the Aluminium foil cap. Disinfect the part of the rubber stopper which is now exposed. Draw patient's blood with the sterile or disposable needle and syringe as explained in specimen collection and disposable column. Transfer the blood sample immediately into the culture bottle by puncturing the rubber stopper with the needle and injecting the blood. Venting: Use sterile venting needle (LA038). Keep the bottle in an upright position preferably in a biological safety cabinet, place an alcohol swab over the rubber stopper and insert the venting needle with filter through it. Insertion and withdrawal of the needle should be done in a straight line. discard the needle and mix the contents by gently inverting the bottle 2-3 times. Do Not vent the bottle for anaerobic cultures. Incubate at 35-37°C for 24-48 hours. and further for seven days.

Principle And Interpretation

Brewer (1) formulated Fluid Thioglycollate Medium for rapid cultivation of aerobes as well as anaerobes including microaerophiles by adding a reducing agent and small amount of agar. The USP (2), BP (3), EP (4) and AOAC (5) have recommended the media for sterility testing of antibiotics, biologicals and foods and for determining the phenol coefficient and sporicidal effect of disinfectants. However, it is intended for the examination of clear liquid or water-soluble materials. Fluid Thioglycollate Medium is also routinely used to check the sterility of stored blood in blood banks (6).

Dextrose, tryptone, yeast extract, L-cystine provide the growth factors necessary for bacterial multiplication. L cystine and sodium thioglycollate allows *Clostridium* to grow in this medium even under aerobic conditions (7). Also the small amount of agar used in the medium favors the growth of aerobes as well as anaerobes in the medium, even if sodium thioglycollate is deleted from the medium (1). Sodium thioglycollate act as a reducing agent and neutralizes the toxic effects of mercurial preservatives and peroxides formed in the medium, thereby promoting anaerobiosis, and making the medium suitable to test materials containing heavy metals. (8,9). Any increase in the oxygen content is indicated by a colour change of redox indicator, resazurin to red (10). The small amount of agar helps in maintaining low redox potential for stabilizing the medium (9). SPS (Sodium polyanethol sulphonate, SPS) is used as a nontoxic anticoagulant which enables bacterial growth and prevents the action of natural bacterial inhibitors of blood. Van Haebler and Miles (11) first demonstrated the usefulness of SPS in Blood culture medium.

Note: If more than upper one-third of the medium has acquired a pink-purple colour, the medium may be restored once by heating in a water bath or in free flowing steam until the pink colour disappears.

Type of specimen

Clinical samples- blood samples.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (12,13).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use. For professional use only. Read the label before opening the container. Wear protective gloves/ protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. It is intended for the examination of clear liquid or water-soluble materials.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Sterile clear Fluid Thioglycollate Medium w/ 0.05% SPS in glass bottle.

Colour

Light straw coloured solution with upper 10% or less medium pink-purple on standing.

Quantity of Medium

70ml of medium in glass bottle, (For Adult use)

pH

6.90-7.30

Sterility Check

Passes release criteria

Cultural response

Cultural characteristics was observed after incubation at 35-37°C for 24-48 hours for bacteria and at 22-28°C for fungus for 48-72 hours

Organism	Inoculum (CFU)	Growth
<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	luxuriant
# <i>Bacillus spizizenii</i> ATCC 6633 (00003*)	50 -100	luxuriant
<i>Streptococcus pyogenes</i> ATCC 19615	50 -100	luxuriant
<i>Neisseria meningitidis</i> ATCC 13090	50 -100	luxuriant
<i>Clostridium sporogenes</i> ATCC 11437	50 -100	luxuriant
<i>Micrococcus luteus</i> ATCC 10240	50 -100	luxuriant

Key : *Corresponding WDCM numbers. # - Formerly known as *Bacillus subtilis* subsp. *spizizenii*

Storage and Shelf Life

Store between 15-30°C. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (12,13).

Reference

1. Brewer, 1940, J. Am. Med. Assoc., 115:598.
2. The United States Pharmacopoeia-National Formulary (USP-NF), 2022.
3. The British Pharmacopoeia, 2022, Medicines and Healthcare products Regulatory Agency.
4. European Pharmacopoeia, 2022, 10 th volume, European Directorate for the quality of medicines & Healthcare.
5. Williams H., (Ed.), 2005, Official Methods of Analysis of the Association of Official Analytical Chemists, 19th Ed., AOAC, Washington, D.C.
6. Portwood, 1944, J. Bact., 48:255. Quastel and Stephenson, 1926, J. Biochem., 20

7. Federal Register, 1992, Fed. Regist., 21:640.
8. MacFaddin J.F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.
9. Marshall, Gunnison and Luxen, 1940, Proc. Soc. Exp. Biol. Med., 43:672.
10. Nungester, Hood and Warren, 1943, Proc. Soc. Exp. Biol. Med., 52:287.
11. Van Haebler T., A.A. Miles, 1938. The action of sodium polyanethol sulphonate (liquoid) on Blood Cultures Journal of Pathology.46:245
12. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
13. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W.(2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 03/2023



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



**In vitro diagnostic
medical device**



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



**Do not use if
package is damaged**

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.