

# НіСтоте<sup>тм</sup> Enterococci Broth

## Intended use

Recommended for identification and differentiation of Enterococci from water samples.

## **Composition\*\***

Ingredients	g / L
Peptone, special	10.000
Sodium chloride	5.000
Sodium azide	0.300
Chromogenic substrate	0.040
Polysorbate 80 (Tween 80)	2.000
Disodium hydrogen phosphate	1.250
Final pH ( at 25°C)	7.5±0.2

\*\*Formula adjusted, standardized to suit performance parameters

## Directions

Suspend 18.59 gram (single-strength) or 37.18 grams (double-strength) in 1000 ml purified/distilled water. Heat if necessary to dissolve the medium completely. Dispense in tubes or flasks as desired and sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes.

## **Principle And Interpretation**

HiCrome<sup>TM</sup> Enterococci Broth is formulated on the basis of the work carried out by Althous et al (1), Amoras (2), Litsky et al (3), and Manafi and Sommer (4) and Snyder and Lichstein (5). These media is recommended for the rapid detection of Enterococci from water samples. The presence of *Enterococcus* group, which is a subgroup of the faecal Streptococci, serves as a valuable bacterial indicator for determining the extent of faecal contamination (1,6) and it is more specific than the detection of coliforms, which may originate from non-faecal sources. The enzyme  $\beta$ -glucosidase produced by Enterococci cleaves the chromogenic substrate, resulting in a bluish green colour.

The medium contains peptone special, which provides carbonaceous, nitrogenous compounds, long chain amino acids, vitamins and other essential nutrients. Sodium chloride maintains the osmotic balance of the medium. Sodium azide inhibits the accompanying microflora, especially gram-negative organisms. Polysorbate 80 acts as a source of fatty acids.

## **Type of specimen**

Water samples

## **Specimen Collection and Handling:**

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards (7). After use, contaminated materials must be sterilized by autoclaving before discarding.

## Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

## Limitations

- 1. Some species may show poor growth due to nutritional variations.
- 2. Slight colour variations may be observed depending upon the utilization of the substrate by the organism.
- 3. Further biochemical and serological test are need to be carried out for confirmation.

## **Performance and Evaluation**

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

M1376

## **Quality Control**

#### Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light amber coloured, clear solution in tubes

#### Reaction

Reaction of 1.86% w/v aqueous solution at 25°C. pH : 7.5±0.2

#### pН

7.30-7.70

#### **Cultural Response**

Cultural characteristics observed after an incubation at 35- 37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Colour of Medium
Escherichia coli ATCC 25922 (00013*)	50-100	none-poor	light yellow
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50-100	luxuriant	Light blue-green
Pseudomonas aeruginosa ATCC 27853 (00025*)	50-100	none-poor	light yellow
Staphylococcus aureus subsp. aureus ATCC 25923 (00034*)	50-100	Fair-good	light yellow

Key : (\*) Corresponding WDCM numbers.

#### Storage and Shelf Life

Store between 15-25°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

#### Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (8,9).

#### Reference

- 1. Althous, H., Dott, W., Havemeister, G, Muller, H.E, a. Sacre, C., 1982, Zbl. Bakt. Hyg. I. Abt. Orig. A. 252:154-165.
- 2. Amoras I, 1995, Poster presentation congress of Spanish Society of Microbiology, Madrid.
- 3. Litsky, W., Mallmann, W.L., a Fifield, C.W. 1953, Amer. J. Pbl. Hlth. 43:873-879.
- 4. Manafi M., and Sommer R, 1993, Wat. Sci. Tech. 27:271-274.
- 5. Snyder M.L., and Lichstein, H.C. 1940, J. Infect. Dis. 67. 113-115
- 6. Standard Methods for the Examination of Water and Wastewater, 20th Edition, Edited by L.S. Clesceri, A.E. Greenberg and A.D. Eaton, Published by APHA, AWWA and WEF (1998).
- 7. Lipps WC, Braun-Howland EB, Baxter TE,eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.
- 8. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
- 9. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 06/2024

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia<sup>TM</sup> publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia<sup>TM</sup> Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Corporate Office : Plot No.C-40, Road No.21Y, MIDC, Wagle Industrial Area, Thane (W) - 400604, India. Customer care No.: 022-6147 1919 Email: techhelp@himedialabs.com Website: www.himedialabs.com