



HiCrome® VRE Agar Base

M1830

Intended Use

Recommended as a selective media for isolation of Vancomycin Resistant Enterococci (VRE) from clinical specimens.

Composition**

Ingredients	g / L
Peptone special	25.000
Chromogenic mixture	0.450
Sodium chloride	5.000
Buffering agent	1.250
Salt mixture	4.250
Agar	15.000
Final pH (at 25°C)	6.5±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 50.95 gram in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C and aseptically add the rehydrated contents of two vials of VF Selective Supplement (FD277). Mix well and pour into sterile Petri plates.

Principle And Interpretation

Enterococci are the common habitants of the normal flora residing in the intestines of mammals (1). Vancomycin Resistant Enterococci are the group of Enterococci that have developed resistance towards many antibiotics particularly vancomycin. Enterococcal infections that result in human disease can be fatal, particularly those caused by strains of vancomycin-resistant enterococci (VRE) (2). Early detection of VRE is important to prevent the emergence of vancomycin resistant in *Enterococcus faecalis*. VRE can be transmitted from person to person, especially in a hospital or chronic-care facility. Microscopic amounts of fecal material from an infected or colonized patient can contaminate the hospital environment and be a reason for the spread of infection. There are many traditional media for the detection of VRE which includes Vancomycin Resistant Enterococci Broth Base/Agar or Bile Esculin Agar supplemented with vancomycin. Peptone special in the medium supplies nitrogenous and carbonaceous compounds, long chain amino acids, vitamins and other necessary nutrients required for the growth of microorganisms. Sodium chloride maintains the osmotic balance. Buffering agents provides buffering to the medium. *Enterococcus faecalis* cleaves the chromogenic substrate in the medium to produce blue coloured colonies, which are clearly visible against the opaque background. The supplement added to the medium allows the selective isolation of Vancomycin Resistant Enterococci. This medium can be inoculated directly from screening swab, isolated colony prepared as a liquid suspension approximately equivalent to 0.5 McFarland turbidity.

Type of specimen

Clinical samples - faeces, urine, etc.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (3,4).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic Use only. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Some intermediate species may show poor growth due to nutritional variations and tolerance to vancomycin.
2. Slight colour variation may be observed depending upon the utilization of the substrate by the organism.
3. Interspecies differentiation between *Enterococcus faecalis* and *Enterococcus faecium* cannot be confirmed.
4. Further confirmation has to be carried using sensitivity testing.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel.

Colour and Clarity of prepared medium

Off white coloured opaque gel forms in Petri plates.

Reaction

Reaction of 5.1% w/v aqueous solution at 25°C. pH : 6.5±0.2

pH

6.30-6.70

Cultural Response

Cultural characteristics observed with added VF Selective Supplement (FD277), after an incubation at 35-37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
<i>Enterococcus faecalis</i> (VRE) ATCC 51299 (00085*)	50-100	luxuriant	≥50%	bluish green
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	≥10 ⁴	inhibited	0%	
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	≥10 ⁴	inhibited	0%	

Key : * Corresponding WDCM numbers.

Storage and Shelf Life

Store between 15-25°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (3,4).

Reference


1. Mara D., Horan NJ: The Handbook of water, wastewater and microbiology, Amsterdam, The Netherlands, Academic Press; 2003.
2. Mascini EM, Bonten MJ: Vancomycin- resistant enterococci: consequences for therapy and infection control. Clin Microbiol Infect. 2005;11 (Suppl.4) :43-56
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



**In vitro diagnostic
medical device**



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



**Do not use if
package is damaged**

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.