



Acetate Differential Agar

M339

Intended Use:

Recommended for the differentiation of *Shigella* species from *Escherichia coli*.

Composition**

Ingredients	Gms / Litre
Sodium acetate	2.000
Magnesium sulphate	0.100
Sodium chloride	5.000
Monoammonium phosphate	1.000
Dipotassium hydrogen phosphate	1.000
Bromothymol blue	0.080
Agar	20.000
Final pH (at 25°C)	6.7±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 29.18 grams in 1000 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Distribute in tubes in sufficient amounts to give butt and slant. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Allow the tubes to cool in a slanted position.

Principle And Interpretation

Acetate Differential Agar was formulated by Trabulsi and Ewing (6). Tatum, Ewing and Weaver (5) modified the medium by replacing sodium citrate by sodium acetate, which enables the differentiation of *Shigella* species from *Escherichia coli*. Organic acids have been used widely as an aid in the differentiation of *Enterobacteriaceae*, usually in formulae that contained organic nitrogen sources. Most bacteria, however, can use citrate and acetate in the presence of organic nitrogen. The differentiation of groups is based on the ability or failure of the test culture to utilize acetate in a medium devoid of trace organic nitrogen. This medium contains sodium acetate as the sole source of carbon. Trabulsi and Ewing demonstrated that *Shigella* and *Proteus* species are unable to utilize acetate and therefore fails to grow. Majority of *Escherichia coli* and closely related organisms grow well within 24-48 hours but some strains grow very slowly and a few strains are unable to utilize acetate as a sole carbon source. Acetate utilization is indicated by formation of blue colour, which is due to the utilization of sodium acetate and subsequent formation of an alkaline reaction detected by the presence of bromothymol blue indicator.

Sodium acetate is utilized as a sole source of carbon by some serotypes of *S.flexneri* such as *Shigella flexneri* 4a (1,4). Magnesium sulphate is essential ion. Sodium chloride maintains osmotic equilibrium and phosphates act as buffers.

Type of specimen

Isolated Microorganism

Specimen Collection and Handling

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Some strains of *Escherichia coli* utilize acetate slowly or not at all and therefore may produce a false negative reaction.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to light green homogeneous free flowing powder

Gelling

Firm, comparable with 2.0% agar gel.

Colour and Clarity of prepared medium

Emerald green coloured clear to slightly opalescent gel forms in tubes as slants

Reaction

Reaction of 2.92% w/v aqueous solution at 25°C. pH : 6.7±0.2

pH

6.50-6.90

Cultural Response

Cultural characteristics observed after an incubation at 25-30°C for upto 1-7 days.

Organism	Inoculum (CFU)	Growth	Acetate utilization
<i>Citrobacter freundii</i> ATCC 8090	50-100	good-luxuriant	positive reaction, blue colour
<i>Enterobacter cloacae</i> ATCC 23355 (00082*)	50-100	good-luxuriant	positive reaction, blue colour
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	good-luxuriant	positive reaction, blue colour
<i>Klebsiella pneumoniae</i> ATCC 13883 (00097*)	50-100	good-luxuriant	positive reaction, blue colour
<i>Proteus vulgaris</i> ATCC 13315	≥10 ⁴	inhibited	
<i>Salmonella</i> Arizonae ATCC 13314	50-100	good-luxuriant	positive reaction, blue colour
<i>Salmonella</i> Typhi ATCC 19430	50-100	poor	negative reaction green colour
<i>Shigella sonnei</i> ATCC 25931	50-100	none-poor	negative reaction, no change, medium remains green

Key : *- Corresponding WDCM numbers

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (2,3).

Reference

1. Ewing, 1986, Edwards and Ewings Identification of *Enterobacteriaceae*, 4th Ed. Elsevier Science Publishing Co., Inc., New York.
2. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
3. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
4. Talukder K. A, Islam M. A., Dutta D.K., Hasan F., Sada A., Nair G. B . and Sack D. A., 2002, J. Clin. Microbiol., 40:2490
5. Tatum H. W., Ewing W. H., and Weaver R. E., 1974, Manual of Clinical Microbiology, 2nd Ed., American Society for Microbiology, Washington D.C. Pg.-270
6. Trabulsi and Ewing, 1962, Public Health Lab., 20:137.

Revision : 02 / 2019

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Reg.office : 23, Vadhani Ind.Est., LBS Marg, Mumbai-400086, India. Customer care No.: 022-6116 9797 Corporate office : A-516, Swastik Disha Business Park, Via Vadhani Ind. Est., LBS Marg, Mumbai-400086, India. Customer care No.: 022-6147 1919 Email: techhelp@himedialabs.com Website: www.himedialabs.com