



Nitrate Broth

M439

Intended Use:

Recommended for detection of nitrate reduction by bacteria. The composition and performance criteria are in accordance with, ISO 7932:1993.

Composition**

ISO 7932:1993 Specification-Nitrate Medium

Ingredients	g / L
Peptone	5.000
Beef extract	3.000
Potassium nitrate (KNO ₃)	1.000
Final pH (at 25°C)	7.00

M439-Nitrate Medium

Ingredients	g / L
Peptone	5.000
HM peptone B #	3.000
Potassium nitrate	1.000
Final pH (at 25°C)	7.0±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Beef extract

Directions

Suspend 9.00 grams in 1000 ml purified / distilled water. Heat if necessary to dissolve the medium completely. Dispense in tubes, flasks or as desired. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes.

Principle And Interpretation

Nitrate Broth is prepared in accordance with the formula published in Pure Culture Study of Bacteria of the Society of American Bacteriologist (1). The ability to reduce nitrate is valuable for differentiating and identifying various types of bacteria especially *Enterobacteriaceae* family (2). Nitrate Broth is also recommended by the ISO Committee for detection of nitrate reduction by bacteria(3). Non-fermenters and other miscellaneous gram-negative bacilli vary in their ability to reduce nitrates. Some members of this group are capable of denitrification which is the reduction of nitrate to nitrogen gas. For the glucose fermenting gram-negative bacilli, the production of nitrogen gas from nitrate is an important differential test (4). Potassium nitrate in the medium acts as a substrate for determining nitrate reduction by bacteria. Certain bacteria convert nitrate to nitrite, ammonia or nitrogen gas. The presence of nitrites can be detected by the addition of 0.5 ml each of sulphanilic acid (R015) and α -naphthylamine solution (R009). The development of red violet colour, due to the formation of a red diazonium dye i.e. p-sulfobenzene-azo-a-naphthylamine, indicates nitrate reduction to nitrite. If no colour develops, it means that either nitrate is not reduced or further reduction to ammonia or nitrogen gas has taken place. This can be verified by adding a pinch of zinc dust to the tube. Zinc reduces nitrate to nitrite resulting in a red colour. The red colour indicates that nitrate is still present and was not reduced previously. An absence of red colour after the addition of zinc dust indicates that no nitrate is present and thus the nitrate was reduced further than nitrite. Therefore the nitrate reduction test is evidenced by either the presence of a catabolic end product or the absence of nitrate in the medium. Members of *Enterobacteriaceae* characteristically reduce nitrate to nitrite which reacts with sulfanilic acid and N, N-dimethyl-1-naphthylamine to produce the red colour. This reaction is known as Griess reaction. If an organism grows rapidly and reduces nitrate actively, the test should be performed after an early incubation period since the nitrite may be further reduced to nitrogen.

Type of specimen

Isolated Microorganisms from clinical and non-clinical samples.

Specimen Collection and Handling

For the test: Add few drops of each reagent i.e. sulphanilic acid (R015) and α -naphthylamine solution (R009) into the tube containing culture to be tested. A distinct red or pink colour indicates nitrate reduction. A control (un-inoculated) tube should also be tested. If there is no pink colour formation, add a pinch of zinc dust to confirm the absence of nitrate in the medium (4).

Warning and Precautions

Please refer disclaimer Overleaf.

In Vitro diagnostic use. For professional use only. Read the label before opening the container. Wear protective gloves/ protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Nitrate reduction is not a confirmatory test. Complete identification of bacteria should include the morphology, gram reaction, biochemical and serological tests. Addition of excess zinc may result in false negative reaction. Also during performance of nitrate reduction test with α -naphthylamine, the colour produced in a positive reaction may fade quickly (6).
2. The nitrate test is very sensitive and an uninoculated nitrate control should be tested with reagents to determine whether medium is nitrite free (5).

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light amber coloured clear solution in tubes

Reaction

Reaction of 0.9% w/v aqueous solution at 25°C. pH : 7.0±0.2

pH

6.80-7.20

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours. Nitrate reduction observed on addition of 0.5ml of sulphanilic acid (R015) and 0.5ml of α -naphthylamine Solution (R009).

Organism	Growth	Nitrate reduction
<i>Acinetobacter calcoaceticus</i> ATCC 19606	luxuriant	Negative reaction
<i>Escherichia coli</i> ATCC 25922 (00013*)	luxuriant	Positive reaction, distinct red-pink colour developed within 1-2 minutes
# <i>Klebsiella aerogenes</i> ATCC 13048 (00175*)	luxuriant	Positive reaction, distinct red-pink colour developed within 1-2 minutes
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	luxuriant	Positive reaction, distinct red-pink colour developed within 1-2 minutes

Key : (*) Corresponding WDCM numbers. (#) Formerly known as *Enterobacter aerogenes*

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

Please refer disclaimer Overleaf.

Reference

1. Society of American Bacteriologist, Pure Culture Study of Bacteria, 1944, 12 : Leaflet 11: 8.
2. Ewing, 1986, Edwards and Ewings Identification of *Enterobacteriaceae*, 4th ed., Elsevier Science Pub. Co., Inc., N.Y.
3. International Organization for Standardization (ISO), ISO 7932:1993.
4. MacFaddin, 2000, Biochemical Tests for the Identification of Medical Bacteria, 3rd ed., Williams and Wilkins, Baltimore.
5. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore
6. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
7. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 06/2024



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



**In vitro diagnostic
medical device**



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



**Do not use if
package is damaged**

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.