

Technical Data

Motility Nitrate Medium, Buffered

Intended Use:

Recommended for isolation and detection of *Clostridium perfringens* on the basis of motility and nitrate test.

Composition**

Ingredients	Gms / Litre
Peptone	5.000
HM peptone B #	3.000
Galactose	5.000
Potassium nitrate	1.000
Disodium hydrogen phosphate	2.500
Agar	3.000
Final pH (at 25°C)	7.4 ± 0.2
**Formula adjusted, standardized to suit performance parameters	

Equivalent to Beef extract

Directions

Suspend 19.5 grams in 1000 ml purified / distilled water containing 5 ml glycerol. Heat to boiling to dissolve the medium completely. Dispense in test tubes to make them half full. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool quickly in cool running water and allow the tubed medium to solidify in an upright position.

Principle And Interpretation

Clostridium perfringens food poisoning is one of the most common type of human foodborne illness. The foods usually involved are cooked meat or poultry products containing large number of viable cells. *Clostridium perfringens* is a gram-positive, rod shaped anaerobic, spore-forming bacteria that produces enterotoxin. This toxin if ingested, can cause food poisoning. Motility Nitrate Medium, Buffered formulated in accordance with FDA (2) and APHA (6), is recommended for the detection of *C. perfringens* on the basis of motility and nitrate test.

Peptone and HM peptone B supply amino acids and other complex nitrogenous substances. Agar is added to obtain a semisolid gel that helps to demonstrate motility of the organism along the stab line of inoculation. Growth of motile organisms extends out from the line of inoculation. The medium contains 0.5% each of glycerol and galactose to improve the consistency of the nitrate reduction reaction with different strains of the organisms (3). Potassium nitrate serves as a base for nitrate reduction. A red or orange colour formation on addition of nitrate reagents indicates reduction of nitrate to nitrite.

Motility is indicated by turbidity extending out from the line of stab inoculation. Non-motile organisms grow only in the inoculated area. After 3-8 hours of incubation, a small puffball of motility may be seen around the line of inoculation. If this is not observed, tubes should be re-incubated for 24-48 hours and compared for turbidity to an un-inoculated tube. Negative motility reactions should be confirmed by a hanging drop preparation.

In the nitrate reduction test, a pink to red color develops after addition of the reagents if nitrite is present. Colour development indicates that nitrate reduction has occurred in the tube. Some organisms further reduce nitrite to ammonia that can be detected by the addition of a small amount of zinc dust to the tubes exhibiting no colour. A pink colour in this part of the test indicates no nitrate reduction. A colourless reaction indicates that nitrates have been completely reduced.

Inoculate 2 grams of food sample in 15 to 20 ml of Chopped Liver Broth (M606) or Tryptone Glucose Yeast Extract Broth (M952). After an incubation at 35-37°C for 20-24 hours, isolate on Perfringens Agar Base (TSC/SFP Agar Base) (M837). Presumptive *C.perfringens* colonies are confirmed biochemically by inoculating into Motility Nitrate Medium, Buffered to detect motility and nitrate reduction.

Type of specimen

Isolated Microorganism from food and water samples

Specimen Collection and Handling:

M630

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (1,6,7). For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(3) After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations:

1. Due to nutritional variations, some strains may show poor growth.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Semisolid, comparable with 0.3% Agar gel.

Colour and Clarity of prepared medium

Light amber coloured clear to slightly opalescent gel forms in tubes as butts

Reaction

Reaction of 1.95% w/v aqueous solution at 25°C. pH : 7.4±0.2

pН

7.20-7.60

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours .

Organism	Inoculum (CFU)	Growth	Motility	Nitrate reduction
Clostridium absonum ATCC 27555	50-100	luxuriant	weakly motile	weak or negative reaction
Clostridium perfringens ATCC 12924	50-100	luxuriant	negative, growth along the stabline, surrounding medium remains clear	positive, red- violet colour developed within 1-2 minutes

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-25°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,5).

Reference

- 1. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
- 2. Bacteriological Analytical Manual, Food and Drug Administration, 1995, 8th Ed., AOAC International, Gaithersburg, Md., USA.
- 3. Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- 4. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
- 5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
- 6. Salfinger Y., and Tortorello M.L., 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
- 7. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.

Revision :02/ 2019

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia[™] publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia[™] Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Reg.office : 23, Vadhani Ind.Est., LBS Marg, Mumbai-400086, India. Customer care No.: 022-6116 9797 Corporate office : A-516,Swastik Disha Business Park,Via Vadhani Ind. Est., LBS Marg, Mumbai-400086, India. Customer care No.: 022-6147 1919 Email: techhelp@himedialabs.com Website: www.himedialabs.com