



## Egg Yolk Agar Base

M808

### Intended Use:

Used for isolation and identification of *Clostridia* and certain other anaerobes.

### Composition\*\*

Ingredients	g/ L
Proteose peptone	40.000
Disodium hydrogen phosphate	5.000
Potassium dihydrogen phosphate	1.000
Sodium chloride	2.000
Magnesium sulphate	0.100
Dextrose (Glucose)	2.000
Hemin	0.005
Agar	25.000
Final pH ( at 25°C)	7.6±0.2

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

Suspend 75.10 grams in 900 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Dispense 90 ml amounts in tubes or flasks as desired. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C and add 10 ml of sterile egg yolk emulsion (FD045) per 90 ml of medium. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

*Clostridium perfringens* food poisoning is one of the most common types of human food borne illness (1). The foods usually involved are cooked meat or poultry products containing large numbers of viable cells. A heat-labile enterotoxin produced only by sporulating cells (2) induces the major symptoms of diarrhea in perfringens poisoning.

Egg Yolk Agar Base is a slight modification (3) of McClung Toabe Agar Base (4) used for isolation and detection of *Clostridium perfringens*. Egg Yolk Agar Base differs from the original formula by the inclusion of hemin.

Proteose peptone provide the essential nutrients along with carbonaceous and nitrogenous substances. Phosphates buffer the medium whereas sodium chloride maintains the osmotic equilibrium. Magnesium sulphate serves as a source of divalent cations along with sulphates. Glucose serves as a source of energy. Hemin help to enhance the growth of anaerobic organisms. Organisms producing lecithinase break down lecithin present in the egg yolk emulsion producing an insoluble opaque precipitate around the colonies. Lipase-producing organisms break down free fatty acids (in the egg yolk emulsion) forming an iridescent sheen on the surface of the colonies. Lipase activity may be delayed, therefore plates should not be discarded as negative before incubation for a week. Proteolytic activity is seen as clear zones around the colonies (5). The media should be directly inoculated with the test specimen.

### Type of specimen

Clinical- stool, abscess, etc.; Food samples.

### Specimen Collection and Handling

Prior to inoculation, media plates should be reduced by placing in an anaerobic jar for 18-24 hours. An enrichment broth should be simultaneously inoculated with the test sample to detect small number of anaerobic organisms. Standard procedures for the isolation of organism should be referred. Incubation should be carried out for 18-48 hours and continued for 7 days. After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

In Vitro diagnostic use. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

## Limitations

1. Lipase activity may be delayed, therefore plates should not be discarded as negative before incubation for a week.
2. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
3. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Cream to yellow homogeneous free flowing powder

### Gelling

Firm, comparable with 2.5% Agar gel.

### Colour and Clarity of prepared medium

Basal medium: Medium amber coloured, clear to slightly opalescent gel After addition of egg yolk emulsion (FD045): Yellow coloured opaque gel forms in Petri plates

### Reaction

Reaction of 7.5% w/v aqueous solution at 25°C. pH : 7.6±0.2

### pH

7.40-7.80

### Cultural Response

Cultural characteristics observed with added Egg yolk emulsion (FD045), after an incubation at 35-37°C for 48-72 hours when incubated anaerobically. (\*- Plates should be incubated up to 7 days before regarding them as negative)

Organism	Inoculum (CFU)	Growth	Recovery	Lecithinase	Lipase activity*	Proteolytic activity
<i>Bacteroides fragilis</i> ATCC 25285	50-100	good-luxuriant	≥50%	negative reaction	negative reaction, no iridescent sheen on the colony surface and medium	negative, no clear zone surrounding colonies
<i>Clostridium botulinum</i> ATCC 25763	50-100	good-luxuriant	≥50%	negative reaction	negative, no iridescent sheen on the colony surface and medium	positive, clear zone surrounding colonies
<i>Clostridium butyricum</i> ATCC 13732	50-100	good-luxuriant	≥50%	negative reaction	negative, no iridescent sheen on the colony surface and medium	positive, clear zone surrounding colonies
<i>Clostridium perfringens</i> ATCC 12924	50-100	good-luxuriant	≥50%	positive, opaque zone around the colony	negative, no iridescent sheen on the colony surface and medium	negative, no clear zone surrounding colonies
<i>Clostridium sporogenes</i> ATCC 11437	50-100	good-luxuriant	≥50%	negative reaction	positive, iridescent sheen on the colony surface and medium.	positive, clear zone surrounding colonies

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

## Reference

1. Labbe R., 1989, *Clostridium perfringens*, In Foodborne Bacterial Pathogens Ed., Doyle M. P., P.191, Marcel Dekker, New York, N.Y.
2. Duncan C. L., 1973, A. J. Bacteriol., 113:932
3. Atlas R. M., 2004, Handbook of Microbiological Media, 3rd Ed., CRC Press.
4. McClung and Toabe, 1947, J. Bacteriol., 53:139
5. Murray P. R., Baron J. H., Tenover F. C., Tenover J. C., (Ed.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C.
6. Isenberg (Ed.), 1992, Clinical Microbiology Procedures Handbook, American Society for Microbiology, Washington, D.C.
7. Jorgensen, J.H., Tenover, F.C., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision :06/2024




HiMedia Laboratories Pvt. Limited,  
Plot No.C-40, Road No.21Y,  
MIDC, Wagle Industrial Area,  
Thane (W) -400604, MS, India



*In vitro* diagnostic  
medical device



Storage temperature



CEpartner4U, Esdoornlaan 13,  
3951DB Maarn, NL  
[www.cepartner4u.eu](http://www.cepartner4u.eu)



CE Marking



Do not use if  
package is damaged

### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.