

Technical Data

Fermentation Medium for Neisseria

M825

Intended Use:

Used for studying fermentation reaction of *Neisseria* species.

Composition	
Ingredients	g/ L
Tryptone	20.000
L-Cystine	0.500
Sodium chloride	5.000
Sodium sulphite	0.500
Phenol red	0.017
Agar	3.500
Final pH (at 25°C)	7.5±0.1

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 29.52 grams in 1000 ml purified / distilled water. Heat to boiling to dissolve the medium completely. Dispense and sterilize by autoclaving at 118°C for 15 minutes. The pressure should not exceed 12lbs. Cool to around 40-45°C and add membrane filter sterilized sugar solutions to final concentration of 1%. (i.e. 5 ml of 20% sugar solution per 100 ml of medium).

Principle And Interpretation

Neisseria species are oxidative i.e. they produce acid from carbohydrate by oxidation. Because these species are oxidative and produce less acid from carbohydrates than do fermentative organisms and because they also produce ammonia from peptones which may neutralize any acid produced from carbohydrates, acid production is determined in a medium with a low protein/carbohydrate ratio and a sensitive indicator such as phenol red (1,2). Fermentation Medium for *Neisseriae* is recommended for studying the fermentation reactions of fastidious organisms such as *Neisseria* (3). This medium is the modification of the medium originally formulated by Vera (4).

Neisseria species oxidize the added carbohydrates to yield acids. The acids thus formed change the colour of the pH indicator, phenol red form orange to yellow. The organism also degrades the peptone source to yield ammonia. The alkalinity thus formed causes the phenol red to change to pink. However, if the acidity formed by carbohydrate metabolism is greater than the alkalinity formed by peptone degradation, the medium remains yellow in colour.

Tryptone supplies the necessary nitrogenous nutrients to the organisms. L-Cystine acts as an amino acid source as well as a reducing agent, which can remove (bind) molecular oxygen thereby preventing the accumulation of peroxides which are lethal to certain microorganisms (5). Small amount of agar in the medium reduces convection currents in the medium and hence contributes to maintaining anaerobic conditions in the depth of the tubes. Sodium chloride maintains the osmotic equilibrium in the medium. Phenol red is the pH indicator, which turns yellow at acidic pH. Observe the inoculated tubes after every 4 hours.

Development of yellow colour throughout the medium indicates that the carbohydrate has been oxidized leading to the production of acids. Development of pink colour indicates that carbohydrates have not been oxidized and only the peptones have been degraded. *Neisseria* species tend to produce acids from carbohydrate in the vicinity of inoculated (stab) area. If accompanying contaminating organisms are present the entire medium may turn yellow.

Type of specimen

Clinical samples - isolated organism from clinical sample.

Specimen Collection and Handling:

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (6,7). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

In Vitro diagnostic use only. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
Neisseria species should be further confirmed by gram staining and oxidase test.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to light pink homogeneous free flowing powder

Gelling

Semisolid, comparable with 0.35% Agar gel.

Colour and Clarity of prepared medium

Straw coloured, clear to slightly opalescent gel forms in tubes as butts

Reaction

Reaction of 2.95% w/v aqueous solution at 25°C. pH : 7.5±0.1

pН

7.40-7.60

Cultural Response

Cultural characteristics observed with added 1% dextrose after an incubation at 35-37°C for 18-24 hours.

Organism	Growth	Acid w/added dextrose	Motility
Escherichia coli ATCC 25922 (00013*)	luxuriant	positive reaction, yellow colour	positive,growth away from stabline causing turbidity
Neisseria gonorrhoeae ATCC 19424	luxuriant	positive reaction, yellow colour	negative, growth along the stabline, surrounding medium remains clear
Streptococcus pneumoniae ATCC 6303	luxuriant	positive reaction, yellow colour	negative,growth along the stabline, surrounding medium remains clear

Key: *Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

Please refer disclaimer Overleaf.

Reference

- 1. Knapp J. S., 1988, Clin. Microbiol., Rev. 1: 415-431.
- Murray P. R., Baron E. J., Jorgensen J. H., Pfaller M. A., Yolken R. H., (Eds.), 8th Ed., 2003, Manual of Clinical Microbiology, ASM, Washington, D.C.
- 3. Atlas R. M., 2004, Handbook of Microbiological Media, 3rd Ed., CRC Press.
- 4. Vera, 1948, J. Bacteriol., 55:531.
- 5. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol.1, Williams and Wilkins, Baltimore
- 6. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition
- 7. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
- 8. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.

Revision : 05/2024



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMediaTM publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMediaTM Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Corporate Office : Plot No.C-40, Road No.21Y, MIDC, Wagle Industrial Area, Thane (W) - 400604, India. Customer care No.: 022-6147 1919 Email: techhelp@himedialabs.com Website: www.himedialabs.com