



## Malt Extract Agar Base, Modified as per Thom and Church

M995

### Intended Use:

Recommended for isolation, detection and enumeration of yeasts and moulds.

### Composition\*\*

Ingredients	g / L
Peptone	0.780
Maltose	12.750
Dextrin	2.750
Agar	15.000
Final pH ( at 25°C)	4.7±0.2

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

Suspend 31.28 grams in 1000 ml purified/distilled water. Add 2.35 gm glycerol. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Avoid overheating. Cool to 45-50°C. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

Malt Extract medium is recommended for the isolation, detection and enumeration of yeasts and moulds. Malt Extract Agar has been used for many years for the detection of yeast and moulds in a wide variety of materials including dairy products and foods (1). The medium is also suitable for maintaining stock cultures of fungi.

Reddish (2) described a medium prepared from malt extract which was an acceptable substitute for wort. Following the formula of Reddish, Thom and Church (3) used Malt extract as a base from which they prepared the complete media. Peptone provide essential growth nutrients for the growth of fungi. Maltose and dextrin are the suitable carbohydrates for the growth of fungi. The low pH inhibits bacterial growth (4).

### Type of specimen

Clinical samples: skin scrapings / skin lesions, etc; Food samples

### Specimen Collection and Handling:

Streak the specimen as soon as possible after it is received in the laboratory. Consult appropriate references for information regarding the processing and inoculation of specimens (5). For isolation of fungi from potentially contaminated specimen, a selective medium should be inoculated along with the non-selective medium. Incubate the plates at 25 to 30°C with increased humidity for upto 7 days. Examine the plates for fungal colonies and for confirmation, perform biochemical test and serological diagnosis. After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

In Vitro diagnostic Use. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations

1. Examine the plates for fungal colonies and for confirmation, perform biochemical test and serological diagnosis.
2. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
3. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.

### Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature

## Quality Control

### Appearance

Cream to yellow homogeneous free flowing powder

### Gelling

Firm, comparable with 1.5% Agar gel

### Colour and Clarity of prepared medium

Yellow coloured clear to slightly opalescent gel forms in Petri plates

### Reaction

Reaction of 3.12% w/v aqueous solution at 25°C. pH : 4.7±0.2

### pH

4.50-4.90

### Cultural Response

Cultural characteristics observed after an incubation at 25 - 30°C for 40 - 48 hours

Organism	Inoculum (CFU)	Growth	Recovery
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50-100	good-luxuriant	
<i>Candida albicans</i> ATCC 10231 (00054*)	50-100	good-luxuriant	≥70%
<i>Saccharomyces cerevisiae</i> ATCC 9763 (00058*)	50-100	good-luxuriant	≥70%

Key: (\*) Corresponding WDCM numbers # - Formerly known as *Aspergillus niger*

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

## Reference

1. Salfinger Y., and Tortorello M.L., 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
2. Reddish, 1919, Abst. Bact., 3:6.
3. Thom and Church, 1926, The Aspergilli.
4. Lennett, Balows, Hausler and Shadomy (Eds.), 1985, Manual of Clinical Microbiology, 4th ed., ASM, Washington, D.C.
5. Ajello L., Georg L. K., Kaplan W. and Kaufman L., 1963, CDC Laboratory Manual for Medical Mycology, Washington, D. C.
6. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
7. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 05/2024



HiMedia Laboratories Pvt. Limited,  
Plot No.C-40, Road No.21Y,  
MIDC, Wagle Industrial Area,  
Thane (W) -400604, MS, India



**In vitro diagnostic  
medical device**



**Storage temperature**



CEpartner4U, Esdoornlaan 13,  
3951DB Maarn, NL  
[www.cepartner4u.eu](http://www.cepartner4u.eu)



**CE Marking**



**Do not use if  
package is damaged**

**Disclaimer :**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.