

MB581PC16200 HiPurA® Pre- filled Cartridges for Shrimp DNA Purification

Kit Contents

Product Code	Reagents provided	MB581PC16200
		48 PR
PF16J1	Pre-filled Cartridges for Shrimp DNA Extraction	48 no
LA1118B	Magnetic Rod Tip	12 no
DS0015	Lysis Solution (AL)	39 ml
DS0013	Proteinase K	1 ml
DS0040	Elution Buffer (ET) [10 mM Tris-Cl, pH 8.5]	12 ml
DS1005A	Magnetic Beads	1 ml

Intended Use

Recommended for isolation of DNA from Shrimp sample.

Introduction

Shrimp is an important commodity in international trade accounting for 15% in terms of value of internationally traded seafood products. Aquaculture contributes to over 50% of global shrimp production. Shrimp production by aquaculture has been seriously impacted by diseases. The major constraints faced by shrimp aquaculture is the loss due to viral and bacterial diseases like white spot syndrome, yellow head disease, Taura syndrome, *Vibrio* spp., etc. HiMedia's HiPurA® Pre- filled Cartridges for Shrimp DNA Purification provides a fast and easy method for purification of total DNA for reliable applications in PCR technique etc. The DNA obtained is compatible with downstream applications such as restriction enzyme digestion, PCR and Southern blotting.

HiPurA® Pre- filled Cartridges for Shrimp DNA Purification

This kit simplifies isolation of DNA from shrimp tissue samples with magnetic based procedure. Shrimp tissue (spliced and digested) is subjected to lysis by Proteinase K in a chaotropic salt solution. Following lysis, is the adsorption of nucleic acid to the magnetic beads, removal of residual contaminants and elution of pure nucleic acid. The magnetic beads have a high binding capacity and high-quality DNA is obtained from shrimp samples.

Elution

The yield of genomic DNA depends on the sample type and the number of cells in the sample. A single elution with 100 µL of Elution Buffer (ET) will provide sufficient DNA to carry out multiple amplification reaction. Elution with volume less than 100 µL will increase the final DNA concentration, but will reduce the overall DNA yield.

Storage

HiPurA® Pre- filled Cartridges for Shrimp DNA Purification can be stored at room temperature (15-25°C) for up to 2 years without showing any reduction in performance. We advise a certain storage temperature for the reagents listed below:

- **On receipt store Proteinase K (DS0013): at -20°C.**
- **On receipt store Magnetic Beads (DS1005A): at 2-8°C.**

Materials needed but not provided

- Tabletop Microcentrifuge (with rotor for 2.0 ml tubes)
- Chilled Ethanol (96-100%)
- Ethanol (70%)
- Insta NX® Mag16 (Product Code: LA1118)
- Cartridge Holder (LA1118CH)
- Insta NX® Mag32 (LA1096)
- Vortex
- Polypropylene sealing film (Product Code: PR21)
- HiPer® Lock Microcentrifuge Tube, 2.0ml (Product Code: MBLA017)
- Sterile plastic sachets and pestle (for tissue homogenization)

General Preparation Instructions

1. Thoroughly mix reagents

Examine the reagents for precipitation. If any kit reagent forms a precipitate (other than enzymes), warm at 55-65°C until the precipitate dissolves and allow cooling to room temperature (15-25°C) before use.

2. Ensure that clean & dry Nuclease-free tubes and tips are used for the procedure.

Procedure

1. Prepare tissue

Weigh 100mg of fresh or frozen tissue and mince quickly. If frozen tissue is used, allow it to thaw slightly before slicing but keep on ice in order to protect degradation. Cut the tissue into small pieces as it enables more efficient lysis.

NOTE: Tissue can be harvested by aliquoting in 2.0 ml collection tubes (not provided) and flash freezing in liquid nitrogen; these can be stored at -70°C for several months before preparing DNA.

2. Digest tissue

In a sterile plastic sachet, add 800µL of Lysis Solution (AL) (DS0015) and 20µL of Proteinase K solution (DS0013) to the tissue. Homogenize the tissue by grinding it with a pestle. Incubate the sample at room temperature (15- 25°C) for 10 minutes.

3. Transfer the homogenized mixture to a 2 ml micro-centrifuge tube (not provided) and centrifuge at 5,000rpm for 5 minutes at room temperature (15- 25°C).
4. Transfer 150µL of supernatant to fresh tube and add 450 µL of 100% chilled ethanol (not provided). Mix well by vortexing and centrifuge at 10,000 rpm for 5 minutes at 4°C.

5. **Wash**

Discard the supernatant and re-suspend the pellet with 400µl of 70% ethanol (not provided) and centrifuge at 13,000 rpm for 5 minutes at 4°C.

6. Discard the supernatant and dry the pellet. Resuspend the DNA pellet in 100µl of Elution Buffer (ET) (DS0040). **This will be pre-processed sample.**

Set up processing plates as follows:

1. Switch on the UV light for 10 minutes prior to use.
2. Open the door of Insta NX® Mag16 machine.
3. Select “**MB581Shrimp**” program.
4. Remove the seal from the **Pre-filled Cartridges for Shrimp DNA Extraction (PF16J1)**. Place the **Pre-filled Cartridges for Shrimp DNA Extraction (PF16J1)** into the Cartridge Holder (LA1118CH).

NOTE: Take care while peeling off the seal. Hold the plate firmly by one hand and then peel off the seal by holding it in your other hand without shaking the plate.

5. Add 100µl of Elution Buffer (ET) [10mM Tris-Cl, pH8.5] (DS0040) into the 6th well of the **Pre-filled Cartridges for Shrimp DNA purification (PF16J1)**.
6. Add 100µl pre- processed sample in the 1st well of the **Pre-filled Cartridges for Shrimp DNA Extraction (PF16J1)**.
7. **Add 20µl Magnetic Beads (DS1005A) in the 1st well of the Pre-filled Cartridges for Shrimp DNA Extraction (PF16J1)**. Place the Cartridge Holder (LA1118CH) along with cartridges on the platform.
8. Place the Magnetic Rods Tip (LA1118B) by sliding onto the machine.

NOTE: After placing the rods ensure that the rods are properly fixed on their place.

9. Close the door of Insta NX® Mag16 machine.
10. Click on the **RUN** option on the home screen.
1. After the run is complete, remove Pre-filled Cartridges for Shrimp DNA Extraction (PF16J1) from the position. Discard the Magnetic rod’s tip (LA1118B). Dispense the eluted DNA from well 6 to a new HiPer® Lock Microcentrifuge Tube, 2.0ml (Product Code: MBLA017) (not provided). The eluate contains pure DNA.

NOTE: If small amount of magnetic beads are observed in the final eluate then keep the cartridges along with cartridge holder on Magnetic pad (not provided) for 4-5 minute and collect supernatant carefully without disturbing beads pellet in new collection tube.

OR

Take out eluate in new collection tube and centrifuge at higher speed for around 1 min to pellet down the traces of Magnetic beads present in the eluate.

***NOTE: If you process less than 4 samples at a time please order LA1118B- Magnetic Rod Tip (Pack size- LA1118B-4no/ LA1118B-40no).**

Storage of the eluate with purified nucleic acid: The eluate contains pure genomic DNA. For short-term storage (24-48 hrs) of the DNA, 2-8°C is recommended. For long-term storage, -20°C or lower temperature (-80°C) is recommended. Avoid repeated freezing and thawing of the sample which may cause denaturing of DNA. The Elution Buffer will help to stabilize the DNA at these temperatures.

Warning

Not for Medicinal Use. Read the procedure carefully before beginning the protocol. Wear protective gloves/protective clothing/eye protection/face protection. Follow good clinical laboratory practices while handling clinical samples. Standard precautions should be followed as per established guidelines. Safety guidelines may be referred in safety data sheets of the product.

Limitations

The yield of nucleic acid depends upon the type and the volume of starting material used.

Performance and Evaluation

The yield and efficiency of purification is determined by performing Real- Time PCR.

Quality Control

Each lot of HiMedia's HiPurA® Pre- filled Cartridges for Shrimp DNA Purification is tested against predetermined specifications to ensure consistent product quality.

Warning

Not for Medicinal Use. Read the procedure carefully before beginning the protocol. Wear protective gloves/protective clothing/eye protection/face protection. Follow good clinical laboratory practices while handling clinical samples. Standard precautions should be followed as per established guidelines. Safety guidelines may be referred in safety data sheets of the product.

Safety Information

The HiPurA® Pre- filled Cartridges for Shrimp DNA Purification is for laboratory use only, not for drug, household or other uses. Take appropriate laboratory safety measures and wear gloves when handling. Not compatible with disinfecting agents containing bleach. Please refer the Safety Data Sheet (SDS) for information regarding hazards and safe handling practices.

Disposal









User must ensure safe disposal by autoclaving and/ or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Technical Assistance

At HiMedia, we pride ourselves on the quality and availability of our technical support. For any kind of technical assistance mail to mb@himedialabs.com.

Please refer disclaimer Overleaf.

Symbols

	Manufacturer		Do not use if package is damaged
	Batch code		Temperature limit
	Date of manufacture (YYYY-MM)		Consult instructions for use
	Use-by date (YYYY-MM)		Catalogue number

Identification No.: PIMB581PC16

Rev. No.: 01

Date of Issue: 2025-04

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Reg.office : Plot No. C-40, Road No. 21Y, MIDC, Wagle Industrial Estate, Thane, (West) 400604, Maharashtra, INDIA.
Customer Care No.: 00-91-22-6116 9797 Tel: 00-91-22-6147 1919, 6903 4800 Email: techhelp@himedialabs.com Website: www.himedialabs.com