

MBPCR226

Hi-PCR® Extended-spectrum β -lactamases (ESBLs) Gene Quantification Probe PCR Kit

Description

In Gram-negative bacteria, the production of extended spectrum β -lactamases (ESBLs) represents the most important contributing factor to resistance against β -lactam antibiotics. These ESBLs have the ability to hydrolyze various types of the newer β -lactam antibiotics, including extended-spectrum cephalosporins of the 3rd and 4th generation (e.g. cefotaxime, ceftriaxone, ceftazidime) and monobactams (e.g. aztreonam), which were assessed as “critically important antimicrobials” by the World Health Organization (WHO). This has resulted in increased morbidity, mortality, higher cost of treatment and longer hospital stay. The ESBL producing Enterobacteriaceae have been marked as a “serious” threat by Centre for Disease Control (CDC). Currently, the predominant ESBL gene families encountered are CTX-M, SHV and OXA-10/11. Nucleic acid amplification-based assays or Polymerase Chain Reaction (PCR) is an alternative method that allows for sensitive and specific detection of ESBL genes i.e. CTX-M, SHV and OXA-10/11 genes from clinical samples / cultured isolates. Real-Time PCR technique is considerably simple and fast with respect to the standard PCR technique. This technique has been successfully used for the rapid detection and identification of a variety of infectious and non-infectious pathogens and genes.

NOTE: HiMedia’s Hi-PCR® Extended-spectrum β -lactamases (ESBLs) Gene Quantification Probe PCR Kit is for *in-vitro* use only.

Intended Use

Recommended for sensitive and specific detection with quantification of SHV, CTX-M and OXA-10/11 families in clinical samples and cultures.

Principle

Real-time polymerase chain reaction, also called quantitative Polymerase Chain Reaction (qPCR) or kinetic Polymerase Chain Reaction, is a laboratory technique based on the principle of PCR. This technique is used to amplify a targeted DNA sequence by use of hydrolysis probes that are short oligonucleotides that have a fluorescent reporter dye attached to the 5' end and a quencher dye to the 3' end. HiMedia’s Hi-PCR® Extended-spectrum β -lactamases (ESBLs) Gene Quantification Probe PCR Kit is designed to detect the specific regions of the genes encoding the ESBLs wherein **SHV, CTX-M and OXA-10/11 is detected in FAM, HEX and Cy5 channels** respectively with **Internal Control in Cy5.5 channel** in a single tube reaction. The kit allows sensitive and specific detection with quantification of single and co-present ESBL encoding genes in a single tube reaction.


Positive Control

This is a control reaction using a known template (target pathogen). A positive control is usually used to check that the primers have been designed properly and the PCR conditions have been set up correctly.

Internal Control

This is a control sequence which is amplified in the same reaction tube along with the target sequence (target species) but detected with a different primer (i.e. Multiplex PCR). An internal control is often used to detect the failure of amplification in cases where the target sequence is not amplified.

Diagrammatic representation of preferential binding of probe specific to DNA fragments in Real-Time PCR



Polymerization: A fluorescent reporter (R) dye and a quencher (Q) are attached to the 5' and 3' end of the probe respectively

Strand displacement: When the probe is intact, the report dye emission is quenched.

Cleavage: During each extension cycle, the DNA polymerase cleaves the reporter dye from the probe

Polymerization completed: Once separated from the quencher, the reporter dye emits its characteristic fluorescence

While the probe is intact, the proximity of the quencher dye greatly reduces the fluorescence emitted by the reporter dye by fluorescence resonance energy transfer (FRET). The probes are designed such that they anneal within a DNA region amplified by a specific set of primers. During PCR amplification, these probes will hybridize to the target sequences located in the amplicon i.e. the DNA. As the *Taq* DNA polymerase replicates the template with the bound probe, the 5'-nuclease activity of the polymerase enzyme cleaves the fluorescent probe. The end result in cleavage of the probe is separation of the reporter dye from the quencher dye and increasing the reporter dye signal. As the probe is removed from the target strand, primer extension continues to the end of the template strand. Hence, fluorescence detected in the quantitative PCR thermal cycler is directly proportional to the fluorophore released and the amount of DNA template present in the PCR. Thus, inclusion of the probe does not inhibit the overall PCR process.

Features

- Fast and simple
- Good sensitivity and specific results
- Guaranteed reproducible results
- Rapid detection of all relevant clinical pathogens

Standards

The kit is supplied with 5 standards of defined copies that can be used as reference for calculating the copy number of the unknown.

Sample Source: Clinical samples / cultured isolates

Storage and Shelf life

The provided kit has a shelf-life of 12 months when stored at -20°C. Repeated thawing and freezing of PCR reagents should be avoided, as this may reduce the sensitivity. If the reagents are to be used multiple times, we recommend storing reagents as aliquots to avoid repeated freeze and thaw. Degradation of sample DNA specimens can also reduce the sensitivity of the assay. HiMedia Laboratories does not recommend using the kit after the expiry date stated on pack.

Kit Contents: The provided PCR kit contains:

Components	Product code	Reagents provided for 25R* (µL)	Reagents provided for 50R* (µL)
2X Super Mastermix	DS0900	338	675
ESBL Primer-Probe Mix	DS1226	81	162
Internal Control Primer-Probe Mix	DS1227	27	54
Internal Control DNA	DS1096	27	54
Molecular Biology Grade Water for PCR	ML065	75	150
ESBL Positive Control Std. 1	DS0967	13	25
ESBL Positive Control Std. 2	DS0968	13	25
ESBL Positive Control Std. 3	DS0969	13	25
ESBL Positive Control Std. 4	DS0970	13	25
ESBL Positive Control Std. 5	DS0971	13	25

* For a 25 µL PCR reaction

Specimen collection and Handling

Follow appropriate techniques for handling specimens; after use, contaminated materials must be sterilized by autoclaving before discarding. Standard precautions as per established guidelines should be followed while handling clinical specimens and items contaminated with blood and other body fluids. Safety guidelines may be referred in individual safety data sheets.

Sample Preparation

Various clinical samples / cultured isolates are routinely examined. For extraction and purification of pure DNA for high yield, perform the nucleic acid purification using HiPurA® Multi-Sample DNA Purification Kit (MB554) for clinical samples and HiPurA® Bacterial Genomic DNA Purification Kit (MB505) for cultured isolates as instructed in the protocol.

Materials needed but not provided

- PCR tubes (Product code PW1255) or PCR Strips (Product code: PR17) or PCR Plates (Product code: PR2 / PR3 / PR19) & Sealing film (PR18)
- Insta Q Real Time PCR System (Product Code: LA1073 / LA1074)
- Barrier Micropipette Tips
- Micropipettes
- For clinical samples: HiPurA® Multi-Sample DNA Purification Kit (MB554)
- For cultured isolates: HiPurA® Bacterial Genomic DNA Purification Kit (MB505)

General Preparation Instructions

- Before use all PCR components should be completely thawed on ice (4°C).
- Perform the amplification reactions in a clean area, preferably in a biosafety cabinet.
- Use of aerosol barrier pipette tips is recommended to reduce contamination risks from extraneous DNA templates.
- Extract and store positive control sample (if used) separately from all other reagents to avoid contamination and add it to the reaction mix in a separate area.

A. Protocol for PCR Master Mix Preparation


Perform PCR reactions for each DNA sample as per the following table:

Components	Product Code	Volume to be added for 1R (for a 25 µL reaction)
2X Super Mastermix	DS0900	12.5 µL
ESBL Primer-Probe Mix	DS1226	3 µL
Internal Control Primer-Probe Mix	DS1227	1 µL
Internal Control DNA	DS1096	1 µL
Molecular Biology Grade Water for PCR	ML065	2.5 µL
Standard DNA / Negative Control / Template DNA	-	5 µL
Total volume	-	25 µL

Centrifuge the tube briefly at 6000 rpm for about 10 seconds. Place the tubes in Real-Time PCR machine and set the recommended PCR program (mentioned below). Interpret the data from the amplification plot (observe the Ct values).

B. Setting up an experiment (InstaQ series)

1. Click on Absolute module



2. Write the name of experiment and / or Username.
3. By default, FAM detector is added. Add the remaining, detectors i.e. FAM, HEX, Cy5 and Cy5.5. Name the detector for the gene to be tested (target gene).
4. Next go on to samples, add the 5 standards and sample IDs respectively to be tested.
5. Setup the plate design based on the placement of the tubes and select the detector appropriately.
6. Select the wells containing Standards. Change the property to Standard [S] and in the adjoining blank enter the values for the respective Standards and press ENTER.

Standard	Concentration (ng)	Number of copies
Standard 1	20	1 x 10 ⁶
Standard 2	2	1 x 10 ⁵
Standard 3	0.2	1 x 10 ⁴
Standard 4	0.02	1 x 10 ³
Standard 5	0.002	1 x 10 ²

NOTE: For Absolute concentration, a standard curve is essential in every run

C. Recommended PCR program

- | | | |
|----------------------------|------------------------------------|------|
| 1. Initial denaturation | : 95°C for 10 minutes | } 45 |
| 2. Denaturation | : 95°C for 15 seconds | |
| 3. Annealing and Extension | : 60°C for 30 seconds (Plate Read) | |
| Sampling | : FAM/HEX/Cy5/Cy5.5 | |
| 4. Hold | : 4°C for ∞ | |

Quality Control

Each lot of HiMedia's Hi-PCR® Extended-spectrum β-lactamases (ESBLs) Gene Quantification Probe PCR Kit is assayed for contaminating endonuclease, exonuclease and non-specific DNase activities. Functionally tested in DNA amplification.

Data Analysis

The following conditions should be met for a valid diagnostic test:

Control	Detection channel			
	FAM (SHV)	HEX (CTX-M)	Cy5 (OXA-10/11)	Cy5.5 (IC)
Positive Control	+	+	+	+
Negative Control	-	-	-	+


Data Interpretation

Target			Result Interpretation
SHV (FAM)	CTX-M (HEX)	OXA-10/11 (Cy5)	
Ct value ≤ 33	Ct value ≤ 33	Ct value ≤ 31	Positive for SHV, CTX-M and OXA-10/11
Ct value >34	Ct value >34	Ct value >32	Negative for SHV, CTX-M and OXA-10/11
No Ct	No Ct	No Ct	PCR inhibition or reagent failure. Repeat PCR or repeat extraction from original sample

*The presence or absence of a signal in the Cy5.5 channel is not relevant for the validity of the test run due to competition between the test template and Internal Control template.

Standards used

Standard	Concentration (ng)	Number of copies
Standard 1	20	1 x 10 ⁶
Standard 2	2	1 x 10 ⁵
Standard 3	0.2	1 x 10 ⁴
Standard 4	0.02	1 x 10 ³
Standard 5	0.002	1 x 10 ²



Warning

Not for Medicinal Use.

Precautions

Read the procedure carefully before beginning the protocol. Wear protective gloves/protective clothing/eye protection/face protection. Follow good clinical laboratory practices while handling clinical samples. Standard precautions should be followed as per established guidelines. Safety guidelines may be referred in safety data sheets of the product.

Performance and Evaluation

Each lot of HiMedia's Hi-PCR® Extended-spectrum β -lactamases (ESBLs) Gene Quantification Probe PCR Kit is tested against predetermined specifications to ensure consistent product quality.

Troubleshooting Guide

Sr. No.	Problem	Cause	Solution
1.	No amplification	Degraded samples	1. Check the integrity of DNA using agarose gel electrophoresis. 2. Use freshly prepared DNA to ensure the availability of intact template sequence for efficient amplification.
		Error in protocol setup	Verify that the correct reagent volumes, dilutions and storage conditions have been used.
2.	Variability between replicates	Error in reaction set-up	Prepare a large volume master mix, vortex thoroughly and aliquot into reaction tubes.
		Air bubbles in reaction mix	Briefly centrifuge reaction samples/plate prior to running on a Real-Time PCR instrument.
		Pipetting error	C _t values of replicates can show increased variation due to poor laboratory technique or imprecise pipettes.
3.	Amplification in negative control	Reagents contaminated	1. Replace all critical solutions. 2. Repeat the analysis of all tests with fresh aliquots of critical reagents.
4.	No signal with positive controls	Incorrect programming of the temperature profile of the thermocycler	Compare the temperature profile to the manual.

Safety Information


HiMedia's Hi-PCR® Extended-spectrum β -lactamases (ESBLs) Gene Quantification Probe PCR Kit is for laboratory use only, not for drug, household or other uses. Take appropriate laboratory safety measures and wear gloves when handling.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Technical Assistance

At HiMedia, we pride ourselves on the quality and availability of our technical support. For any kind of technical assistance, mail at mb@himedialabs.com.



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Private Limited,
Reg. Off: Plot No. C-40, Road No. 21Y,
MIDC, Wagle Industrial Estate, Thane,
(West) 400604, Maharashtra, INDIA.
Web: www.himedialabs.com



04/2027

PIMBPCR226_O/0424

MBPCR226-04

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

HiMedia Laboratories Pvt. Ltd. Reg.office : Plot No. C-40, Road No. 21Y, MIDC, Wagle Industrial Estate, Thane, (West) 400604, Maharashtra, INDIA. Customer Care No.: 00-91-22-6116 9797 Tel: 00-91-22-6147 1919, 6903 4800 Email: mb@himedialabs.com Website: www.himedialabs.com