



Sabouraud Dextrose Medium (without Membrane Filter)

MF003

For yeast and mould detection and enumeration.

Composition**

Ingredients	Gms / Litre
Dextrose	20.000
Peptone, special	10.000

**Formula adjusted, standardized to suit performance parameters

Directions

The test sample should be filtered through a sterile membrane filter having pore size of 0.22 μ / 0.45 μ . Rehydrate the nutrient pad with 2.0 - 2.5 ml sterile distilled / purified water. After filtration, remove the membrane filter aseptically using sterile forceps. Place the membrane filter on rehydrated nutrient pad. Incubate the inoculated nutrient. Interpret the results qualitatively by observing the presence or absence of growth and quantitatively by counting the number of colonies on the surface of the membrane filter and calculating CFU/ml.

Principle And Interpretation

Field of Application: Waste water, pharmaceuticals, cosmetics, packing material. DriFilter Membrane Nutrient Pad Medium is ready to use sterile culture media in the form of a 50 mm biological inert absorbent pads impregnated with Sabouraud Dextrose medium, then dried and sterilized in 55 mm petri plate. They eliminate the need of laborious media preparation and autoclaving procedures. The nutrient pads are to be just rewetted with sterile distilled water and are ready to use. Use of nutrient pads allows larger sample volumes to be tested at a time. Interpretation of results is directly by counting the CFUs and also quantifies the microbial load present in the sample. Sabouraud Dextrose Medium is Carliers modification (1) of the formulation described by Sabouraud (2) for the cultivation of fungi (yeasts, moulds), particularly useful for the fungi associated with skin infections. This medium is also employed to determine microbial contamination in food, cosmetics, and clinical specimens (3). Mycological peptone provides nitrogenous compounds. Dextrose provides an energy source. High dextrose concentration and low pH favours fungal growth and inhibits contaminating bacteria from test samples (4). Some pathogenic fungi may produce infective spores which are easily dispersed in air, so examination should be carried out in safety cabinet. For heavily contaminated samples, the plate must be supplemented with inhibitory agents for inhibiting bacterial growth with lower pH.

Quality Control

Appearance

Dry filter membrane pad of 50mm diameter

Colour

Pale coloured nutrient pad

Sterility test

Passes release criteria

Cultural response

Cultural characteristics observed after incubation at 35-37°C for 18-48 hours.

Organism	Growth	Colour of colony
<i>Sacchromyces cerevisiae</i> ATCC 9763	Luxuriant	Colourless
<i>Candida albicans</i> ATCC 10231	Luxuriant	Colourless

Storage and Shelf Life

Please refer disclaimer Overleaf.

Store between 10-30°C. Use before expiry date on the label.

Reference

1. Carlier G. I. M., 1948, Brit. J. Derm. Syph., 60:61. 2. Sabouraud K., 1892, Ann. Dermatol. Syphilol, 3:1061. 3. Bacteriological Analytical Manual, 8th Edition, Revision A, 1998. AOAC, Washington D.C. 4. Murray PR, Baren EJ, Jorgensen JH, Pfaller MA, Tenover FC, Tenover MC (editors) 2003, Manual of clinical Microbiology, 8th ed., ASM, Washington, D.C.



Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.