

Soyabean Casein Digest Medium, Sterile powder

MH011G

Intended Use:

It is λ irradiated sterile powder recommended for evaluation of sterility in manufacturing process.

Composition**

Ingredients	Gms / Litre
Tryptone #	17.000
Soya peptone ##	3.000
Sodium chloride	5.000
Dipotassium hydrogen phosphate	2.500
Glucose monohydrate	2.500
Final pH (at 25°C)	7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Pancreatic digest of casein ,## Papaic digest of soybean (soyabean)

Directions

Sterile powder can be used directly for the evaluation of sterility in manufacturing process. For sterile liquid medium aseptically add 29.77 grams (the equivalent weight of dehydrated medium per litre) in 1000 ml sterile distilled water. Heat if necessary to dissolve the medium completely. Excessive heating is detrimental. Dispense aseptically in sterile flasks or tubes as desired. Recommended **NOT TO AUTOCLAVE OR OVERHEAT**.

Principle And Interpretation

Soyabean Casein Digest Medium is recommended as a sterility testing medium in accordance with the harmonized method of USP/EP/BP/JP/IP (1,2,3,4,5). It is used for the sensitivity testing of antimicrobial agents by the tube dilution method (6). It is also employed in diagnostic research in microbiology. This medium is used as a diluent and suspending medium for preparation of samples or test strains. It is also employed in sample preparation for testing of products, wherein incubation is carried out, only to serve sufficient resuscitation of the cell, while avoiding multiplication of the organism. The combination of tryptone and soya peptone makes this medium nutritious by providing amino acids and long chain peptides for the growth of microorganisms. Natural sugars in soybean promote growth of fastidious organism. Glucose is the fermentable source of carbon and dibasic potassium phosphate serves as the buffer in the medium. Sodium chloride maintains the osmotic balance of the medium.

This medium is recommended for sterility checking and for studying total aerobic microbial count in verification of microbiological testing procedures employed for sterility checking.

Type of specimen

Pharmaceutical samples

Specimen Collection and Handling

For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per pharmaceutical guidelines (1,2,3,4,5). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Biochemical characterization is necessary to be performed on colonies from pure cultures for further identification.
2. This medium is general purpose medium and may not support the growth of fastidious organisms.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light yellow coloured clear solution without any precipitate.

Reaction

pH of 2.98% w/v aqueous solution at 25°C (after sterilization). pH : 7.3±0.2

pH

7.10-7.50

Sterility Testing

No growth is observed after 14 days for Bacteria at 30-35°C and for fungi at 20-25°C.

Test for Mycoplasma

None detected.

Stability test

Light yellow coloured clear solution without any precipitation or sedimentation at room temperature for 7 days.

Growth Promotion Test

In accordance with the harmonized method of USP/EP/BP/IP.

Growth promoting properties

Clearly visible growth of microorganism comparable to that previously obtained with previously tested and approved lot of medium occurs at the specified temperature for not more than the shortest period of time specified inoculating ≤100 cfu (at 30-35°C for 18-24 hours).

Sterility Testing + Validation

The medium is tested with suitable strains of microorganisms inoculating ≤100cfu and incubating at 20-25°C for not more than 3 days in case of bacteria and not more than 5 days in case of fungi.

Cultural Response

Organism	Inoculum (CFU)	Growth	Incubation period	Incubation temperature
Growth promoting				
<i>Salmonella</i> Abony NCTC 6017 (00029*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Escherichia coli</i> NCTC 9002	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (00003*)	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Micrococcus luteus</i> ATCC 9341	50 -100	luxuriant	18 -24 hrs	30 -35 °C
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	luxuriant	18 -24 hrs	30-35 °C
<i>Streptococcus pneumoniae</i> ATCC 6305	50 -100	luxuriant	18-24 hrs	30-35 °C
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50 -100	luxuriant	18-24 hrs	30-35 °C
Sterility Testing- Growth promotion+Validation				
<i>Streptococcus pneumoniae</i> ATCC 6305	50 -100	luxuriant	≤3 d	20 -25 °C
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	50 -100	luxuriant	≤3 d	20 -25 °C

<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Escherichia coli</i> NCTC 9002	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (00003*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Micrococcus luteus</i> ATCC 9341	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	luxuriant	<=3 d	20 -25 °C
<i>Salmonella</i> Abony NCTC 6017 (00029*)	50 -100	luxuriant	<=3 d	20 -25 °C
* <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50 -100	luxuriant	<=5 d	20 -25 °C
<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	luxuriant	<=5 d	20 -25 °C
<i>Candida albicans</i> ATCC 2091 (00055*)	50 -100	luxuriant	<=5 d	20 -25 °C

Key : (#) Formerly known as *Aspergillus niger*, (*) Corresponding WDCM numbers

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-25°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference

1. The United States Pharmacopoeia-National Formulary (USP-NF), 2022.
2. European Pharmacopoeia, 2022, 10 th volume, European Directorate for the quality of medicines & Healthcare.
3. The British Pharmacopoeia, 2022, Medicines and Healthcare products Regulatory Agency.
4. The Japanese Pharmacopoeia, 17th edition, 2016, The Ministry of Health, Labour and welfare.
5. Indian Pharmacopoeia, 2022, Indian Pharmacopoeia Commission, Ministry of Health and Family Welfare Government of India.
6. Wright and Welch, 1959-60, Antibiotics Ann., 61.
7. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
8. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock.,D.W.(2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 04/2022

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.