Silver Staining Kit

Product Name	Product Code	Kit Packing
Silver Staining Kit	ML123-1KT	1 kit

Introduction: In the mid-17th century it was discovered that silver has ability to develop images. This property of silver was exploited in the development of photography and in several histological procedures. Silver staining for detection of proteins following gel electrophoresis was first reported in 1979 by Switzer et al. resulting in a major increase in the sensitivity of protein detection. This staining method is generally accepted to be several times more sensitive than methods using Coomassie Brilliant Blue R-250.

All silver staining methods depend upon the reduction of ionic silver to its metallic form. For increase in sensitivity the gel is treated with reducing or sulfiding agents like Glutaraldehyde, Formaldehyde and Sodium thiosulphate which has affinity for proteins prior to silver impregnation. Glutaraldehyde forms a Schiff's base with the proteins which results in the cross linking of amine groups and aldehydes which reduces silver during development. Protein bands can be visualized on the gel due to the difference in oxidation/reduction potential between the sites occupied by protein and the adjacent sites not occupied by the protein on the gel. The silver cations complex with protein amino groups, particularly the amino group of lysine, and with sulphur residues of cysteine and methionine. The gel is impregnated with silver nitrate in a weakly acidic (pH 6) solution. Any free silver nitrate is washed out of the gel prior to development, as precipitation of silver oxide will result in high background staining. Development is subsequently achieved by selective reduction of ionic silver by formaldehyde made alkaline by sodium carbonate. Sodium carbonate acts as a buffer to the formic acid produced by the oxidation of formaldehyde, which results in reduction of silver ion till protein bands appear in the gel.

HiMedia's Silver Staining Kit is based upon the method developed by M. Swain and N.W. Ross in 1995 which is highly sensitive for the detection of proteins on a clear background.

Kit Contents

Product Code	Reagents	Quantity	Storage
DS0105	1.2% Glutaraldehyde	100 ml	2 – 8°C
DS0106	0.5% Sodium thiosulphate	100 ml	2 – 8°C
DS0107	2.5% Silver nitrate	100 ml	2 – 8°C
DS0108	0.19% Formaldehyde	2 X 100 ml	2 – 8°C
DS0109	Sodium carbonate	50 g	2 – 8°C

ISO 9001-2015 CERTIFIED
WHO GMP CERTIFIED
ISO 13485-2016 CERTIFIED

Registered Office
HiMedia Laboratories Pvt Ltd.
Plot No. C-40, Road No. 21Y, MIDC, Wagle Industrial Area,
Thane, (West) 400604, Maharashtra, INDIA.
Customer Care No.: 00-91-22-6116 9797
Tel : 00-91-22-6147 1919, 6903 4800
Fax : 6147 1920
Web : www.himedialabs.com
Email : info@himedialabs.com
mb@himedialabs.com

The information contained herein is believed to be accurate and complete. However no warranty or guarantee whatsoever is made or is to be implied with respect to such information or with respect to any product, method or apparatus referred to herein

Materials needed but not provided

- Ethanol
- Molecular Biology Grade Water (Product Code: ML024)
- Acetic acid (MB052)

Preparation of Reagents (100 ml)

- Fixing Solution: Ethanol……40 ml
  Acetic acid……..10 ml
  Distilled water…….50 ml
- Fix/Sensitizing Solution: Ethanol……..40 ml
  1.2% Glutaraldehyde……..4 ml
  0.19% Formaldehyde……..2 ml
  Distilled water…….54 ml
- Ethanol Rinsing Solution: Ethanol……..40 ml
  Distilled water……..60 ml
- Sensitizing Solution: 0.5% Sodium thiosulphate….4 ml
  Distilled water……96 ml
- Silver Solution: 2.5% Silver nitrate…… 4 ml
  Distilled water……96 ml
- Developing Solution: Sodium carbonate……2.5 g
  0.19% Formaldehyde……..8 ml
  Distilled water……92 ml
- Stop Solution: Acetic acid……5 ml
  Distilled water……95 ml

Procedure:

- After completion of SDS-PAGE, carefully remove the gel from in-between the glass plates into a clean tray. Add 50 ml of Fixing solution and incubate for 10 minutes with constant moderate shaking.
- Discard the Fixative Solution. Wash with distilled water for 10 minutes.
- Add 50 ml of Fix/Sensitizing Solution and shake for 5 minutes.
- Discard the Fix/Sensitizing Solution. Wash the gel with 40% Ethanol for 20 minutes.
- Wash the gel with distilled water for 20 minutes.

- Add Sensitizing Solution and shake for one minute.
- Discard the Sensitizing Solution and wash the gel twice with distilled water for one minute.
- Incubate the gel in 50 ml of Silver Solution for 20 minutes ensuring that the gel is completely submerged in the liquid. Incubate with constant moderate shaking.
- Remove the gel from the Staining Solution and wash it with distilled water for one minute.
- Add the Developing solution and shake for 1minute. Remove this solution and add fresh. Incubate the gel in Developing Solution until color of separated protein bands develop and are distinctly visible.
- Remove the gel from the Developing Solution and immerse the gel in Stopping Solution for 5 minutes to obtain sharp distinct bands.
- Discard the Stop Solution and wash the gel with distilled water 2 times for 5 minutes.

Troubleshooting Guide:

Sr. No	Problem	Possible Cause	Solution
1	Gray or brown precipitate appears as smudges	Non-specific deposition of silver	Increase the washing steps.
2	Large discolored spots on the gel	Pressure on the gel during handling	Keep minimal contact with the gel while performing the protocol.
3	Slow or no development of bands	Developer solution is not mixed properly	Ensure the Developing Solution is mixed thoroughly before use
4	Low band sensitivity	Long exposure of Reducer Solution	Do not place the gel in reducer solution for more than 1 minute

Safety Information

The Silver Staining Kit is for laboratory use only, not for drug, household or other uses. Please refer the Material Safety Data Sheet (MSDS) for information regarding hazards and safe handling practices. Follow good clinical laboratory practices while handling clinical samples. Standard precautions should be followed as per established guidelines. Safety guidelines may be referred in safety data sheets of the product.

Performance and Evaluation

Performance of the kit is expected when the kit is stored at recommended temperature and within the expiry period.

Product Use Limitation & Warranty

HiMedia guarantees the performance of product in the manner described in the product literature. Silver Staining Kit is designed and sold for research and in vitro purposes only. The product is not to be used for human diagnostic or drug purposes or to be administered to humans unless expressed clearly for that purpose by the Food and Drug Administration or the appropriate regulatory authorities in the country of use. All due care and attention should be exercised in the handling of many of the materials described in the text.

Technical Assistance

At HiMedia, we pride ourselves on the quality and availability of our technical support. For any kind of technical assistance, mail at mb@himedialabs.com.



Storage temperature

Do not use if package is damaged

HiMedia Laboratories Private Limited,
Reg. Off: Plot No. C-40, Road No. 21Y,
MIDC, Wagle Industrial Area, Thane,
(West) 400604, Maharashtra, INDIA.
Web: www.himedialabs.com

02/2025

ML123_3/0222 ML123-03

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory,diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.