



## MacConkey Agar w/ 0.15% Bile salts, CV and NaCl Plate

MP081

### Intended Use:

Recommended for the selective isolation and differentiation of coliform organisms and other enteric pathogens from clinical and non-clinical samples.

### Composition\*\*

Ingredients	g / L
Gelatin peptone	17.000
Tryptone	1.500
Peptone	1.500
Lactose	10.000
Bile salts	1.500
Sodium chloride	5.000
Neutral red	0.030
Crystal violet	0.001
Agar	15.000
Final pH ( at 25°C)	7.1±0.2

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

### Principle And Interpretation

MacConkey agars are slightly selective and differential plating media mainly used for the detection and isolation of gram-negative organisms from clinical (1), dairy (2), food (3,4), water (5), pharmaceutical (6,7) and industrial sources (8). It is also recommended for the selection and recovery of the *Enterobacteriaceae* and related enteric gram-negative bacilli.

USP recommends this medium for use in the performance of Microbial Limit Tests (7).

These agar media are selective since the concentration of bile salts, which inhibit gram-positive microorganisms, is low in comparison with other enteric plating media. The medium M081, which corresponds with, that recommended by APHA can be used for the direct plating of water samples for coliform bacilli, for the examination of food samples for food poisoning organisms (4) and for the isolation of *Salmonella* and *Shigella* species in cheese (2). Other than that this medium is also used for count of coli-aerogenes bacteria in cattle and sheep faeces (9), the count of coli-aerogenes and non-lactose fermenters in poultry carcasses (9), bacterial counts on irradiated canned minced chicken (10) and the recognition of coli-aerogenes bacteria during investigations on the genus *Aeromonas* (11,12).

MacConkey Agar is the earliest selective and differential medium for cultivation of enteric microorganisms from a variety of clinical specimens (13,14). The original medium contains protein, bile salts, sodium chloride and two dyes. The selective action of this medium is attributed to crystal violet and bile salts, which are inhibitory to most species of gram-positive bacteria. Gram-negative bacteria usually grow well on the medium and are differentiated by their ability to ferment lactose. Lactose-fermenting strains grow as red or pink colonies and may be surrounded by a zone of acid precipitated bile. The red colour is due to production of acid from lactose, absorption of neutral red and a subsequent colour change of the dye when the pH of medium falls below 6.8. Lactose non-fermenting strains, such as *Shigella* and *Salmonella* are colourless, transparent and typically do not alter appearance of the medium.

Peptone, Tryptone and gelatin peptone are sources of nitrogen, carbon, long chain amino acids and other nutrients. Lactose is a fermentable carbohydrate, Sodium chloride maintains the osmotic equilibrium. Bile salts and crystal violet are selective agents that inhibit growth of gram-positive organisms. Neutral red is the pH indicator dye.

### Type of specimen

Clinical - faeces, urine and other pathological material, foodstuffs and dairy samples, water samples, pharmaceutical samples.

## Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (1,15). For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (2,4). For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards (5). For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per guidelines (6,7). After use, contaminated materials must be sterilized by autoclaving before discarding.

## Warning and Precautions

In Vitro diagnostic use. For professional use only. Read the label before opening the pack. Wear protective gloves/ protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

## Limitations

1. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium
2. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
3. It is recommended to store the plates at 24-30°C to avoid condensation.
4. Though the medium is recommended for selective isolation, further biochemical and serological testing must be carried out for further confirmation.
5. The surface of the medium should be dry when inoculated.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Sterile MacConkey Agar w/ 0.15% Bile salts, CV and NaCl in 90 mm disposable plates with smooth surface and absence of black particles/cracks/bubbles

### Colour of medium

Red with purplish tinge coloured medium

### Quantity of medium

27 ml of medium in 90 mm disposable plates.

### pH

6.90-7.30

### Sterility Check

Passes release criteria

### Cultural Response

Cultural response was observed after an incubation at 30-35°C for 18-72 hours. Recovery rate is considered as 100% for bacteria growth on Soybean Casein Digest Agar.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
<i>Corynebacterium diphtheriae type gravis</i>	$\geq 10^3$	inhibited	0%	
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50 -100	fair to good	30 -40 %	colourless
<i>Salmonella</i> Paratyphi A ATCC 9150	50 -100	luxuriant	$\geq 50$ %	colourless
<i>Salmonella</i> Abony NCTC 6017 (00029*)	50 -100	luxuriant	$\geq 50$ %	colourless
## <i>Proteus hauseri</i> ATCC 13315	50 -100	luxuriant	$\geq 50$ %	colourless
<i>Salmonella</i> Typhi ATCC 6539	50 -100	luxuriant	$\geq 50$ %	colourless

<i>Staphylococcus aureus</i> <i>subsp.aureus</i> ATCC 6538 (00032*)	$\geq 10^3$	Inhibited	0%	
<i>Salmonella</i> Paratyphi B ATCC 8759	50-100	luxuriant	$\geq 50\%$	colourless pink to red with bile precipitate
<i>Escherichia coli</i> ATCC 25922(00013*)	50-100	luxuriant	$\geq 50\%$	pink to red
# <i>Klebsiella aerogenes</i> ATCC 13048(00175*)	50-100	luxuriant	$\geq 50\%$	pink to red
<i>Salmonella</i> Typhimurium ATCC 14028(00031*)	50-100	luxuriant	$\geq 50\%$	colourless
<i>Enterococcus faecalis</i> ATCC 29212(00087*)	50-100	None-poor	$< 10\%$	colourless to pale pink
<i>Salmonella</i> Enteritidis ATCC 13076(00030*)	50-100	luxuriant	$\geq 50\%$	colourless
<i>Staphylococcus aureus</i> <i>subsp.aureus</i> ATCC 25923(00034*)	$\geq 10^3$	inhibited	0%	
<i>Staphylococcus</i> <i>epidermidis</i> ATCC 12228 (00036*)	$\geq 10^3$	inhibited	0%	

Key: (\*) Corresponding WDCM numbers,

(##) Formerly known as *Proteus vulgaris*

(#) Formerly known as *Enterobacter aerogenes*

## Storage and Shelf Life

On receipt store between 20-30°C. Use before expiry date on the label. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,15).

## Reference

- Murray P. R, Baron E, J., Jorgensen J. H., Tenover F. C., Tenover F. C., (Eds.), 8th Ed., 2003, Manual of Clinical Microbiology, ASM, Washington, D.C.
- Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
- FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, D.C.
- Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
- Lipps WC, Braun-Howland EB, Baxter TE, eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.
- The British Pharmacopoeia, 2022, Medicines and Healthcare products Regulatory Agency.
- The United States Pharmacopoeia-National Formulary (USP-NF), 2023.
- Williams, (Ed.), 2005, Official Methods of Analysis of the Association of Official Analytical Chemists, 19th Ed., AOAC, Washington, D.C
- Medrek T. F and Barnes Ella M., 1962, J. Appl. Bacteriol., 25(2),159-168.
- Thornley Margaret J., 1957, J. Appl. Bacteriol., 20(2), 273-285.
- Barnes Ella M. and Shrimpton D. H., 1957, J. Appl. Bacteriol., 20(2),273-285.
- Eddy B. P., 1960, J. Appl. Bacteriol., 23(2).216-249.
- MacConkey A., 1905, J. Hyg., 5:333.
- MacConkey A., 1900, The Lancet, ii:20.
- Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition


Revision: 04/2024



HiMedia Laboratories Pvt. Limited,  
Plot No.C-40, Road No.21Y,  
MIDC, Wagle Industrial Area,  
Thane (W) -400604, MS, India



*In vitro* diagnostic  
medical device



Storage temperature



CEpartner4U, Esdoornlaan 13,  
3951DB Maarn, NL  
[www.cepartner4u.eu](http://www.cepartner4u.eu)



CE Marking



Do not use if  
package is damaged

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.