



Technical Data

Soyabean Casein Digest Agar Plate w/Lecithin and Polysorbate 80 MP449GT (Tryptone Soya Agar Plate w/ Lecithin & Polysorbate 80) (γ -irradiated, Triple Pack)

Intended use

Recommended for determining efficiency of sanitization of containers, equipment, surfaces, water miscible cosmetics etc.

Composition**

Ingredients	g/ L
Tryptone	15.000
Soya Peptone	5.000
Sodium chloride	5.000
Lecithin	0.700
Polysorbate 80 (Tween 80)	5.000
Agar	15.000
Final pH (at 25°C)	7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

Principle And Interpretation

Soyabean Casein Digest Agar Plate w/Lecithin and Polysorbate 80 used in RODAC (Replicate Organism Detection and Counting) plates (1) for the detection and enumeration of microorganisms present on surfaces of sanitary importances (2,3). Tryptone and Soya peptone provide nitrogenous compounds and other nutrients essential for microbial replication. Lecithin and polysorbate 80 (Tween 80) are neutralizers reported to inactivate residual disinfectants from where the sample is collected (4). Lecithin neutralizes quaternary ammonium compounds and polysorbate 80 neutralizes phenolic disinfectants, hexachlorophene, formalin and with lecithin ethanol (5).

Collection of samples from areas before and after the treatment with disinfectant evaluates cleaning procedures in environmental sanitation. The presence and number of microorganisms is determined by the appearance of colonies on the agar surface (6). After counting the colonies, carry out biochemical testing for identification.

Type of specimen

Pharmaceutical samples, Swabs of containers, Equipment surfaces, Water miscible cosmetics etc.

Specimen Collection and Handling

For swabs of containers, equipment surfaces, water miscible cosmetics samples follow appropriate techniques for handling specimens as per established guidelines (1).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Individual strain of a microorganism may have unique growth requirements with respect to nutrients and physical conditions. Based on which the growth pattern of each varies on a medium and some even may display significant delay in development.
2. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
3. It is recommended to store the plates at 24-30°C to avoid minimum condensation.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Sterile Soyabean Casein Digest Agar w/Lecithin & Polysorbate 80 in 90 mm disposable plates with smooth surface and absence of black particles/cracks/bubbles. (γ irradiated, Triple Pack)

Colour of medium

Light yellow coloured medium

Quantity of medium

32 ml of medium in 90 mm disposable plates.

Reaction

7.10-7.50

Sterility Check

Passes release criteria

Dose of irradiation (Kgy)

13.00- 20.00

Cultural Response

Growth Promotion was observed after an incubation at 30-35°C for ≤ 3 days for bacteria and ≤ 5 days for fungus .

Recovery rate

Recovery rate is considered 100% for bacteria growth on Soyabean Casein Digest Agar and fungus growth on Sabouraud Dextrose Agar.

Growth promoting properties

Growth of microorganism comparable to that previously obtained with previously tested and approved lot of medium occurs at the specified temperature for not more than the shortest period of time specified inoculating ≤ 100 cfu (at 30-35°C for ≤ 18 hours).

Neutralization Test

The smaller zone of inhibition compared to SCDA indicates neutralization of quaternary ammonium compounds by this medium.

Organism	Inoculum (CFU)	Growth	Observed Lot value (CFU)	Recovery	Incubation temperature	Incubation period
Growth promoting						
** <i>Bacillus spizizenii</i> ATCC 6633 (00003*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
^ <i>Pseudomonas paraeruginosa</i> ATCC 9027 (00026*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Salmonella</i> Abony NCTC 6017	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
\$ <i>Kokuria rhizophila</i> ATCC 9341	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	18 -24 hrs
<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	≤ 5 d
<i>Candida albicans</i> ATCC 2091 (00055*)	50 -100	luxuriant	35 -100	≥ 70 %	30 -35 °C	≤ 5 d

# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50 -100	luxuriant	35 -100	>=70 %	30 -35 °C	<=5 d
<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	luxuriant	35 -100	>=70 %	20 -25 °C	<=5 d
<i>Candida albicans</i> ATCC 2091 (00055*)	50 -100	luxuriant	35 -100	>=70 %	20 -25 °C	<=5 d
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50 -100	luxuriant	35 -100	>=70 %	20 -25 °C	<=5 d
<i>Clostridium sporogenes</i> ATCC 19404 (00008*)	50 -100	luxuriant	35 -100	>=70 %	30 -35 °C	<=48 hours

Key : (*) Corresponding WDCM numbers.

**Formerly known as *Bacillus subtilis* subsp. *spizizenii*

^ Formerly known as *Pseudomonas aeruginosa*

\$ Formerly known as *Micrococcus luteus*

Formerly known as *Aspergillus niger*

Storage and Shelf Life

On receipt store between 20-30°C. Use before expiry date on the label stated. Product performance is best if used within expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference

- Hall and Hartnett, 1964, Public Hlth. Rep., 79:1021.
- MacFaddin J.F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Wilkins, Baltimore.
- Richardson (Ed), 1985, Standard Methods for the Examination of Dairy Products, 15th ed., APHA, Washington, D.C.
- Favero (Chairman), 1967, Biological Contamination Control Committee, a state of the art report., Am. Assoc. for contamination control.
- Brummer, 1976, Appl. Environ. Microbiol., 32:80.
- Murray PR, Baron, Pfaller, and Tenover (Eds.), 2003, In Manual of Clinical Microbiology, 8th ed., ASM, Washington, D.C.
- Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 05/ 2024

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.