

Candida Selective Agar Plate

MP5477

Intended Use

Recommended for isolation of *Candida* species (appears as white coloured) from clinical samples.

Composition**

Ingredients	g / L
Peptonized SM powder	25.000
Malt extract	10.000
Dextrose (Glucose)	15.000
Selective mixture	0.500
Agar	15.000
Can Selective Supplement, Modified (FD824)	1 vial
Selective agent	450mg
pH after sterilization(at 25°C)	5.6±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

Principle And Interpretation

Candidiasis is a fungal infection caused by a yeast (a type of fungus) called *Candida*. White fungus causes an infection in the lungs similar to COVID-19. *Candida* species responsible for the infection are *Candida albicans* and *Candida auris*. It usually affects nose, mouth, lungs and stomach or nail beds. People suffering from COVID-19, HIV/AIDS and other viral diseases, congenital bone marrow disease, cancers and untreated or irregularly treated diabetes have reduced immunity and are prone to acquire the infection.

This medium is specifically designed to promote rapid and selective growth of *Candida* strains. Peptonized SM powder and malt extract serves as a source of nitrogenous and carbonaceous compounds, long chain amino acids, vitamins and other essential nutrients. Dextrose in the medium serves as a rich source of energy. Selective mixture suppresses the growth of mycelial fungi and inhibits bacterial flora.

Type of specimen

Clinical samples : eye lesion, nasal swabs, other site of infections etc.

Specimen Collection and Handling

Specimens from the eye, nose, nasopharynx and other sites of infection are usually collected with the help of sterile swab and transported to the lab by using HiFungal™ Transport medium w/swab (MS5478) which contains the transport medium along with the swab. Specimen collection should be carried out by trained personnel. After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precaution

In Vitro Diagnostic use only. For professional use only. Read the label before opening the pack. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling clinical specimens and culture. Standard guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Individual fungi differ in their growth requirement and therefore show variable growth patterns on the medium.
2. Further biochemical and serological studies are to be carried out for complete identification.

Performance and Evaluation

Performance of the medium is expected when used as per the direction within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Sterile Candida Selective Agar in 90 mm disposable plates with smooth surface and absence of black particles/cracks/bubbles

Colour of medium

Light yellow to amber coloured medium

Quantity of medium

25 ml of medium in 90 mm disposable plates.

pH

5.40-5.80

Sterility Check

Passes release criteria

Growth Promotion Test

Cultural characteristics observed after an incubation at 25-30°C for 24-48 hours.

Cultural Response

Organism	Growth
<i>Candida albicans</i> ATCC 10231 (00054*)	luxuriant
<i>Candida auris</i> CDC B 11903	luxuriant
<i>Mucor racemosus</i> ATCC 42647	none-poor
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	none-poor
<i>Escherichia coli</i> ATCC 25922 (00013*)	inhibited
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	inhibited
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00087*)	inhibited

Key : (#) - Formerly known as *Aspergillus niger*, (*) - corresponding WDCM numbers

Storage and Shelf Life

On receipt store between 20-30°C. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1,2).

Reference

1. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
2. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 01/2024



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



In vitro diagnostic
medical device



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



Do not use if
package is damaged

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.