

## Tryptone Soya Agar Plate

MPH290

### Intended use

Recommended as a general purpose medium used for cultivation of a wide variety of microorganisms from pharmaceutical products in accordance with harmonized method of USP/EP/BP/JP/IP (Medium 2).

### Composition\*\*

Ingredients	Gms / Litre
Tryptone #	15.000
Soya peptone ##	5.000
Sodium chloride	5.000
Agar	15.000
pH after sterilization ( at 25°C)	7.3±0.2

\*\*Formula adjusted, standardized to suit performance parameters

# Pancreatic digest of casein

## Papaic digest of soyabean (soybean)

### Directions

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

### Principle And Interpretation

Various pharmacopoeias recommend Tryptone Soya Agar Plate as sterility testing medium. It is also used in validation of sterility checking procedure in accordance with the microbial limit testing harmonized methodology of USP/EP/BP/JP/IP (1,2,3,4,5). This medium is used in microbial limit test and antimicrobial preservative- effective test. Gunn et al (5) used this medium for the growth of fastidious organisms and study of haemolytic reaction after addition of 5% v/v blood. The combination of tryptone and soya peptone makes these media nutritious by providing amino acids and long chain peptides for the growth of microorganisms. Natural sugars of soy enhance growth of microorganism. Sodium chloride maintains the osmotic balance in the medium. Agar is the solidifying agent.

The total aerobic count is considered to be equal to the number of colony forming units found on this medium, if colonies of fungi are detected on this medium they are counted along with total aerobic count.

### Type of specimen

Pharmaceutical samples

### Specimen Collection and Handling

For pharmaceutical samples, follow appropriate techniques for sample collection, processing as per pharmaceutical guidelines (1,2,3,4,5).

After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

Read the label before opening the pack. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations

1. Individual strain of a microorganism may have unique growth requirements with respect to nutrients and physical conditions. Based on which the growth pattern of each varies on a medium and some even may display significant delay in development.

2.Environmental Monitoring Test : Exposure of media plates for 4 h as a settle plate or in air sampler or even under laminar air flow may lead reduction in some available moisture on the surface. This may cause development of tiny cracks in the agar or slight shrinkage. This however, does not impact the performance of the media.

3. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.

4. It is recommended to store the plates at 24-30°C to avoid minimum condensation.

### Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

### Quality Control

#### Appearance

Sterile Tryptone Soya Agar in 90 mm disposable plates.

#### Colour of medium

Light yellow coloured medium

#### Quantity of medium

25 ml of medium in 90 mm disposable plates.

#### pH

7.10-7.50

#### Sterility Test

Passes release criteria

#### Growth Promotion Test

Growth Promotion was carried out in accordance with the harmonized method of USP/EP/BP/JP, and growth was observed after an incubation at 30-35°C for 18-24 hours. Recovery rate is considered 100% for bacteria growth on Blood Agar and fungus growth on Sabouraud Dextrose Agar.

#### Growth promoting properties

Growth of microorganism comparable to that previously obtained with previously tested and approved lot of medium occurs at the specified temperature for not more than the shortest period of time specified inoculating  $\leq 100$  cfu (at 30-35°C for 18 hours).

#### Cultural Response

Organism	Inoculum (CFU)	Observed Lot value (CFU)	Recovery	Incubation period
<i>Bacillus subtilis subsp. spizizenii</i> ATCC 6633 (00003*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Staphylococcus aureus subsp. aureus</i> ATCC 25923 (00034*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Staphylococcus aureus subsp. aureus</i> ATCC 6538 (00032*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Escherichia coli</i> ATCC 25922 (00013*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Escherichia coli</i> ATCC 8739 (00012*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Escherichia coli</i> NCTC 9002	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs
<i>Salmonella</i> Abony NCTC 6017 (00029*)	50 -100	35 -100	$\geq 70$ %	18 -24 hrs

<i>Micrococcus luteus</i> ATCC 9341	50 -100	35 -100	>=70 %	18 -24 hrs
<i>Streptococcus pneumoniae</i> ATCC 6305	50 -100	35 -100	>=70 %	18 -24 hrs
<i>Salmonella</i> Typhimurium ATCC 14028 (00031*)	50 -100	35 -100	>=70 %	18 -24 hrs
<i>Candida albicans</i> ATCC 10231 (00054*)	50 -100	35 -100	>=70 %	<=5 d
<i>Candida albicans</i> ATCC 2091 (00055*)	50 -100	35 -100	>=70 %	<=5 d
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053*)	50 -100	25 -70	50-70 %	<=5 d

Key : (#) Formerly known as *Aspergillus niger*, (\*) Corresponding WDCM numbers

## Storage and Shelf Life

On receipt store between 20-30°C Use before expiry date on the label. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

## Reference

1. British Pharmacopoeia, 2019, The Stationery office British Pharmacopoeia
2. European Pharmacopoeia, 2019, European Dept. for the quality of Medicines.
3. Indian Pharmacopoeia, 2018, Govt. of India, the controller of Publication, Delhi, India.
4. The United States Pharmacopoeia, 2019, The United States Pharmacopoeial Convention. Rockville, MD.
5. Japanese Pharmacopoeia, 2016.
6. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
7. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 04/ 2022

## Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.