

## Streptococcus Selection HiVeg™ Broth

MV303

### Intended Use:

Recommended for selective isolation and cultivation of Streptococci including group A β- haemolytic strains.

### Composition\*\*

Ingredients	g / L
HiVeg™ hydrolysate	15.000
Soya peptone	5.000
Sodium chloride	4.000
Sodium citrate	1.000
L-Cystine	0.200
Sodium sulphite	0.200
Dextrose (Glucose)	5.000
Sodium azide	0.200
Crystal violet	0.0002
Final pH ( at 25°C)	7.4±0.2

\*\*Formula adjusted, standardized to suit performance parameters

### Directions

Suspend 30.6 grams in 1000 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Mix well and dispense into sterile tubes or sterile flasks as desired. Autoclaving is not required if medium is used on the same day. If storage is desired, sterilize by autoclaving at  $\Delta 118^{\circ}\text{C}$  for 15 minutes. Avoid overheating which causes the medium to become more inhibitory.  $\Delta$  Corresponds to 12lbs pressure

### Principle And Interpretation

Streptococcus Selection Broth is based on the suggestion of Pike (1), for the selective isolation of Streptococci from various materials, specially those which are heavily contaminated with accompanying microbial flora (2). Welch et al (3) also reported the abilities of this medium to recover group A β-haemolytic Streptococci. Streptococcus Selection HiVeg™ Broth is prepared by using vegetable peptones in place of animal based peptones which make the media free of BSE/TSE risks.

HiVeg™ hydrolysate, Soya peptone, dextrose and salts provide nutrients essential for the growth of Streptococci. Sodium azide, sodium sulphite inhibits gram-negative rods and the crystal violet suppresses Staphylococci. However, Streptococci are not affected by these inhibitors at these concentrations. Due to this reason, these media are useful in studies of streptococcal flora from nutritional, dental and epidemiological research. Growth of coliforms, *Proteus*, *Pseudomonas* and *Bacillus* species is markedly suppressed in this medium. However, some strains of Staphylococci and Pneumococci may grow in this medium.

### Type of specimen

Water samples

### Specimen Collection and Handling:

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(5) After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions :

Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

### Limitations :

1. All presumptive streptococci must be confirmed for identification.
2. Further biochemical and serological tests need to be carried out for confirmation.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Cream to yellow homogeneous free flowing powder

### Colour and Clarity of prepared medium

Light to medium amber coloured clear to slightly opalescent solution in tubes

### Reaction

Reaction of 3.06% w/v aqueous solution at 25°C. pH : 7.4±0.2

### pH

7.20-7.60

### Cultural Response

Cultural characteristics observed after an incubation at 35 - 37°C for 18 - 24 hours.

Organism	Inoculum (CFU)	Growth
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (00003*)	≥10 <sup>4</sup>	inhibited
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	none-poor
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50-100	luxuriant
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	≥10 <sup>4</sup>	inhibited
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	50-100	none-poor
<i>Streptococcus pyogenes</i> ATCC 19615	50-100	luxuriant

Key : \*Corresponding WDCM numbers.

## Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-25°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

## Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

## Reference

1. Pike R.M., 1945, Am. J. Hyg., 41:211.
2. Facklam and Carly, 1985, Manual of Clinical Microbiology, Lennette and others (Eds.), 4th ed., ASM, Washington D.C.
3. Welch D.F. et al, 1991, Am. J. Clin. Pathol., 95:587.
4. Lipps WC, Braun-Howland EB, Baxter TE, eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.
5. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

Revision : 03/2024

**Disclaimer :**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.